

Service Repair Sheet SRS68298

Contact Name

Michael Carrick

Company/ Hospital Name

Royal London Hospital

Department

Position

Senior Clinical Technologist

Direct Phone

02035942240

General Phone

Opera Account

00003030

Email

Michael.carrick@nhs.net

Order Number

Date Received

29/Sep/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC

Date: 29/09/22

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 26/Sep/2022 Kate Griffiths

26/Sep/2022 Kate Griffiths

To send in V1000, PR065A1 for service and calibration

29/Sep/2022 Robert Connor

Received 1 x V1000 s/n PR065A17, with 4 x AA battery and blue fabric case.

Ready For quote

CGreen 30-9-22

Repair Complete Signed

CGreen 5-10-22 CGreen 17-10-22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35473	Foetal Simulator	1410000	PR065A17	N

1480000 x 1 @ £45

SIN, SRS, SRN

1430309 x 1 FOC — Added 4-10-22
missed off by mistake CG 4-10-22

SRS, SRN

UPS x 1 @ £10 if done today

or £12 if making quote on 31/10/22.

Service Repair Sheet 68298

Contact Name	Michael Carrick
Company/ Hospital Name	Royal London Hospital
Position	Senior Clinical Technologist
Direct Phone	02035942240
Opera Account	00003030
Email	Michael.carrick@nhs.net
Order Number	40807782
Date Received	29/Sep/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35473	1410000	PR065A17	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

Barts Health NHS Trust

OFFICIAL ORDER

Order Date : 07-10-2022

Order No : **40807782**

Must be quoted on all correspondence.

Deliver To :

CLINICAL EQUIPMENT MANAGEMENT (CEMS)

Royal London Hospital

Whitechapel Road

Whitechapel

GTL

E1 1FR

GB

Requested delivery date: 13-10-2022

Location ID: AN1647 CLINICAL EQUIPMENT
MANAGEMENT (CEMS)

Invoice and Payment Enquiries To

TREASURY & PAYMENTS DEPARTMENT

8TH FLOOR, 20 CHURCHILL PLACE

EORI: GB654 9470 05 000

LONDON

E14 5HJ

GB

Fax: apinvoices.bartshealth@nhs.net

All enquiries regarding this order to:

Contact : Carrick, Michael

Telephone :

Facsimile No. :

Email Address : michael.carrick@nhs.net

Warning : Order has at least one service line. Please process manually.

Buyer Contact : Franz, Graeme

Buyer Email : barts.procurement@nhs.net

Supplier

Viamed Ltd

Customer's Supplier Name:
VIAMED

Conditions

THIS ORDER IS SUBJECT TO STANDARD NHS TERMS AND CONDITIONS. IF PRICES STATED ON THIS ORDER ARE INCORRECT ANY REVISED PRICES MUST BE AUTHORISED BY THE BUYER PRIOR TO ORDER EXECUTION. PAYMENT WILL BE MADE AT THE PRICES STATED HEREIN. DO NOT ASSIGN THIS ORDER SPECIAL INSTRUCTIONS.

Line	Goods or Services Required	Quantity	UOM	Contract Ref.	Unit Price	Line Value	VAT
1	N/A Foetal Heart Simulator V1000 - ServiceQuotation QVM139516Customer Reference SRS68298	45	EA		£1.00	£45.00	-

Net Total : £45.00

Carriage : -

Tax : -

Total : £45.00