

## Classification Rationale

<b>Product Family</b>	Microstim DBS Mark3		
<b>Type</b>	<b>Rule</b>	<b>Class</b>	<b>Additional Information</b>
<b>Non-Invasive</b>	Yes	IIa	
<b>Active</b>	Yes - Rule 10		Supplies electric energy to make muscles twitch, as an indication of depth of Anaesthesia. Has been compared to a reflex hammer
<b>Invasive</b>	No		
<b>Special</b>	No		

BSI Flow chart includes : **Yes** or No

Classified by: .....

Date: ...../...../.....