

Service Repair Sheet SRS68250

Contact Name

Magnus Persson

Company/ Hospital Name

Swevet - AB

Department

Position

Direct Phone

46 416 258 21

General Phone

46 771215500

Opera Account

00006308

Email

magnus.persson@swevet.se

Order Number

Date Received

22/Aug/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC

Date: 22/8/22

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes Parcel on its way via postnord, tracking no. 03015130012954

22/Aug/2022 Robert Connor

Received 2 x VM-2500-S-Vet s/n 70840141, with 4 x AA battery and sampling line, s/n A1906160004, with li-poly battery, clear silicone cover and sampling line.

Ready For quote

CGneen 25.8.22

Repair Complete Signed

CGneen 26.10.22

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35219	Capnograph	4410540	70840141	N
SRN35220	Capnograph	4410540	A1906160004	N
SRN35221	Pulse Oximetry Y Sensor	0015002	VYV200015	N

€
Dist

4440000 x 2 @ 0.5hr each €70 per hr
S/N's, SRS, SRN's

4430201 x 2 @ €17.99 each.

SRS's, SRN's

~~446~~ 4430200 x 2 @ €14.68.

SRS's x SRN's

0040001 x 1 @ €0 → Please add to order cg 26.10.22
S/N, SRS, SRN

UPS x 1
32x24x16cm



Proforma invoice No F8011

Ship to: Date 2022-08-08

Shippers name:

Viamed Ltd
15 Station Road
Cross Hill Keighley
BD20 7DT West Yorkshire
Contact : Zoey Teal

Swevet AB
Forskaregatan 1D

SE-275 37 SJÖBO
Contact: Magnus Persson
Mail: magnus.persson@swevet.se

<http://www.viamed.co.uk>
Email: zoey.teal@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

This is veterinary equipment for repair, Equipment will return.

Nr	Description	
4410540	Viamed CO2 & SpO2 VM-2500-S	2 Pc
	S/n: A1906160004 & 70840141	Value : 3 580 EUR
	Ref. SRS68250	



Magnus Persson
2022-08-08

Tel +46 416-25821
Orgnr 556621-5843
VAT SE556621584301

Tillverk land : GB
Vikt : 3kg
Tariffnr : 90189084
VAT Reg No: GB287389593
Eori No: GB287389593000

magnus.persson@swevet.se

Service Repair Sheet 68250

Contact Name	Magnus Persson
Company/ Hospital Name	Swevet - AB
Direct Phone	46 416 258 21
General Phone	46 771215500
Opera Account	00006308
Email	magnus.persson@swevet.se
Date Received	22/Aug/2022
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN35219	4410540	70840141	Capnograph	Complete - Repaired Time :0.5 Hour(s)

Parts Replaced

Qty: 1	4430200
Qty: 1	4430201
Qty: 0.5	4440000

The Capnograph VM2500-S Vet has been repaired. The exhaust assembly has been replaced.

Repair	Ref	S/N	Equipment Type	
SRN35220	4410540	A1906160004	Capnograph	Complete - Repaired Time :0.5 Hour(s)

Parts Replaced

Qty: 1	4430200
Qty: 1	4430201
Qty: 0.5	4440000

The Capnograph VM2500-S Vet has been repaired. The exhaust assembly has been replaced.

Repair	Ref	S/N	Equipment Type	
SRN35221	0015002	VYV200015	Pulse Oximetry Y Sensor	No Fault Found Time :0 Hour(s)

The pulse oximetry Y sensor has been tested. No fault was found.
