



Risk Assessment iaw EN ISO 14971:2000 Annex D: Possible hazards with medical devices.

Ref.	Hazard.	Related part /	Sev	Like	Risk.	Solution.	Document referenced.	Sev	Like	Risk.
		Component posing	of	of				of	0f	
		risk.	Haz.	Haz.				Haz.	Haz.	
	Energy hazards and con									
D.2.1	Electricity	N/A	1	1	1			1	1	1
D.2.2	Heat	If battery polarity	3	1	3	Protection diode added in	(O) Design file	1	1	1
		reversed				circuit.				
D.2.3	Mechanical force	Instrument	2	1	2			2	1	2
D.2.4	Ionising radiation	N/A	1	1	1			1	1	1
D.2.5	Non ionising radiation	N/A	1	1	1			1	1	1
D.2.6	Moving parts	N/A	1	1	1			1	1	1
D.2.7	Unintended motion	N/A	1	1	1			1	1	1
D.2.8	Suspended masses	N/A	1	1	1			1	1	1
D.2.9	Patient support failure	N/A	1	1	1			1	1	1
D.2.10	Pressure (vessel rupture)	N/A	1	1	1			1	1	1
D.2.11	Acoustic pressure	N/A	1	1	1			1	1	1
D.2.12	Vibration	N/A	1	1	1			1	1	1
D.2.13	Magnetic fields	N/A	2	2	4		F. User manual / Technical	2	1	2
	(eg. MRI)						manual			
	Biological hazards and c									
D.3.1	Bio-contamination	Instrument & leads	2	2	4	Easy to clean – cleaning	F. User manual / Technical	1	2	2
						instructions in user manual	manual			
D.3.2	Bio-incompatibility	Headbox	1	1	1		Manufacturers data	1	1	1
D.3.3	Incorrect formulation	N/A	1	1	1			1	1	1
	(chemical composition)									
D.3.4	Toxicity	N/A	1	1	1		Manufacturers data	1	1	1
D.3.5	Allergenicity	N/A	1	1	1		Manufacturers data	1	1	1
D.3.6	Mutagenicity	N/A	1	1	1		Manufacturers data	1	1	1
D.3.7	Oncogenicity	N/A	1	1	1		Manufacturers data	1	1	1

T:\Simon Watmough\Projects\CE files\Microstim\E. Risk Management BS14971 Annex D_Microstim_incomplete.doc Page 1 23/04/03 Severity of hazard: 1 = Insignificant, 2 = Tolerable, 3 = Critical, 4 = Intolerable. Probability of event: 1 = Improbable, 2 = Occasional, 3 = Likely, 4 = Highly likely. Risk calculated as severity of hazard x probability of event, 1 - 16. Further explanation of risk management policy - see Risk Management Policy & Definitions.



D.3.8	Carcinogenicity	N/A	1	1	1		Manufacturers data	1	1	1
	Re and/or cross infection	Instrument & leads	1	1	1		Manufacturers data	1	1	1
	Pyrogenicity	N/A	1	1	1		Manufacturers data	1	1	1
	Inability to maintain hygienic standards	Instrument & leads	2	2	4	Easy to clean – cleaning instructions in user manual	F. User manual / Technical manual	2	1	2
D.3.12	Degradation	Instrument & leads	2	1	2	Care instructions given in the user manual	F. User manual / Technical manual	1	1	1
D.4	Environmental hazards	and contributory fac	tors		•					
D.4.1	Electromagnetic fields	N/A	1	1	1			1	1	1
D.4.2	Susceptibility to electromagnetic interference	N/A	1	1	1			1	1	1
D.4.3	Emissions of electromagnetic interference	N/A	1	1	1			1	1	1
D.4.4	Inadequate supply of power	N/A	1	1	1			1	1	1
D.4.5	Inadequate supply of coolant	N/A	1	1	1			1	1	1
D.4.6	Storage / operation outside prescribed environmental conditions	Instrument	2	1	2	No specific environmental storage / operating conditions imposed		2	1	2
D.4.7	Incompatibility with other devices with which the product is intended to be used	ECG pads	1	1	1	Standard sized ECG press studs on leads		1	1	1
D.4.8	Accidental mechanical damage	Instrument & leads	2	1	2	Relatively robust material used. If damaged, user to assess level of damage – return for repair or spares parts available.	F. User manual / Technical manual	1	1	1

T:\Simon Watmough\Projects\CE files\Microstim\E. Risk Management BS14971 Annex D_Microstim_incomplete.doc Page 2 23/04/03 Severity of hazard: 1 = Insignificant, 2 = Tolerable, 3 = Critical, 4 = Intolerable. Probability of event: 1 = Improbable, 2 = Occasional, 3 = Likely, 4 = Highly likely. Risk calculated as severity of hazard x probability of event, 1 - 16. Further explanation of risk management policy - see Risk Management Policy & Definitions.





D.4.9	Contamination due to	Instrument & leads	1	1	1	No special disposal		1	1	1
	waste products and/or					required. Alkaline				
	device disposal					batteries used.				
D.5	Hazards resulting from	incorrect output of e	nergy a	nd sub	stance	S				
D.5.1	Electricity	N/A	1	1	1			1	1	1
D.5.2	Radiation	N/A	1	1	1			1	1	1
D.5.3	Volume	N/A	1	1	1			1	1	1
D.5.4	Pressure	N/A	1	1	1			1	1	1
D.5.5	Supply of medical gases	N/A	1	1	1			1	1	1
D.5.6	Supply of anaesthetic	N/A	1	1	1			1	1	1
	agents									
D.6	D.6 Hazards related to the use of the medical device and contributory factors									
D.6.1	Inadequate labelling	Label	2	1	2	Product easy to use	F. Labelling	1	1	1
						Label				
D.6.2	Inadequate operating	User manual	2	1	2	Product easy to use	F. User manual / Technical	1	1	1
	instructions					User manual / Technical	manual			
						manual				
D.6.3	Inadequate specification	N/A	1	1	1			1	1	1
-	of accessories			-	<u> </u>			-		
D.6.4	Inadequate specification	User manual / label	2	2	4	Product easy to use	F. User manual / Technical	2	1	2
	of pre-use checks				10	User manual / label	manual / label			
D.6.5	Over-complicated	User manual	2	1	2	Product easy to use	F. User manual / Technical	1	1	1 1
	operating instructions					User manual / Technical	manual			
			_			manual				
D.6.6	Inadequate specification	N/A	1	1	1	Instrument set up at		1	1	1
	of service and					manufacturer. Does not				
	maintenance					require regular service.				
						Maintenance only required				
						if performance degrades.				

T:\Simon Watmough\Projects\CE files\Microstim\E. Risk Management BS14971 Annex D_Microstim_incomplete.doc Page 3 23/04/03 Severity of hazard: 1 = Insignificant, 2 = Tolerable, 3 = Critical, 4 = Intolerable. Probability of event: 1 = Improbable, 2 = Occasional, 3 = Likely, 4 = Highly likely. Risk calculated as severity of hazard x probability of event, 1 - 16. Further explanation of risk management policy - see Risk Management Policy & Definitions.



D.6.7	Use by unskilled / untrained personnel	Instrument & leads	3	2	6	Product easy to use Warning in User manual / Technical manual – not for	F. User manual / Technical manual	2	1	2
D.6.8	Reasonable foreseeable misuse	Instrument & leads	3	2	6	use by untrained personnel Product easy to use Warning in User manual / Technical manual – not for use by untrained personnel	F. User manual / Technical manual	2	1	2
D.6.9	Insufficient warning of side effects	N/A	1	1	1			1	1	1
D.6.10	Inadequate warnings of hazards likely with reuse of single use devices	N/A	1	1	1			1	1	1
D.6.11	Incorrect measurement and other metrological aspects	N/A	1	1	1			1	1	1
D.6.12	Misrepresentation of results	Instrument & leads	1	1	1	Only for use by qualified & trained personnel – warning in user / technical manual. In depth explanation in user / technical manual.	F. User manual / Technical manual	1	1	1
D.6.13	Incompatibility with consumables / accessories / other devices	Battery & leads	1	1	1	Standard battery used. Leads colour coded – spares available	F. User manual / Technical manual	1	1	1
	Sharp edges or points	Instrument & leads	2	2	4	Only if damaged, user to assess level of damage – return for repair or spares parts available.	F. User manual / Technical manual	1	1	1
D.7	Inappropriate, inadequa	ite or overcomplicate	d user	interfa	ce (ma	an/machine communication	1)			

T:\Simon Watmough\Projects\CE files\Microstim\E. Risk Management BS14971 Annex D_Microstim_incomplete.doc Page 4 23/04/03 Severity of hazard: 1 = Insignificant, 2 = Tolerable, 3 = Critical, 4 = Intolerable. Probability of event: 1 = Improbable, 2 = Occasional, 3 = Likely, 4 = Highly likely. Risk calculated as severity of hazard x probability of event, 1 - 16. Further explanation of risk management policy - see Risk Management Policy & Definitions.





D 5 1	[3.6" . 1	>T/ i		-		[37] 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	- 4		4
D.7.1	Mistakes & judgement errors	N/A	1	1	1	No complex user interface	1	1	1
D.7.2	Lapses and cognitive recall errors	N/A	1	1	1	No complex user interface	1	1	1
D.7.3	Slips & blunders (mental or physical)	N/A	1	1	1	No complex user interface	1	1	1
D.7.4	Violation or abbreviation of instructions, procedures etc	N/A	1	1	1	No complex user interface	1	1	1
D.7.5	Complex or confusing control system	N/A	1	1	1	No complex user interface	1	1	1
D.7.6	Ambiguous or unclear device state	N/A	1	1	1	No complex user interface	1	1	1
D.7.7	Ambiguous or unclear presentation of settings, measurement, or other information	N/A	1	1	1	No complex user interface	1	1	1
D.7.8	Misrepresentation of results	N/A	1	1	1	No results presented	1	1	1
D.7.9	Insufficient visibility, audibility or tactility	N/A	1	1	1	No complex user interface	1	1	1
D.7.10	Poor mapping of controls to action or of displayed information to actual state	N/A	1	1	1	No complex user interface	1	1	1
D.7.11	Controversial modes or mappings as compared to existing equipment	N/A	1	1	1	No complex user interface	1	1	1
D.8	Hazards arising from fun	iction failure, main	tenance	and ag	geing a	nd contributory factors			
D.8.1	Erroneous data transfer	N/A	1	1	1		1	1	1

T:\Simon Watmough\Projects\CE files\Microstim\E. Risk Management BS14971 Annex D_Microstim_incomplete.doc Page 5 23/04/03 Severity of hazard: 1 = Insignificant, 2 = Tolerable, 3 = Critical, 4 = Intolerable. Probability of event: 1 = Improbable, 2 = Occasional, 3 = Likely, 4 = Highly likely. Risk calculated as severity of hazard x probability of event, 1 - 16. Further explanation of risk management policy - see Risk Management Policy & Definitions.





D.8.2	Lack of, or inadequate specification for maintenance including post maintenance functional tests	N/A	1	1	1		1	1	1
D.8.4	Inadequate maintenance	N/A	1	1	1	Instrument set up at manufacturer. Does not require regular service. Maintenance only required if performance degrades.	1	1	1
D.8.5	Lack of adequate determination of end of device life	Instrument	1	1	1	User decision based on damage or cost effectiveness of repair.	1	1	1
D.8.6	Loss of electrical integrity	Instrument	2	2	4	Failure of any electrical component will result in loss of audio or visual output indication	1	1	1
D.8.7	Loss of mechanical integrity	Instrument	1	1	1	User decision based on damage or cost effectiveness of repair.	1	1	1
D.8.8	Inadequate packaging (contamination and / or deterioration of the device)	Instrument &leads	2	2	4	Boxed in protective plastic case. Soft and hard carrying cases available. Bubble wrap / cardboard box through mail / courier	1	1	1
D.8.9	Re-use and/or improper re-use	Instrument &leads	1	1	1	User decision on suitability for next use based on pre use check & tolerable damage to casing.	1	1	1

T:\Simon Watmough\Projects\CE files\Microstim\E. Risk Management BS14971 Annex D_Microstim_incomplete.doc Page 6 23/04/03 Severity of hazard: 1 = Insignificant, 2 = Tolerable, 3 = Critical, 4 = Intolerable. Probability of event: 1 = Improbable, 2 = Occasional, 3 = Likely, 4 = Highly likely. Risk calculated as severity of hazard x probability of event, 1 - 16. Further explanation of risk management policy - see Risk Management Policy & Definitions.



CE file: Microstim.

D.8.10	Deterioration in function	Instrument &leads	1	1	1	Degradation of electrical	1	1	1
	(gradual occlusion of					components will			
	fluid / gas path or change					eventually result in loss of			
	in resistance to flow,					audio or visual output			
	electrical conductivity)					indication			
	as a result of repeated								
	use								