



Kate Griffiths <viamed.kate.griffiths@gmail.com>

Prosol order 528138

1 message

Steve Hardaker <office@viamed.co.uk>
Reply-To: steve.hardaker@viamed.co.uk
To: kate.griffiths@viamed.co.uk

3 November 2022 at 15:06

Hi Kate,

I spoke to Steve Machin at Prosol about their order 528138 for 20x R-21A and 50x R-17A-LVG as we have neither in stock.

I explained that due to the intermediate thread, they can use their Prosol branded 0111365 / GOC5223 sensor in place of the R-21A, so he said to cancel those off the order.

He wants to leave the R-17-LVG for now as they still have some in stock, but is aware of the Teledyne option and will ring me when he gets down to a week or so left in stock.

Best Regards,

Steve Hardaker
Technical Support Manager
Viamed Ltd.

Please note: Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

Website: www.viamed.co.uk
email: steve.hardaker@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up
Twitter: twitter.com/ViamedLtd
Facebook: facebook.com/ViamedLtdUK
Company registered in England, No. 1291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542
GDPR / Privacy Policy available at www.viamed.co.uk/policy
