Service Repair Sheet SRS68348

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Paul Denyer

Crawley Hospital

**EME** Department

**EME Services Technician** 

01293 600300

00001300

paul.denyer@nhs.net

19/Dec/2022

Robert Connor

Viamed

Quote

Goods In Only Decontamination certificate provided by customer  Cleaned by Viamed, f no declaration certificate rom customer Signed: Date:  Goods Out Only Cleaned by Viamed before returning to customer	VIAN	1€D <b>cl</b> e	ean
f no declaration certificate from customer Signed: Date:  Goods Out Only Cleaned by Viamed before	Decontam certificate	ination provided	Z
Cleaned by Viamed before	f no declaration custom Signed:	tion certificate	1
			oforo
Signed:			

Notes 13/Dec/2022 Sophie Lines

13/Dec/2022 Sophie Lines

Returning Microstim DB3 - serial number M0005254 - has the following faults:

When the P.T.C 1 Hz switch is activated you don't always get an output voltage, this is the case if the switch is operated in either direction. The other switch also occasionally fails to operate.

I advised the fixed cost repair service

19/Dec/2022 Robert Connor

Received 1 x DB3 s/n M0005254, with 9v battery

Ready For quote

CGreen 3-1-23

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35559	Microstim DB3	2510000	M0005254	7

2540010 X1 EAS 5/N, 5RS, SRN

UPS X10 E12.



## **EME Services Equipment Work Ticket**



**EME Services** 

Email Address: sash.eme.helpdesk@nhs.net

13 December 2022

Equipment:

STIMULATOR NERVE

Asset Number:

ES27082

Safety Bulletins Outstanding: 0

Serial Number:

M0005254

Reported Date: 08/12/2022

Model:

MICROSTIM DB3

Batteries last changed: Bat1

Class and Type:

BF

Bat2

Manufacturer:

VIAMED LTD

Next PPM: 26/10/2023

Equipment Location:

THEATRES - CH

Service Book Location: 0446 - ESH 0000 - C

Reg No:

EME Contact Name: PAUL DENYER

Return Address:

EME SERVICES - CRAWLEY HOSPITAL

Email Address:

sash.eme.helpdesk@nhs.net

Fault:

NOT CONDUCTING NO GREEN INDICATION LIGHTS WORKING

**OUTPUT IS VERY INTERMITTENT 13/12/2022PD** 

Stores Ordered / Action Taken:

SENT TO VIAMED FOR REPAIR QUOTATION 13/12/2022PD

Case External Checked:

NO

Case Internal Checked: