

Service Repair Sheet SRS68365

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Purchasing

Medisafe Ltd

01962 885777

01962 885777

00011608

richard@medisafe.org.uk

11/Jan/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC

Date: 11/01/23

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 09/Jan/2023 Janine Gill

09/Jan/2023 Janine Gill

Emailed for serial number

09/Jan/2023 Janine Gill

SN - 2586A14 V1000 calibration. Email sent to the customer with SRS number.

11/Jan/2023 Robert Connor

Received 1 x V1000 s/n PR02586A14, with 4 x AA battery and grey silicone cover.

Ready For quote

CGreen 13-1-23

Repair Complete Signed

CGreen 16-1-23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35609	Foetal Simulator	1410000	PR02586A14	2

1480000 x 1 @ £45

S/N, SRS, SRN

1430309 x 1 FOC

SRS, SRN

UPS x 1 @ £12.

PURCHASE ORDER

TO:

Viamed Limited
15 Station Road
Cross Hills Keighley
West Yorkshire
BD20 7DT

Tel: 01535 634542

Fax: 01535 635582

Please deliver to:

Medisafe Ltd
Medisafe
Unit 7 Ringtower Centre
Moorside Road, Winnall
Winchester
Hampshire
SO23 7RZ

Purchase Order Number: 5587

Purchase Order Date: 10 Jan 2023

Customer Order No: SRS68365

QUANTITY	MODEL NO	DESCRIPTION	PART NO.	COST	SUBTOTAL
		V1000 Foetal heart simulator for calibration SN 100879			
				SUBTOTAL:	£0.00
				CARRIAGE:	£0.00
				VAT @ 20.0%	£0.00
				TOTAL:	£0.00

Medisafe offers more than just an annual service visit

1 Unit 7 • Ringtower Centre • Moorside Road • Winchester • SO23 7RZ
Tel: 01962 885777 • Fax: 01962 885678 • Email: service@medisafe.org.uk • Web: www.medisafe.org.uk

MediSafe Ltd. Registered in England No. 03491833



PURCHASE ORDER

TO:

Viamed Limited
15 Station Road
Cross Hills Keighley
West Yorkshire
BD20 7DT

Tel: 01535 634542

Fax: 01535 635582

Please deliver to:

Medisafe Ltd
Medisafe
Unit 7 Ringtower Centre
Moorside Road, Winnall
Winchester
Hampshire
SO23 7RZ

Purchase Order Number: **5587**

Purchase Order Date: **10 Jan 2023**

Customer Order No: **SRS68365**

QUANTITY	MODEL NO	DESCRIPTION	PART NO.	COST	SUBTOTAL
1		V1000 Foetal heart simulator for calibration SN 100879		£45.00	£45.00
1	CARR	Carriage	P01000	£12.00	£12.00
				SUBTOTAL:	£57.00
				CARRIAGE:	£0.00
				VAT @ 20.0%	£11.40
				TOTAL:	£68.40

Medisafe offers more than just an annual service visit

1

1 Unit 7 • Ringtower Centre • Moorside Road • Winchester • SO23 7RZ

Tel.: 01962 885777 • Fax: 01962 885678 • Email: service@medisafe.org.uk • Web: www.medisafe.org.uk



Service Repair Sheet 68365

Contact Name	Purchasing
Company/ Hospital Name	Medisafe Ltd
Direct Phone	01962 885777
General Phone	01962 885777
Opera Account	00011608
Email	richard@medisafe.org.uk
Order Number	5587
Date Received	11/Jan/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35609	1410000	PR02586A14	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked
