






Hypertronics Male 9 Pin

Finger Probe



Hypertronics Male 9 Pin


Finger Probe




PI - 1.180 x 0.015 x 0.015 PUSHER
 * WARP SPEED AND BUMP SPEEDS TO BE TENSILE TOGETHER.

Hypertronics Male 9 Pin

Finger Probe




Enlarged front view



ø6 x 43 Clear
heatshrink. For
serial no.

Hypertronics male 9 pin



R1 - 1.58 K OHM 1%, 1/8W RESISTOR

* INNER SHIELD AND OUTER SHIELD TO BE TWISTED TOGETHER.

Title P866RA
Simed
Baxter



Dim in mm

Tol ± 0.2

Drawn J.Nirwan

Scale Not To Scale

Date 17/02/03


Dwg No. SPF-866.1

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT


REV	Date	N°	Drawn	Approv	Material :	Part No. 0018660
-----	------	----	-------	--------	------------	------------------

Hypertronics Male 9 Pin

Finger Probe





Enlarged front view



Pin Housing Rear View

{3 & 4} R1 = 1.58k Ω



Title: P866RA
Simed Baxter



Dim in	mm
Tol	± 0.2
Drawn	J.Nirwan

Scale: Not To Scale

Date: 30/01/02

Dwg No.: SPF-866



VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

Material:

Part No. 0018660

REV	Date	N°	Drawn	Approv
-----	------	----	-------	--------