




Hypertronics Male 9 Pin

Y- Probe



Enlarged front view

Hyptrtronics male 9 pin



R1 - 1.58 K OHM 1%, 1/8W RESISTOR

* INNER SHIELD AND OUTER SHIELD TO BE TWISTED TOGETHER.

Title P866YA
Simed Baxter



Dim in mm

Tol ± 0.2

Drawn J.Nirwan




Scale Not To Scale
Date 13/05/02
Dwg No. SPY-860.1

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	Nº	Drawn	Approv	Material :
					Part No. 0018662

Hypertronics Male 9 Pin

Y- Probe



REV	Date	N*	Drawn	Approv	Material :
					Part No. 0018662


Title P866YA Simed Baxter			
Scale	Not To Scale	Dim in	mm
Date	04/11/03	Tol	± 0.2
		Drawn	K.J. Rush
		Dwg No.	SPY-866



VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT


Hypertronics Male 9 Pin

Y- Probe



Pin Housing Rear View

(3 & 4) R1 = 1.98 k



Title P866YA
Simed
Baxter



Dim in mm

Tol ± 0.2

Drawn J.Nirwan

VIAMED

Ltd.


15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	N°	Drawn	Approv	Material :
					Part No. 0018662


Scale	Not To Scale	Date	13/05/02	Dwg No.	SPY-860
-------	--------------	------	----------	---------	---------

Hypertronics Male 9 Pin


Y- Probe



Enlarged front view



Hypertronics male 9 pin



R1 - 1.58 K OHM 1%, 1/8W RESISTOR

* INNER SHIELD AND OUTER SHIELD TO BE TWISTED TOGETHER.



Title P866YA				Dim in	
Simed				mm	
Baxter				Tol	
Scale Not To Scale				± 0.2	
Date 13/05/02				Drawn J.Nirwan	
Dwg No. SPY-860.1					

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	N°	Drawn	Approv	Material :
					Part No. 0018662

Hypertronics Male 9 Pin

Y- Probe




Hypertronics Male 9 Pin

Y- Probe



Hypertronics male 9 pin



P1 = 1.36 k (5 & 4) P1 = 1.36 k

* SHIELD SHIELD AND SHIELD SHIELD TO BE TAPPED TOGETHER