




P1 - 7.5 x 0.44 x 1.0 mm MICRODIP
* WIRE SHIELD AND CABLE SHIELD TO BE TIGHTLY TIGHTENED



D-Type Male 9 Pin

Y- Probe



Title P871YA
Nonin



Dim in	mm
Tol	± 0.2
Drawn	J.Nirwan





VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	N°	Drawn	Approv	Material:	Part No. 0018712
-----	------	----	-------	--------	-----------	------------------

Scale	Not To Scale	Date	13/05/02	Dwg No.	SPY-871
-------	--------------	------	----------	---------	---------


D-Type Male 9 Pin

Y- Probe



REV	Date	N*	Drawn	Approv	Material :
					Part No. 0018712


Title P871YA Nonin					VIAMED 15 Station Rd Cross Hills, Keighley West Yorkshire BD20 7DT	Ltd.
		Dim in	mm			
Scale Not To Scale		Tol	± 0.2			
Date 04/11/03		Drawn	K.J. Rush			
Dwg No.		SPY-871				




Title P871YA Nonin				VIAMED Ltd. 15 Station Rd Cross Hills, Keighley West Yorkshire BD20 7DT	
Scale		Not To Scale		Drawn J.Nirwan	
Date 13/05/02		Dwg No.		SPY-871	

D-Type Male 9 Pin

Y- Probe



D-type male 9 pin



R1 - 7.5 K OHM 1%, 1/8W RESISTOR

* INNER SHIELD AND OUTER SHIELD TO BE TWISTED TOGETHER.

Title		P871YA				VIAMED Ltd. 15 Station Rd Cross Hills, Keighley West Yorkshire BD20 7DT					
Nonin		Dim in								mm	
		Tol								± 0.2	
Scale		Not To Scale		Drawn						J.Nirwan	
Date		13/05/02		Dwg No.		SPY-871.1					