

Service Repair Sheet SRS68368

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Mike Brampton
Thames Medical Ltd
Accounts

01903 522911
01903 522911
00011982
mike@thamesmedical.com
PO-01871
25/Jan/2023
Robert Connor
Viamed
Warranty

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer



Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 11/Jan/2023 Steve Hardaker

11/Jan/2023 Steve Hardaker

Returning a VM2500M - Mainstream Vet Capnograph SN-A2106150004 with IRMA SN-IR00007674.

The display has a lot of missing pixels.

Sold 3/12/21 - RVM133737-1, still under warranty.

25/Jan/2023 Robert Connor

Received 1 x VM-2500-M-Vet s/n A2106150004, with silicone cover and universal mounting plate. No IRMA included.

Ready For quote

Repair Complete Signed

C Green 1.2.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35622	Capnograph	4410530	A2106150004	Yes

4440000 x 1 FOC

S/N, SRS, SRN

4430120 x 1 FOC

SRS, SRN

UPS x 1 FOC

Thames Medical Ltd

8 A'Becket Gardens

Worthing, West Sussex BN13 2BW

www.thamesmedical.com

email info@thamesmedical.com

tel 01903 522911

fax 01903 522922

Experts in patient monitoring and anaesthesia

**PURCHASE ORDER**

order #

PO-01871

date

20/01/2023

vendor address **Viamed Ltd**
15 Station Road
Cross Hills
Keighley, West Yorkshire
BD20 7DT

contact Kate
phone 01535 634542
fax 01535 635582

carrier
UPS

payment terms
Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SPECIAL TAX RATE	SUB-TOTAL
TML-V-VM2500 M1-RL2WR	VM2500M - Warranty Repair - Service Level 2 SN-A2106150004	Warranty Repair of VM2500M Mainstream Capnograph	1	£0.00	Taxable	£0.00
remarks	SRS68368				SUB-TOTAL	£0.0
	FAO Steve Hardaker				VAT	£0.0
	Pixels faulty on screen				TOTAL	£0.0

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:
info@thamesmedical.com

Service Repair Sheet 68368

Contact Name	Mike Brampton
Company/ Hospital Name	Thames Medical Ltd
Department	Accounts
Direct Phone	01903 522911
General Phone	01903 522911
Opera Account	00011982
Email	mike@thamesmedical.com
Order Number	PO-01871
Date Received	25/Jan/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN35622	4410530	A2106150004	Capnograph	Complete - Repaired
				Time :0 Hour(s)
Parts Replaced				
		Qty: 1	4430120	

The Capnograph VM 2500-M Vet has been repaired. The OLED display has been replaced.
