

Service Repair Sheet SRS68370

Contact Name Steve Lawrence
 Company/ Hospital Name St Bartholomews Hospital
 Department
 Position
 Direct Phone 02034656417
 General Phone 020 7480 4622
 Opera Account 00003204
 Email steve.lawrence2@nhs.net
 Order Number
 Date Received 23/Jan/2023
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Quote

VIAMEDclear

Goods In Only
 Decontamination
 certificate provided
 by customer

Cleaned by Viamed,
 if no declaration certificate
 from customer
 Signed: _____
 Date: _____

Goods Out Only
 Cleaned by Viamed before
 returning to customer
 Signed: _____
 Date: _____

Notes 18/Jan/2023 Janine Gill
 18/Jan/2023 Janine Gill
 Enquired about calibration of a V1000, email sent with SRS Number and instructions.
 23/Jan/2023 Robert Connor
 Received 1 x V1000 s/n PR02940A13, with 4 x AA battery and grey plastic carry case.

Ready For quote

CGreen 25.1.23

Repair Complete Signed

CGreen 3.2.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35621	Foetal Simulator	1410000	PR02940A13	N

1480000 x 1 £45

S/N, SRS, SRN

1430309 x 1 £0

SRS, SRN

UPS x 1 £12

Barts Health NHS Trust

OFFICIAL ORDER

Order Date : 01-02-2023

Order No : **40843887**

Must be quoted on all correspondence.

Deliver To:
CLINICAL EQUIP (C E M S) BNB-B1-205
All Deliveries to: ST BARTS STORES FM YARD
Bnt Level B1, KGV Wing
West Smithfield
GTL
EC1A 7BE
GB
Requested delivery date: 07-02-2023
Location ID: AN0038 CLINICAL EQUIP (C E M S)
BNB-B1-263

Invoice and Payment Enquiries To
TREASURY & PAYMENTS DEPARTMENT
8TH FLOOR, 20 CHURCHILL PLACE
EORI: GB654 9470 05 000
LONDON
E14 5HJ
GB
Fax: apinvoices.bartshealth@nhs.net

Warning : Order has at least one service line. Please process manually.
Buyer Contact : Begum, Khadijah
Buyer Email : barts.procurement@nhs.net

All enquiries regarding this order to:
Contact : Lawrence, Steve
Telephone :
Facsimile No. :
Email Address : steve.lawrence2@nhs.net

Warning : Order has at least one service line. Please process manually.
Buyer Contact : Begum, Khadijah
Buyer Email : barts.procurement@nhs.net

Supplier

Viamed Ltd

Customer's Supplier Name:
VIAMED

Conditions

THIS ORDER IS SUBJECT TO STANDARD NHS TERMS AND CONDITIONS. IF PRICES STATED ON THIS ORDER ARE INCORRECT ANY REVISED PRICES MUST BE AUTHORISED BY THE BUYER PRIOR TO ORDER EXECUTION. PAYMENT WILL BE MADE AT THE PRICES STATED HEREIN. DO NOT ASSIGN THIS ORDER SPECIAL INSTRUCTIONS.

Line	Goods or Services Required	Quantity	UOM	Contract Ref.	Unit Price	Line Value	VAT
1	N/A Please Calibrate V1000 Foetal heart simulator S/N: PR02940A13. A/N: 118550.	57	EA		£1.00	£57.00	-

Net Total : £57.00
Carriage : -
Tax : -
Total : £57.00

Service Repair Sheet 68370

Contact Name	Steve Lawrence
Company/ Hospital Name	St Bartholomews Hospital
Direct Phone	02034656417
General Phone	020 7480 4622
Opera Account	00003204
Email	steve.lawrence2@nhs.net
Order Number	40843887
Date Received	23/Jan/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN35621	1410000	PR02940A13	Foetal Simulator	Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
