



New Information Data Sheet



Date	27 11 98	Technician	DB
Original Manufacturer	BETA BIOMED	Original part Number	?
Work requested by	JSL	Length	4' (TIP TO TIP)
Description	MELLOX EQUIV.	Time taken mins	60
New part number if required	—	Photographs	Yes No

Connector manufacturer AMP	Observations on the cable:-				manufacturer	
View SOLDER SIDE					View SOLDER SIDE	
Connector Diagram	Colour	Pins	Cross wiring diagram	Pin	Colour	Diagram
	RES.	1		1	BLACK	
	RED + GREEN	2		2	WHITE	
	YELLOW + ORANGE	3		3		
	NOT USED	4		4	GREEN	
	WHITE	5		5	ORANGE	
	RES.	6		6	YELLOW	
	BOTH SHIELDS	7		7	RED	
	NOT USED	8		8		
	BLACK	9		9		
		10		10		
		11		11		
		12		12		
		13		13		
		14		14		
		15		15		
		16		16		
Electronic components		Pin	Are there any changes from Viamed records	Pin		Electronic components
		②				
		③				
		⑤				
		⑨				
		①				
		⑥				


This diagram has been validated by :

Date


Cable works on original equipment	Cable diagram correct	Components correct	Colours correct
-----------------------------------	-----------------------	--------------------	-----------------

BETA BIOMED DATASCOPE 8 PIN DIN


SEN RV



LED RV



CONN RV



$$R1 = 47K\Omega$$

$$R2 = 22\Omega$$