

Service Repair Sheet SRS68398

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Artwell Bere

Blackpool Victoria Hospital

Medical Engineering

Medical Engineer

01253957799

01253 300000

00000570

artwell.bere@bfwml.co.uk

21/Feb/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer



Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 21/Feb/2023 Robert Connor

21/Feb/2023 Robert Connor

Received 1 x V1000 s/n PR02676A12, with 4 x AA battery and hard plastic carry case.

Ready For quote

CCGreen 21.2.23

Repair Complete Signed

CCGreen 28.2.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35679	Foetal Simulator	1410000	PR02676A12	2

14800000 x 1 @ £45

S/N, SRS, SRN

1430309 x 1 FOC

SRS, SRN

UPS x 1 @ £12.

Purchase Order No. RXLA400023080		Rev No. 0	Page 1 of 1	
Date of Order 23-FEB-2023		Revision Date		

Supplier: /amed Ltd 15 Station Road Cross Hills Keighley 3D20 7DT Tel:		Deliver To: VICTORIA HOSPITAL STORES WHINNEY HEYS ROAD BLACKPOOL FY3 8NR United Kingdom	Invoice To: Atlas BFW Management Limited Victoria Hospital, Accounts Payable Services Home 7, Whinney Heys Road Blackpool FY3 8NR United Kingdom Tel: 5055207387886 Email: bfwmpayables@bfwml.co.uk	Enquiries To: Eden Stafford WORKS STORES (BVH) BLACKPOOL VICTORIA HOSPITAL WHINNEY HEYS ROAD BLACKPOOL FY3 8NR Tel: Email: Eden.Stafford@bfwml.co.uk
---	--	---	---	---

Important Information:

1. This order is issued in accordance with the Terms and Conditions of Contracts from www.dh.gov.uk/en/Publicationsandstatistics/Publications/PublicationsPolicyAndGuidance/DH_1212602. DELIVERY NOTES to accompany all deliveries of goods.3. NO VARIATION to this order without written authority. Any alteration in quantity or price must be agreed in writing by the ordering officer before any goods are supplied.4. The Goods/Services Are Ordered On 2020 Incoterm DDP (Delivered Duty Paid) Basis With Delivery Required To Blackpool Victoria Hospital Whinney Heys Rd, Blackpool FY3 8NR, United Kingdom.If An Alternate Location Is Required, This Is Shown In The Delivery Text. EORI No. GB2647418840005. COSHH 2002 (Amended 2003) REGS: If any of the items detailed on this order could be hazardous to health the supplier must provide a detailed Product Safety Data Sheet.6. Any invoice received incorrectly addressed and/or without a valid trust purchase order number will be rejected.

Line No.	Product Code	Description of Goods or Services	Qty	Unit of Measure	Unit Price	Line Total	Deliver by Date	Contract/Quote Reference
1	SRS68398	Service and Calibration of Viamed Foetal Heart Simulator V1000, Serial Number: 100857 Note: artwell.bere@bfwml.co.uk				57.00	02/03/2023	
						Total GBP:	57.00	

Service Repair Sheet 68398

Contact Name	Artwell Bere
Company/ Hospital Name	Blackpool Victoria Hospital
Department	Medical Engineering
Position	Medical Engineer
Direct Phone	01253957799
General Phone	01253 300000
Opera Account	00000570
Email	artwell.bere@bfwml.co.uk
Order Number	RXLA400023080
Date Received	21/Feb/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35679	1410000	PR02676A12	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked
