



New Information Data Sheet



Extension Cables

Date	6-10-98	Technician	APB
Original Manufacturer	SPACELABS	Original part Number	
Work requested by	EPIC DEPT.	Length	
Description	E.T. CABLE	Time taken mins	
New part number if required		Photographs	Yes No

Connector manufacturer	Observations on the cable:- A BLUE WIRE IS ATTACHED TO FROM THE SHIELD TO PIN 3 OF THE HYPERBOLIC END, MAKING THE CONNECTION FROM THE SHIELD TO THE PIN 3.						Connector manufacturer
View <u>REAR SOLDER</u>							View <u>REAR SOLDER</u>
Connector Diagram	Colour	Pins	Cross wiring diagram	Pin	Colour	Connector Diagram	
	RED	1		1	BLACK		
	PURPLE	2		2	BROWN		
	GREEN	3		3	SHIELD		
	N.C	4		4	N.C		
	BLACK	5		5	N.C		
	SHIELD	6		6	N.C		
		7		7			
	N.C	8		8	RED		
	BROWN	9		9	N.C		
		10		10	N.C		
		11		11	GREEN		
		12		12	PURPLE		
		13		13			
		14		14			
		15		15			
		16		16			
Electronic components		Pin	Are there any changes from Viamed records	Pin		Electronic components	
<u>LINK</u>		6		2		10-2 RESISTOR	
		7		3			61K RESISTOR
				7			

This diagram has been validated by :

Date

Cable works on original equipment	Cable diagram correct	Components correct	Colours correct
-----------------------------------	-----------------------	--------------------	-----------------