

Service Repair Sheet SRS68389

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Aneta Szubarga

Euromed Medical Solution

0048618439059

504 078 773

00006582

a.szubarga@euromed.net.pl

21/Feb/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RS
Date: 21/02/23

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 13/Feb/2023 Kate Griffiths

13/Feb/2023 Kate Griffiths

V1000, s/n PR02724A13 - they are ``sending it in for inspection tomorrow``

21/Feb/2023 Robert Connor

Received 1 x V1000 s/n PR02724A13, with 4 x AA battery and hard plastic carry case.

Ready For quote

CGreen 21.2.23

Repair Complete Signed

CGreen 22.3.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35680	Foetal Simulator	1410000	PR02724A13	N

480000 x 1 @ €44.20

SIN, SRS, SRN

1430309 x 1 @ €0

SRS, SRN

BINS x 1

UPS x 1

32x26x12cm 0.90kg

€ DIS.

Service Repair Sheet 68389

Contact Name	Aneta Szubarga
Company/ Hospital Name	Euromed Medical Solution
Direct Phone	0048618439059
General Phone	504 078 773
Opera Account	00006582
Email	a.szubarga@euromed.net.pl
Date Received	21/Feb/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN35680	1410000	PR02724A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked
