



OPTIONAL APPROXIMATE FORM
 1.000 2.000 3.000 4.000 5.000 6.000 7.000 8.000 9.000 10.000



LSR, SHORE 40 A,
 MEDICAL GRADE, WHITE
 REMOVE ALL
 FLASHES


- A. REMOVE ALL BURRS
- B. MACHINE SURFACES AS
- C. BREAK ALL SHARP EDGES .010 MAX
- D. DIMENSIONS PER ANSI Y14.5
- E. DIMENSIONS ARE IN INCHES
- F. DO NOT SCALE DRAWINGS

G. TOLERANCES ON
 DECIMALS
 .XX 2X TOLERANCE
 .XXX 3X TOLERANCE
 ANGLES
 ANGLE TOLERANCE


VIAMED Ltd

TITLE		SOFT FINGER PAD (LED) VPA006	
DWG NO.		VPA006/a	
SCALE:		SIZE: B	SHEET 1 OF 1

NOTES: UNLESS OTHERWISE SPECIFIED.



SECTION VIEW A-A



SECTION VIEW B-B



NOTES: UNLESS OTHERWISE SPECIFIED.

Form ENG023 (04/93)



FINISH		VIAMED Ltd	
MATERIAL			
A. REMOVE ALL BURRS B. MACHINE SURFACES GS C. BREAK ALL SHARP EDGES .010 MAX D. DIMENSIONS PER ANSI Y14.5 E. DIMENSIONS ARE IN INCHES F. DO NOT SCALE DRAWING		TITLE SOFT FINGER PAD (LED)	
		DWG NO. VPA006_B	
		REV 1	
G. TOLERANCES ON DECIMALS .XX 2X TOLERANCE .XXX 3X TOLERANCE ANGLES ANGLE TOLERANCE		SCALE:	SIZE: B
		SHEET	SHEET OF 1

DOS FILE NAME: DOS FILE NAME

Section View AA



Section View BB



Sheet 2 Of 2

REV	Date	Details	Approv	Part No. VPA002/a
				Material: LSR, Shore 40 A
				Medial Grade, White

Title Top Pad (Emitter)			
Scale	Not To Scale	Dim in	Inches
Date	12/10/03	Tol	± 0.01
Dwg No.	FPD-003	Drawn	K.J. Rush

VIAMED Ltd.

15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT