

DATE	I.D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
27-Feb-98	2440	BCI	98	60	160	60	98	60	160	60
27-Feb-98	2441	BCI	98	60	160	60	97	60	160	60
27-Feb-98	2442	BCI	98	61	160	60	98	60	160	60
27-Feb-98	2443	BCI	98	60	159	61	99	61	159	60
27-Feb-98	2444	BCI	97	60	160	60	98	60	160	60
27-Feb-98	2445	BCI	98	60	160	60	98	61	160	60
27-Feb-98	2446	BCI	99	60	160	61	98	59	160	60
27-Feb-98	2447	CSI	97	59	160	59	97	59	160	59
27-Feb-98	2448	CSI	98	61	160	60	98	59	160	60
27-Feb-98	2449	CSI	98	60	160	60	98	59	160	61
27-Feb-98	2450	CSI	98	60	160	60	98	60	160	60
02-Mar-98	2451	CSI	98	60	160	60	97	60	160	60
02-Mar-98	2452	CSI	98	61	160	60	98	60	160	60
02-Mar-98	2453	CSI	99	60	160	60	98	60	160	59
02-Mar-98	2454	CSI	98	60	160	60	98	59	160	60
02-Mar-98	2455	CSI	97	60	160	60	98	60	160	60
02-Mar-98	2456	CSI	99	61	160	60	98	59	160	59
02-Mar-98	2457	CSI	99	61	160	59	99	60	159	60
02-Mar-98	2458	CSI	98	62	159	60	98	60	160	61
02-Mar-98	2459	CSI	98	59	160	61	98	61	160	60
02-Mar-98	2460	CSI	97	60	160	60	98	61	160	61
02-Mar-98	2461	CSI	98	62	160	60	97	60	159	60
03-Mar-98	2462	OHMEDA	98	60	161	60	98	60	159	59
03-Mar-98	2463	OHMEDA	97	61	159	59	97	59	160	60
03-Mar-98	2464	OHMEDA	98	59	160	60	98	60	160	60
03-Mar-98	2465	OHMEDA	99	60	160	60	98	58	160	59
03-Mar-98	2466	OHMEDA	98	61	160	60	98	59	161	60
04-Mar-98	2467	OHMEDA	98	59	161	60	98	59	160	60
04-Mar-98	2468	OHMEDA	98	60	160	60	98	60	160	60
04-Mar-98	2469	OHMEDA	99	60	160	60	99	60	160	60
04-Mar-98	2470	OHMEDA	98	61	161	60	98	59	160	59
04-Mar-98	2471	OHMEDA	98	60	160	60	98	60	160	60
04-Mar-98	2472	OHMEDA	99	60	160	61	98	59	160	60
04-Mar-98	2473	OHMEDA	97	59	160	59	97	59	160	59

CROSS COMPARISON VIAMED DL3000 vs BIO-TEK INDEX

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
05-Jan-98	2005	NELLCOR	98	61	160	60	98	60	160	60
05-Jan-98	2006	NELLCOR	98	60	159	61	99	61	159	60
05-Jan-98	2007	NELLCOR	97	60	160	60	98	60	160	60
05-Jan-98	2008	NELLCOR	98	60	160	60	98	61	160	60
05-Jan-98	2009	NELLCOR	99	60	160	61	98	59	160	60
05-Jan-98	2010	NELLCOR	97	59	160	59	97	59	160	59
05-Jan-98	2011	NELLCOR	98	61	160	60	98	59	160	60
05-Jan-98	2012	NELLCOR	98	60	160	60	98	59	160	61
05-Jan-98	2013	NELLCOR	98	60	160	60	98	60	160	60
05-Jan-98	2014	NELLCOR	98	60	160	60	97	60	160	60
05-Jan-98	2015	NELLCOR	98	61	160	60	98	60	160	60
07-Jan-98	2016	NELLCOR	99	60	160	60	98	60	160	59
07-Jan-98	2017	NELLCOR	98	60	160	60	98	59	160	60
07-Jan-98	2018	NELLCOR	97	60	160	60	98	60	160	60
07-Jan-98	2019	NELLCOR	99	61	160	60	98	59	160	59
07-Jan-98	2020	OHMEDA	99	61	160	59	99	60	159	60
07-Jan-98	2021	OHMEDA	98	62	159	60	98	60	160	61
08-Jan-98	2022	OHMEDA	98	59	160	61	98	61	160	60
08-Jan-98	2023	OHMEDA	97	60	160	60	98	61	160	61
08-Jan-98	2024	OHMEDA	98	62	160	60	97	60	159	60
08-Jan-98	2025	OHMEDA	98	60	161	60	98	60	159	59
08-Jan-98	2026	OHMEDA	97	61	159	59	97	59	160	60
12-Jan-98	2027	DATASCOPE	98	59	160	60	98	60	160	60
12-Jan-98	2028	DATASCOPE	99	60	160	60	98	58	160	59
12-Jan-98	2029	DATASCOPE	98	61	160	60	98	59	161	60
12-Jan-98	2030	DATASCOPE	98	59	161	60	98	59	160	60
12-Jan-98	2031	DATASCOPE	98	60	160	60	98	60	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
14-Jan-98	2066	OHMEDA	99	60	160	61	98	59	160	60
14-Jan-98	2067	CSI	97	59	160	59	97	59	160	59
14-Jan-98	2068	CSI	98	61	160	60	98	59	160	60
14-Jan-98	2069	CSI	98	60	160	60	98	59	160	61
14-Jan-98	2070	CSI	98	60	160	60	98	60	160	60
14-Jan-98	2071	CSI	98	60	160	60	97	60	160	60
14-Jan-98	2072	DATASCOPE	98	61	160	60	98	60	160	60
14-Jan-98	2073	DATASCOPE	99	60	160	60	98	60	160	59
14-Jan-98	2074	BCI	98	60	160	60	98	59	160	60
14-Jan-98	2075	DATASCOPE	97	60	160	60	98	60	160	60
15-Jan-98	2076	NOVAMETRIX	99	61	160	60	98	59	160	59
15-Jan-98	2077	CSI	99	61	160	59	99	60	159	60
15-Jan-98	2078	NELLCOR	98	62	159	60	98	60	160	61
15-Jan-98	2079	CSI	98	59	160	61	98	61	160	60
15-Jan-98	2080	DATASCOPE	97	60	160	60	98	61	160	61
15-Jan-98	2081	NOVAMETRIX	98	62	160	60	97	60	159	60
15-Jan-98	2082	OHMEDA	98	60	161	60	98	60	159	59
15-Jan-98	2083	OHMEDA	97	61	159	59	97	59	160	60
15-Jan-98	2084	OHMEDA	98	59	160	60	98	60	160	60
16-Jan-98	2085	OHMEDA	99	60	160	60	98	58	160	59
16-Jan-98	2086	OHMEDA	98	61	160	60	98	59	161	60
16-Jan-98	2087	NELLCOR	98	59	161	60	98	59	160	60
16-Jan-98	2088	NELLCOR	98	60	160	60	98	60	160	60
16-Jan-98	2089	NELLCOR	99	60	160	60	99	60	160	60
16-Jan-98	2090	NELLCOR	98	61	161	60	98	59	160	59
16-Jan-98	2091	NELLCOR	98	60	160	60	98	60	160	60
16-Jan-98	2092	NELLCOR	99	60	160	61	98	59	160	60
16-Jan-98	2093	NELLCOR	97	59	160	59	97	59	160	59
16-Jan-98	2094	NELLCOR	98	61	160	60	98	59	160	60
16-Jan-98	2095	NELLCOR	98	60	160	60	98	59	160	61
16-Jan-98	2096	NELLCOR	98	60	160	60	98	60	160	60
16-Jan-98	2097	NELLCOR	98	60	160	60	97	60	160	60
16-Jan-98	2098	NELLCOR	98	61	160	60	98	60	160	60
19-Jan-98	2099	NELLCOR	99	60	160	60	98	60	160	59

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
19-Jan-98	2100	INIVO	98	60	160	60	98	59	160	60
19-Jan-98	2101	OHMEDA	97	60	160	60	98	60	160	60
19-Jan-98	2102	OHMEDA	99	61	160	60	98	59	160	59
19-Jan-98	2103	OHMEDA	99	61	160	59	99	60	159	60
19-Jan-98	2104	OHMEDA	98	62	159	60	98	60	160	61
19-Jan-98	2105	OHMEDA	98	59	160	61	98	61	160	60
19-Jan-98	2106	BCI	97	60	160	60	98	61	160	61
19-Jan-98	2107	BCI	97	60	160	60	98	60	160	60
20-Jan-98	2108	SENSORMEDIC	98	60	160	60	98	60	160	60
20-Jan-98	2109	NONIN	98	60	160	60	98	60	160	60
20-Jan-98	2110	NOVAMETRIX	98	60	159	60	98	59	160	60
20-Jan-98	2111	NELCOR	99	60	160	59	98	60	160	60
20-Jan-98	2112	NELCOR	98	61	160	60	98	59	160	60
20-Jan-98	2113	NELCOR	98	60	160	60	98	59	160	61
20-Jan-98	2114	NELCOR	98	60	160	60	98	60	160	60
20-Jan-98	2115	NELCOR	98	60	160	60	97	60	160	60
20-Jan-98	2116	NELCOR	98	61	160	60	98	60	160	60
20-Jan-98	2117	NELCOR	99	60	160	60	98	60	160	59
20-Jan-98	2118	NELCOR	98	60	160	60	98	59	160	60
20-Jan-98	2119	NOVAMETRIX	97	60	160	60	98	60	160	60
21-Jan-98	2120	NOVAMETRIX	99	61	160	60	98	59	160	59
21-Jan-98	2121	NOVAMETRIX	98	61	160	59	99	60	159	60
21-Jan-98	2122	NOVAMETRIX	98	62	159	60	98	60	160	61
21-Jan-98	2123	NOVAMETRIX	98	59	160	61	98	61	160	60
21-Jan-98	2124	NOVAMETRIX	98	60	160	60	97	61	160	60
21-Jan-98	2125	NELCOR	98	60	160	60	97	60	160	60
21-Jan-98	2126	NELCOR	98	60	160	60	98	60	160	60
22-Jan-98	2127	NELCOR	98	60	160	60	98	60	160	60
22-Jan-98	2128	NELCOR	98	60	159	60	98	59	160	60
22-Jan-98	2129	NELCOR	98	60	160	60	97	60	160	60
22-Jan-98	2130	NELCOR	98	60	160	60	98	60	160	60
22-Jan-98	2131	NELCOR	98	60	160	60	97	60	160	60
22-Jan-98	2132	NELCOR	98	61	160	60	98	60	160	60
22-Jan-98	2133	NELCOR	99	60	160	60	98	60	160	59

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
22-Jan-98	2134	NELLCOR	98	60	160	60	98	59	160	60
22-Jan-98	2135	NELLCOR	97	60	160	60	98	60	160	60
22-Jan-98	2136	NELLCOR	99	61	160	60	98	59	160	59
22-Jan-98	2137	OHMEDA	98	61	160	59	99	60	159	60
22-Jan-98	2138	OHMEDA	98	62	159	60	98	60	160	61
22-Jan-98	2139	OHMEDA	98	59	160	61	98	61	160	60
22-Jan-98	2140	OHMEDA	98	60	160	60	97	61	160	60
22-Jan-98	2141	OHMEDA	98	60	160	60	97	60	160	60
22-Jan-98	2142	NOVAMETRIX	98	60	160	60	98	60	160	60
22-Jan-98	2143	NOVAMETRIX	98	60	160	60	98	60	160	60
22-Jan-98	2144	CSI	98	60	159	60	98	59	160	60
22-Jan-98	2145	CSI	98	60	160	60	97	60	160	60
22-Jan-98	2146	CSI	98	60	160	60	98	60	160	60
22-Jan-98	2147	DATASCOPE	98	60	159	60	98	59	160	60
22-Jan-98	2148	DATASCOPE	98	60	160	60	97	60	160	60
23-Jan-98	2149	OHMEDA	98	60	160	60	98	60	160	60
23-Jan-98	2150	NELLCOR	98	61	160	60	98	60	160	60
23-Jan-98	2151	NELLCOR	99	60	160	60	98	60	160	59
26-Jan-98	2152	NELLCOR	98	60	160	60	98	59	160	60
26-Jan-98	2153	NELLCOR	97	60	160	60	98	60	160	60
26-Jan-98	2154	NELLCOR	99	61	160	60	98	59	160	59
26-Jan-98	2155	NELLCOR	98	61	160	59	99	60	159	60
26-Jan-98	2156	NELLCOR	98	62	159	60	98	60	160	61
26-Jan-98	2157	NELLCOR	98	59	160	61	98	61	160	60
26-Jan-98	2158	NELLCOR	98	60	160	60	97	61	160	60
26-Jan-98	2159	NELLCOR	98	60	160	60	97	60	160	60
26-Jan-98	2160	NELLCOR	98	60	160	60	98	60	160	60
26-Jan-98	2161	NELLCOR	98	60	160	60	98	60	160	60
26-Jan-98	2162	NELLCOR	98	60	159	60	98	59	160	60
26-Jan-98	2163	NELLCOR	98	60	160	60	97	60	160	60
26-Jan-98	2164	OHMEDA	98	60	160	60	98	60	160	60
27-Jan-98	2165	OHMEDA	98	61	160	60	98	60	160	60
27-Jan-98	2166	OHMEDA	99	60	160	60	98	60	160	59
27-Jan-98	2167	OHMEDA	98	60	160	60	98	59	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
27-Jan-98	2168	OHMEDA	97	60	160	60	98	60	160	60
27-Jan-98	2169	NOVAMETRIX	99	61	160	60	98	59	160	59
27-Jan-98	2170	NOVAMETRIX	98	61	160	59	99	60	159	60
27-Jan-98	2171	NOVAMETRIX	98	62	159	60	98	60	160	61
27-Jan-98	2172	NOVAMETRIX	98	59	160	61	98	61	180	60
27-Jan-98	2173	DATASCOPE	98	60	160	60	97	61	160	60
27-Jan-98	2174	DATASCOPE	98	60	160	60	97	60	160	60
27-Jan-98	2175	DATASCOPE	98	60	160	60	98	60	160	60
27-Jan-98	2176	DATASCOPE	98	60	160	60	98	60	160	60
27-Jan-98	2177	DATASCOPE	98	60	159	60	98	59	180	60
27-Jan-98	2178	DATASCOPE	98	60	160	60	97	60	160	60
27-Jan-98	2179	DATASCOPE	98	60	160	60	98	60	160	60
28-Jan-98	2180	DATASCOPE	98	61	160	60	98	60	160	60
28-Jan-98	2181	DATASCOPE	98	60	159	61	99	61	159	60
28-Jan-98	2182	CSI	97	60	160	60	98	60	160	60
28-Jan-98	2183	CSI	98	60	160	60	98	61	160	60
28-Jan-98	2184	CSI	99	60	160	61	98	59	160	60
28-Jan-98	2185	NELLCOR	97	59	160	59	97	59	160	59
28-Jan-98	2186	NELLCOR	98	61	160	60	98	59	160	60
28-Jan-98	2187	NELLCOR	98	60	160	60	98	59	160	61
28-Jan-98	2188	NELLCOR	98	60	160	60	98	60	160	60
28-Jan-98	2189	NELLCOR	98	60	160	60	97	60	160	60
28-Jan-98	2190	NELLCOR	98	61	160	60	98	60	160	60
28-Jan-98	2191	NELLCOR	99	60	160	60	98	60	160	59
28-Jan-98	2192	NELLCOR	98	60	160	60	98	59	160	60
28-Jan-98	2193	NELLCOR	97	60	160	60	98	60	160	60
28-Jan-98	2194	NELLCOR	99	61	160	60	98	59	160	59
28-Jan-98	2195	NELLCOR	99	61	160	59	99	60	159	60
28-Jan-98	2196	NELLCOR	98	62	159	60	98	60	160	61
28-Jan-98	2197	NELLCOR	98	59	160	61	98	61	160	60
29-Jan-98	2198	NELLCOR	97	60	160	60	98	61	160	61
29-Jan-98	2199	NELLCOR	98	62	160	60	97	60	159	60
29-Jan-98	2200	NELLCOR	98	60	161	60	98	60	159	59
29-Jan-98	2201	NELLCOR	97	61	159	59	97	59	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
28-Jan-98	2202	NELLCOR	98	59	160	60	98	60	160	60
29-Jan-98	2203	NELLCOR	99	60	160	60	98	58	160	59
29-Jan-98	2204	NELLCOR	98	61	160	60	98	59	161	60
30-Jan-98	2205	NELLCOR	98	59	161	60	98	59	160	60
30-Jan-98	2206	NELLCOR	98	60	160	60	98	60	160	60
30-Jan-98	2207	NELLCOR	99	60	160	60	99	60	160	60
30-Jan-98	2208	NELLCOR	98	61	161	60	98	59	160	59
30-Jan-98	2209	NELLCOR	98	60	160	60	98	60	160	60
30-Jan-98	2210	NELLCOR	99	60	160	61	98	59	160	60
30-Jan-98	2211	NELLCOR	97	59	160	59	97	59	160	59
30-Jan-98	2212	NELLCOR	98	61	160	60	98	59	160	60
30-Jan-98	2213	NELLCOR	98	60	160	60	98	59	160	61
30-Jan-98	2214	NELLCOR	98	60	160	60	98	60	160	60
30-Jan-98	2215	NELLCOR	98	60	160	60	97	60	160	60
02-Feb-98	2216	NELLCOR	98	61	160	60	98	60	160	60
02-Feb-98	2217	NELLCOR	99	60	160	60	98	60	160	59
02-Feb-98	2218	NELLCOR	98	60	160	60	98	59	160	60
02-Feb-98	2219	NELLCOR	97	60	160	60	98	60	160	60
02-Feb-98	2220	NELLCOR	99	61	160	60	98	59	160	59
02-Feb-98	2221	NELLCOR	99	61	160	59	99	60	159	60
02-Feb-98	2222	OHMEDA	98	62	159	60	98	60	160	61
02-Feb-98	2223	OHMEDA	98	59	160	61	98	61	160	60
02-Feb-98	2224	OHMEDA	97	60	160	60	98	61	160	61
02-Feb-98	2225	OHMEDA	97	60	160	60	98	60	160	60
02-Feb-98	2226	OHMEDA	98	60	160	60	98	60	160	60
03-Feb-98	2227	OHMEDA	98	60	160	60	98	60	160	60
03-Feb-98	2228	OHMEDA	98	60	159	60	98	59	160	60
03-Feb-98	2229	OHMEDA	99	60	160	59	98	60	160	60
03-Feb-98	2230	OHMEDA	98	61	160	60	98	59	160	60
03-Feb-98	2231	NELLCOR	98	60	160	60	98	59	160	61
04-Feb-98	2232	NELLCOR	98	60	160	60	98	60	160	60
04-Feb-98	2233	NELLCOR	98	60	160	60	97	60	160	60
04-Feb-98	2234	NELLCOR	98	61	160	60	98	60	160	60
04-Feb-98	2235	NELLCOR	99	60	160	60	98	60	160	59

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
04-Feb-98	2236	NELLCOR	98	60	160	60	98	59	160	60
04-Feb-98	2237	NELLCOR	98	61	160	60	98	60	160	60
04-Feb-98	2238	NELLCOR	98	60	159	61	99	61	159	60
04-Feb-98	2239	NOVAMETRIX	97	60	160	60	98	60	160	60
05-Feb-98	2240	NOVAMETRIX	98	60	160	60	98	61	160	60
05-Feb-98	2241	DATASCOPE	99	60	160	61	98	59	160	60
05-Feb-98	2242	DATASCOPE	97	59	160	59	97	59	160	59
05-Feb-98	2243	CSI	98	61	160	60	98	59	160	60
06-Feb-98	2244	BCI	98	60	160	60	98	59	160	61
06-Feb-98	2245	BCI	98	60	160	60	98	60	160	60
06-Feb-98	2246	OHMEDA	98	60	160	60	97	60	160	60
06-Feb-98	2247	OHMEDA	98	61	160	60	98	60	160	60
06-Feb-98	2248	OHMEDA	98	60	160	60	98	60	160	59
06-Feb-98	2249	OHMEDA	98	60	160	60	98	59	160	60
09-Feb-98	2250	OHMEDA	97	60	160	60	98	60	160	60
09-Feb-98	2251	NELLCOR	99	61	160	60	98	59	160	59
09-Feb-98	2252	NELLCOR	99	61	160	59	99	60	159	60
09-Feb-98	2253	NELLCOR	98	62	159	60	98	60	160	61
09-Feb-98	2254	NELLCOR	98	59	160	61	98	61	160	60
09-Feb-98	2255	NELLCOR	97	60	160	60	98	61	160	61
09-Feb-98	2256	NELLCOR	98	62	160	60	97	60	159	60
09-Feb-98	2257	NELLCOR	98	60	161	60	98	60	159	59
09-Feb-98	2258	NELLCOR	97	61	159	59	97	59	160	60
09-Feb-98	2259	NELLCOR	98	59	160	60	98	60	160	60
09-Feb-98	2260	NELLCOR	99	60	160	60	98	58	160	59
09-Feb-98	2261	NELLCOR	98	61	160	60	98	59	161	60
09-Feb-98	2262	NELLCOR	98	59	161	60	98	59	160	60
09-Feb-98	2263	NELLCOR	98	60	160	60	98	60	160	60
09-Feb-98	2264	NELLCOR	99	60	160	60	99	60	160	60
09-Feb-98	2265	NELLCOR	98	61	161	60	98	59	160	59
09-Feb-98	2266	NELLCOR	98	60	160	60	98	60	160	60
09-Feb-98	2267	NELLCOR	98	60	160	61	98	59	160	60
10-Feb-98	2268	NELLCOR	97	59	160	59	97	59	160	59
10-Feb-98	2269	NELLCOR	98	61	160	60	98	59	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
10-Feb-98	2270	NELLCOR	98	60	160	60	98	59	160	61
10-Feb-98	2271	NELLCOR	98	80	160	60	98	60	160	60
10-Feb-98	2272	DATASCOPE	98	60	160	60	97	60	160	60
10-Feb-98	2273	DATASCOPE	98	61	160	60	98	60	160	60
10-Feb-98	2274	DATASCOPE	99	60	160	60	98	60	160	59
10-Feb-98	2275	DATASCOPE	98	60	160	60	98	59	160	60
10-Feb-98	2276	DATASCOPE	97	60	160	60	98	60	160	60
10-Feb-98	2277	DATASCOPE	99	61	160	60	98	59	160	59
10-Feb-98	2278	DATASCOPE	99	61	160	59	99	60	159	60
10-Feb-98	2279	OHMEDA	98	62	159	60	98	60	160	61
10-Feb-98	2280	OHMEDA	98	59	160	61	98	61	160	60
10-Feb-98	2281	OHMEDA	97	60	160	60	98	61	160	61
10-Feb-98	2282	OHMEDA	97	60	160	60	98	60	160	60
10-Feb-98	2283	OHMEDA	98	60	160	60	98	60	160	60
10-Feb-98	2284	OHMEDA	98	60	160	60	98	60	160	60
10-Feb-98	2285	OHMEDA	98	60	159	60	98	59	160	60
11-Feb-98	2286	OHMEDA	99	60	160	59	98	60	160	60
11-Feb-98	2287	OHMEDA	98	61	160	60	98	59	160	60
11-Feb-98	2288	OHMEDA	98	60	160	60	98	59	160	61
11-Feb-98	2289	OHMEDA	98	60	160	60	98	60	160	60
11-Feb-98	2290	OHMEDA	98	60	160	60	97	60	160	60
11-Feb-98	2291	OHMEDA	98	61	160	60	98	60	160	60
11-Feb-98	2292	OHMEDA	99	60	160	60	98	60	160	59
11-Feb-98	2293	NELLCOR	98	60	160	60	98	59	160	60
11-Feb-98	2294	NELLCOR	97	60	160	60	98	60	160	60
11-Feb-98	2295	NELLCOR	99	61	160	60	98	59	160	59
12-Feb-98	2296	NELLCOR	98	61	160	59	99	60	159	60
12-Feb-98	2297	NELLCOR	98	62	159	60	98	60	160	61
12-Feb-98	2298	NELLCOR	98	59	160	61	98	61	160	60
12-Feb-98	2299	NELLCOR	98	61	160	60	98	60	160	60
12-Feb-98	2300	NELLCOR	98	60	159	61	98	61	159	60
12-Feb-98	2301	NELLCOR	97	60	160	60	98	60	160	60
12-Feb-98	2302	NELLCOR	98	60	160	60	98	61	160	60
12-Feb-98	2303	NELLCOR	99	60	160	61	98	59	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
12-Feb-98	2304	NELLCOR	97	59	160	59	97	59	160	59
12-Feb-98	2305	NELLCOR	98	61	160	60	98	59	160	60
12-Feb-98	2306	NELLCOR	98	60	160	60	98	59	160	61
13-Feb-98	2307	NELLCOR	98	60	160	60	98	60	160	60
13-Feb-98	2308	NELLCOR	98	60	160	60	97	60	160	60
13-Feb-98	2309	NELLCOR	98	61	160	60	98	60	160	60
13-Feb-98	2310	NELLCOR	99	60	160	60	98	60	160	59
13-Feb-98	2311	NELLCOR	98	60	160	60	98	59	160	60
13-Feb-98	2312	NELLCOR	97	60	160	60	98	60	160	60
13-Feb-98	2313	NELLCOR	99	61	160	60	98	59	160	59
13-Feb-98	2314	NELLCOR	99	61	160	59	99	60	159	60
13-Feb-98	2315	NELLCOR	98	62	159	60	98	60	160	61
13-Feb-98	2316	NELLCOR	98	59	160	61	98	61	160	60
13-Feb-98	2317	NELLCOR	97	60	160	60	98	61	160	61
13-Feb-98	2318	NELLCOR	98	62	160	60	97	60	159	60
13-Feb-98	2319	NELLCOR	98	60	161	60	98	60	159	59
13-Feb-98	2320	NELLCOR	97	61	159	59	97	59	160	60
13-Feb-98	2321	NELLCOR	98	59	160	60	98	60	160	60
13-Feb-98	2322	NELLCOR	99	60	160	60	98	58	160	59
13-Feb-98	2323	NELLCOR	98	61	160	60	98	59	161	60
13-Feb-98	2324	NELLCOR	98	59	161	60	98	59	160	60
13-Feb-98	2325	NELLCOR	98	60	160	60	98	60	160	60
13-Feb-98	2326	NELLCOR	99	60	160	60	99	60	160	60
13-Feb-98	2327	NELLCOR	98	61	161	60	98	59	160	59
16-Feb-98	2328	NELLCOR	98	60	160	60	98	60	160	60
16-Feb-98	2329	NELLCOR	99	60	160	61	98	59	160	60
16-Feb-98	2330	NELLCOR	97	59	160	59	97	59	160	59
16-Feb-98	2331	NELLCOR	98	61	160	60	98	59	160	60
16-Feb-98	2332	NELLCOR	98	60	160	60	98	59	160	61
17-Feb-98	2333	NELLCOR	98	60	160	60	98	60	160	60
17-Feb-98	2334	NELLCOR	98	60	160	60	97	60	160	60
17-Feb-98	2335	NELLCOR	98	61	160	60	98	60	160	60
17-Feb-98	2336	NELLCOR	99	60	160	60	98	60	160	59
17-Feb-98	2337	NELLCOR	98	60	160	60	98	59	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
17-Feb-98	2338	NELLCOR	97	60	160	60	98	60	160	60
17-Feb-98	2339	NELLCOR	99	61	160	60	98	59	160	59
17-Feb-98	2340	NELLCOR	99	61	160	59	99	60	159	60
17-Feb-98	2341	NELLCOR	98	62	159	60	98	60	160	61
17-Feb-98	2342	NELLCOR	98	59	160	61	98	61	160	60
17-Feb-98	2343	CSI	97	60	160	60	98	61	160	61
17-Feb-98	2344	OHMEDA	97	60	160	60	98	60	160	60
17-Feb-98	2345	OHMEDA	98	61	160	60	98	60	160	60
17-Feb-98	2346	OHMEDA	98	60	159	61	99	61	159	60
18-Feb-98	2347	DATASCOPE	97	60	160	60	98	60	160	60
18-Feb-98	2348	DATASCOPE	98	60	160	60	98	61	160	60
18-Feb-98	2349	DATASCOPE	99	60	160	61	98	59	160	60
18-Feb-98	2350	DATASCOPE	97	59	160	59	97	59	160	59
18-Feb-98	2351	NELLCOR	98	61	160	60	98	59	160	60
18-Feb-98	2352	NELLCOR	98	60	160	60	98	59	160	61
18-Feb-98	2353	NELLCOR	98	60	160	60	98	60	160	60
18-Feb-98	2354	NELLCOR	98	60	160	60	97	60	160	60
18-Feb-98	2355	NELLCOR	98	61	160	60	98	60	160	60
18-Feb-98	2356	NELLCOR	99	60	160	60	98	60	160	59
18-Feb-98	2357	NELLCOR	98	60	160	60	98	59	160	60
18-Feb-98	2358	NELLCOR	97	60	160	60	98	60	160	60
18-Feb-98	2359	NELLCOR	99	61	160	60	98	59	160	59
18-Feb-98	2360	NELLCOR	99	61	160	59	99	60	159	60
18-Feb-98	2361	OHMEDA	98	62	159	60	98	60	160	61
18-Feb-98	2362	OHMEDA	98	59	160	61	98	61	160	60
18-Feb-98	2363	OHMEDA	97	60	160	60	98	61	160	61
18-Feb-98	2364	OHMEDA	98	62	160	60	97	60	159	60
18-Feb-98	2365	OHMEDA	98	60	161	60	98	60	159	59
19-Feb-98	2366	OHMEDA	97	61	159	59	97	59	160	60
19-Feb-98	2367	OHMEDA	98	59	160	60	98	60	160	60
19-Feb-98	2368	OHMEDA	99	60	160	60	98	58	160	59
19-Feb-98	2369	OHMEDA	98	61	160	60	98	59	161	60
19-Feb-98	2370	OHMEDA	98	59	161	60	98	59	160	60
19-Feb-98	2371	OHMEDA	98	60	160	60	98	60	160	60

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
19-Feb-98	2372	OHMEDA	99	60	160	80	99	60	160	60
20-Feb-98	2373	OHMEDA	98	61	161	80	98	59	160	59
20-Feb-98	2374	OHMEDA	98	60	160	60	98	60	160	60
20-Feb-98	2375	OHMEDA	99	60	160	61	98	59	160	60
20-Feb-98	2376	OHMEDA	97	59	160	59	97	59	160	59
20-Feb-98	2377	OHMEDA	98	61	160	60	98	59	160	60
23-Feb-98	2378	OHMEDA	98	60	160	60	98	59	160	61
23-Feb-98	2379	OHMEDA	98	60	160	60	98	60	160	60
23-Feb-98	2380	NELLCOR	98	60	160	60	97	60	160	60
23-Feb-98	2381	NELLCOR	98	61	160	60	98	60	160	60
23-Feb-98	2382	NELLCOR	99	60	160	60	98	60	160	59
23-Feb-98	2383	NELLCOR	98	60	160	60	98	59	160	60
23-Feb-98	2384	NELLCOR	97	60	160	60	98	60	160	60
23-Feb-98	2385	NELLCOR	99	61	160	60	98	59	160	59
23-Feb-98	2386	NELLCOR	99	61	160	59	99	60	159	60
23-Feb-98	2387	NELLCOR	98	62	159	80	98	60	160	61
24-Feb-98	2388	NELLCOR	98	59	160	61	98	61	160	60
24-Feb-98	2389	NELLCOR	97	60	160	60	98	61	160	61
24-Feb-98	2390	NOVAMETRIX	97	60	160	60	98	60	160	60
24-Feb-98	2391	NOVAMETRIX	98	60	160	60	98	60	160	60
24-Feb-98	2392	NOVAMETRIX	98	60	160	60	98	60	160	60
24-Feb-98	2393	NOVAMETRIX	98	60	159	60	98	59	160	60
24-Feb-98	2394	NOVAMETRIX	99	60	160	59	98	60	160	60
24-Feb-98	2395	NOVAMETRIX	98	61	160	60	98	59	160	60
24-Feb-98	2396	NOVAMETRIX	98	60	160	60	98	59	160	61
24-Feb-98	2397	NOVAMETRIX	98	60	160	60	98	60	160	60
24-Feb-98	2398	NOVAMETRIX	98	60	160	60	97	60	160	60
24-Feb-98	2399	NOVAMETRIX	98	61	160	60	98	60	160	60
24-Feb-98	2400	NOVAMETRIX	99	60	160	60	98	60	160	59
24-Feb-98	2401	CSI	98	60	160	60	98	59	160	60
25-Feb-98	2402	CSI	97	60	160	60	98	60	160	60
25-Feb-98	2403	NELLCOR	99	61	160	60	98	59	160	59
25-Feb-98	2404	NELLCOR	98	61	160	59	99	60	159	60
25-Feb-98	2405	NELLCOR	98	62	159	60	98	60	160	61

DATE	I. D. #	MANUFACTURE	BIO-TEK INDEX				VIAMED DL3000			
			SPO ₂		PULSE RATE		SPO ₂		PULSE RATE	
			98	60	160	60	98	60	160	60
25-Feb-98	2406	NELLCOR	98	61	160	60	98	60	160	60
25-Feb-98	2407	NELLCOR	98	60	159	61	99	61	159	60
25-Feb-98	2408	NELLCOR	97	60	160	60	98	60	160	60
25-Feb-98	2409	NELLCOR	98	60	160	60	98	61	160	60
25-Feb-98	2410	NELLCOR	99	60	160	61	98	59	160	60
25-Feb-98	2411	NELLCOR	97	59	160	59	97	59	160	59
25-Feb-98	2412	OHMEDA	98	61	160	60	98	59	160	60
25-Feb-98	2413	OHMEDA	98	60	160	60	98	59	160	61
25-Feb-98	2414	OHMEDA	98	60	160	60	98	60	160	60
26-Feb-98	2415	OHMEDA	98	60	160	60	97	60	160	60
26-Feb-98	2416	OHMEDA	98	61	160	60	98	60	160	60
26-Feb-98	2417	OHMEDA	98	60	160	60	98	60	160	59
26-Feb-98	2418	OHMEDA	98	60	160	60	98	59	160	60
26-Feb-98	2419	OHMEDA	97	60	160	60	98	60	160	60
26-Feb-98	2420	OHMEDA	99	61	160	60	98	59	160	59
26-Feb-98	2421	OHMEDA	99	61	160	59	99	60	159	60
26-Feb-98	2422	NOVAMETRIX	98	62	159	60	98	60	160	61
26-Feb-98	2423	NOVAMETRIX	98	59	160	61	98	61	160	60
26-Feb-98	2424	NOVAMETRIX	97	60	160	60	98	61	160	61
26-Feb-98	2425	NOVAMETRIX	98	62	160	60	97	60	159	60
26-Feb-98	2426	DATASCOPE	98	60	161	60	98	60	159	59
26-Feb-98	2427	BCI	97	61	159	59	97	59	160	60
26-Feb-98	2428	BCI	98	59	160	60	98	60	160	60
26-Feb-98	2429	BCI	99	60	160	60	98	58	160	59
26-Feb-98	2430	BCI	98	61	160	60	98	59	161	60
26-Feb-98	2431	BCI	98	59	161	60	98	59	160	60
27-Feb-98	2432	BCI	98	60	160	60	98	60	160	60
27-Feb-98	2433	BCI	99	60	160	60	99	60	160	60
27-Feb-98	2434	BCI	98	61	161	60	98	59	160	59
27-Feb-98	2435	BCI	98	60	160	60	98	60	160	60
27-Feb-98	2436	BCI	99	60	160	61	98	59	160	60
27-Feb-98	2437	BCI	97	59	160	59	97	59	160	59
27-Feb-98	2438	BCI	98	61	160	60	98	59	160	60
27-Feb-98	2439	BCI	98	60	160	60	98	59	160	61