

A6.1 Test Results P856RA
Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

SpO₂ Testing
Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9732-1	7H00001	98	94	80	70	60	0	0	0	0	0
9732-1	7H00002	98	94	80	70	60	0	0	0	0	0
9732-1	7H00003	98	94	80	70	60	0	0	0	0	0
9732-1	7H00004	98	94	80	70	60	0	0	0	0	0
9732-1	7H00005	98	94	80	70	60	0	0	0	0	0
9732-1	7H00006	98	94	80	70	60	0	0	0	0	0
9732-1	7H00007	98	94	80	70	60	0	0	0	0	0
9732-1	7H00008	98	94	80	69	60	0	0	0	-1	0
9732-1	7H00009	98	94	80	70	60	0	0	0	0	0
9732-1	7H00010	98	94	80	70	60	0	0	0	0	0
9732-1	7H00011	98	94	80	70	60	0	0	0	0	0
9732-1	7H00012	98	94	80	70	60	0	0	0	0	0
9732-1	7H00013	98	94	80	70	60	0	0	0	0	0
9732-1	7H00014	98	94	80	70	60	0	0	0	0	0
9732-1	7H00015	98	94	80	70	60	0	0	0	0	0
9732-1	7H00016	98	94	80	70	60	0	0	0	0	0
9732-1	7H00017	98	94	80	70	60	0	0	0	0	0
9732-1	7H00018	98	94	80	70	60	0	0	0	0	0
9732-1	7H00019	98	94	80	70	60	0	0	0	0	0
9732-1	7H00020	98	94	80	70	60	0	0	0	0	0
9732-1	7H00021	98	94	80	70	60	0	0	0	0	0
9732-1	7H00022	98	94	80	70	60	0	0	0	0	0
9732-1	7H00023	98	94	80	70	60	0	0	0	0	0
9732-1	7H00024	98	94	80	70	60	0	0	0	0	0
9732-1	7H00025	98	94	80	70	60	0	0	0	0	0
9732-1	7H00026	98	94	80	70	60	0	0	0	0	0
9732-1	7H00027	98	94	80	70	60	0	0	0	0	0
9732-1	7H00028	98	94	80	70	60	0	0	0	0	0
9732-1	7H00029	98	94	80	70	60	0	0	0	0	0
9732-1	7H00030	98	94	80	70	60	0	0	0	0	0
9732-1	7H00031	98	94	80	70	60	0	0	0	0	0
9732-1	7H00032	98	94	80	70	60	0	0	0	0	0
9732-1	7H00033	98	94	80	70	60	0	0	0	0	0
9732-1	7H00034	98	94	80	70	60	0	0	0	0	0
9732-1	7H00035	98	94	80	70	60	0	0	0	0	0
9732-1	7H00036	98	94	80	70	60	0	0	0	0	0
9732-1	7H00037	98	94	80	70	60	0	0	0	0	0
9732-1	7H00038	98	94	80	70	60	0	0	0	0	0
9732-1	7H00039	98	94	80	70	60	0	0	0	0	0
9732-1	7H00040	98	94	80	70	60	0	0	0	0	0
9732-1	7H00041	98	94	80	70	60	0	0	0	0	0
9732-1	7H00042	98	94	80	70	60	0	0	0	0	0
9732-1	7H00043	98	94	80	70	60	0	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-1	7H00044	98	94	80	70	60	0	0	0	0	0
9732-1	7H00045	98	94	80	70	60	0	0	0	0	0
9732-1	7H00046	98	94	80	70	60	0	0	0	0	0
9732-1	7H00047	98	94	80	70	60	0	0	0	0	0
9732-1	7H00048	98	94	80	70	60	0	0	0	0	0
9732-1	7H00049	98	94	80	70	60	0	0	0	0	0
9732-1	7H00050	98	94	80	70	60	0	0	0	0	0
9732-2	7H00051	98	94	80	70	60	0	0	0	0	0
9732-2	7H00052	98	94	80	70	60	0	0	0	0	0
9732-2	7H00053	98	94	80	70	60	0	0	0	0	0
9732-2	7H00054	98	94	80	70	60	0	0	0	0	0
9732-2	7H00055	98	94	80	70	60	0	0	0	0	0
9732-2	7H00056	98	94	80	70	60	0	0	0	0	0
9732-2	7H00057	98	94	80	70	60	0	0	0	0	0
9732-2	7H00058	98	94	80	70	60	0	0	0	0	0
9732-2	7H00059	98	94	80	70	60	0	0	0	0	0
9732-2	7H00060	98	94	80	70	60	0	0	0	0	0
9732-2	7H00061	98	94	80	70	60	0	0	0	0	0
9732-2	7H00062	98	94	80	70	60	0	0	0	0	0
9732-2	7H00063	98	94	80	70	60	0	0	0	0	0
9732-2	7H00064	98	94	80	70	60	0	0	0	0	0
9732-2	7H00065	98	94	80	70	60	0	0	0	0	0
9732-2	7H00066	98	94	80	70	60	0	0	0	0	0
9732-2	7H00067	98	94	80	70	60	0	0	0	0	0
9732-2	7H00068	98	94	80	70	60	0	0	0	0	0
9732-2	7H00069	98	94	80	70	60	0	0	0	0	0
9732-2	7H00070	98	94	80	70	60	0	0	0	0	0
9732-2	7H00071	98	94	80	70	60	0	0	0	0	0
9732-2	7H00072	98	94	80	70	60	0	0	0	0	0
9732-2	7H00073	98	94	80	70	60	0	0	0	0	0
9732-2	7H00074	98	94	80	70	60	0	0	0	0	0
9732-2	7H00075	98	94	80	70	60	0	0	0	0	0
9732-2	7H00076	98	94	80	70	60	0	0	0	0	0
9732-2	7H00077	98	94	80	70	60	0	0	0	0	0
9732-2	7H00078	98	94	80	70	60	0	0	0	0	0
9732-2	7H00079	98	94	80	70	60	0	0	0	0	0
9732-2	7H00080	98	94	80	70	60	0	0	0	0	0
9732-2	7H00081	98	94	80	70	60	0	0	0	0	0
9732-2	7H00082	98	94	80	70	60	0	0	0	0	0
9732-2	7H00083	98	94	80	70	60	0	0	0	0	0
9732-2	7H00084	98	94	80	70	60	0	0	0	0	0
9732-2	7H00085	98	94	80	70	60	0	0	0	0	0
9732-2	7H00086	98	94	80	70	60	0	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	98	94	80	70	60	0	0	0	0	0
9732-2	7H00088	98	94	80	70	60	0	0	0	0	0
9732-2	7H00089	98	94	80	70	60	0	0	0	0	0
9732-2	7H00090	98	94	80	70	60	0	0	0	0	0
9732-2	7H00091	98	94	80	70	60	0	0	0	0	0
9732-2	7H00092	98	94	80	70	60	0	0	0	0	0
9732-2	7H00093	98	94	80	70	60	0	0	0	0	0
9732-2	7H00094	98	94	80	70	60	0	0	0	0	0
9732-2	7H00095	98	94	80	70	60	0	0	0	0	0
9732-2	7H00096	98	94	80	70	60	0	0	0	0	0
9732-2	7H00097	98	94	80	70	60	0	0	0	0	0
9732-2	7H00098	98	94	80	70	60	0	0	0	0	0
9732-2	7H00099	98	94	80	70	60	0	0	0	0	0

Tester: [Signature]

Date: 9/23/97
10/10/97

Q.A.: [Signature]

Date: 9/23/97

A6.1 Test Results P856RA

Pulse Rate Testing

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9732-1	7H00001	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00002	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00003	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00004	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00005	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00006	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00007	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00008	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00009	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00010	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00011	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00012	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00013	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00014	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00015	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00016	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00017	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00018	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00019	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00020	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00021	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00022	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00023	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00024	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00025	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00026	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00027	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00028	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00029	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00030	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00031	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00032	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00033	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00034	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00035	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00036	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00037	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00038	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00039	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00040	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00041	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00042	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00043	249	150	95	80	50	-1	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-1	7H00044	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00045	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00046	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00047	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00048	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00049	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00050	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00051	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00052	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00053	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00054	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00055	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00056	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00057	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00058	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00059	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00060	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00061	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00062	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00063	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00064	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00065	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00066	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00067	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00068	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00069	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00070	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00071	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00072	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00073	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00074	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00075	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00076	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00077	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00078	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00079	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00080	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00081	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00082	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00083	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00084	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00085	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00086	249	149	95	80	50	-1	-1	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00088	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00089	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00090	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00091	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00092	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00093	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00094	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00095	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00096	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00097	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00098	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00099	249	150	95	80	50	-1	0	0	0	0

Tester: [Signature]

Date: 9/23/97

Q.A.: [Signature]

Date: 9/23/97

A6.1 Test Results P856RA**Pulse Oximeter Monitor: Nellcor N200**
Monitor S/N: 20619220**SpO₂ Testing****Finger Probe Model: P856RA**
BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9732-1	7H00001	98	94	80	70	60	0	0	0	0	0
9732-1	7H00002	98	94	80	70	60	0	0	0	0	0
9732-1	7H00003	98	94	80	70	60	0	0	0	0	0
9732-1	7H00004	98	94	80	70	60	0	0	0	0	0
9732-1	7H00005	98	94	80	70	60	0	0	0	0	0
9732-1	7H00006	98	94	80	70	60	0	0	0	0	0
9732-1	7H00007	98	94	80	70	60	0	0	0	0	0
9732-1	7H00008	98	94	80	69	60	0	0	0	-1	0
9732-1	7H00009	98	94	80	70	60	0	0	0	0	0
9732-1	7H00010	98	94	80	70	60	0	0	0	0	0
9732-1	7H00011	98	94	80	70	60	0	0	0	0	0
9732-1	7H00012	98	94	80	70	60	0	0	0	0	0
9732-1	7H00013	98	94	80	70	60	0	0	0	0	0
9732-1	7H00014	98	94	80	70	60	0	0	0	0	0
9732-1	7H00015	98	94	80	70	60	0	0	0	0	0
9732-1	7H00016	98	94	80	70	60	0	0	0	0	0
9732-1	7H00017	98	94	80	70	60	0	0	0	0	0
9732-1	7H00018	98	94	80	70	60	0	0	0	0	0
9732-1	7H00019	98	94	80	70	60	0	0	0	0	0
9732-1	7H00020	98	94	80	70	60	0	0	0	0	0
9732-1	7H00021	98	94	80	70	60	0	0	0	0	0
9732-1	7H00022	98	94	80	70	60	0	0	0	0	0
9732-1	7H00023	98	94	80	70	60	0	0	0	0	0
9732-1	7H00024	98	94	80	70	60	0	0	0	0	0
9732-1	7H00025	98	94	80	70	60	0	0	0	0	0
9732-1	7H00026	98	94	80	70	60	0	0	0	0	0
9732-1	7H00027	98	94	80	70	60	0	0	0	0	0
9732-1	7H00028	98	94	80	70	60	0	0	0	0	0
9732-1	7H00029	98	94	80	70	60	0	0	0	0	0
9732-1	7H00030	98	94	80	70	60	0	0	0	0	0
9732-1	7H00031	98	94	80	70	60	0	0	0	0	0
9732-1	7H00032	98	94	80	70	60	0	0	0	0	0
9732-1	7H00033	98	94	80	70	60	0	0	0	0	0
9732-1	7H00034	98	94	80	70	60	0	0	0	0	0
9732-1	7H00035	98	94	80	70	60	0	0	0	0	0
9732-1	7H00036	98	94	80	70	60	0	0	0	0	0
9732-1	7H00037	98	94	80	70	60	0	0	0	0	0
9732-1	7H00038	98	94	80	70	60	0	0	0	0	0
9732-1	7H00039	98	94	80	70	60	0	0	0	0	0
9732-1	7H00040	98	94	80	70	60	0	0	0	0	0
9732-1	7H00041	98	94	80	70	60	0	0	0	0	0
9732-1	7H00042	98	94	80	70	60	0	0	0	0	0
9732-1	7H00043	98	94	80	70	60	0	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-1	7H00044	98	94	80	70	60	0	0	0	0	0
9732-1	7H00045	98	94	80	70	60	0	0	0	0	0
9732-1	7H00046	98	94	80	70	60	0	0	0	0	0
9732-1	7H00047	98	94	80	70	60	0	0	0	0	0
9732-1	7H00048	98	94	80	70	60	0	0	0	0	0
9732-1	7H00049	98	94	80	70	60	0	0	0	0	0
9732-1	7H00050	98	94	80	70	60	0	0	0	0	0
9732-2	7H00051	98	94	80	70	60	0	0	0	0	0
9732-2	7H00052	98	94	80	70	60	0	0	0	0	0
9732-2	7H00053	98	94	80	70	60	0	0	0	0	0
9732-2	7H00054	98	94	80	70	60	0	0	0	0	0
9732-2	7H00055	98	94	80	70	60	0	0	0	0	0
9732-2	7H00056	98	94	80	70	60	0	0	0	0	0
9732-2	7H00057	98	94	80	70	60	0	0	0	0	0
9732-2	7H00058	98	94	80	70	60	0	0	0	0	0
9732-2	7H00059	98	94	80	70	60	0	0	0	0	0
9732-2	7H00060	98	94	80	70	60	0	0	0	0	0
9732-2	7H00061	98	94	80	70	60	0	0	0	0	0
9732-2	7H00062	98	94	80	70	60	0	0	0	0	0
9732-2	7H00063	98	94	80	70	60	0	0	0	0	0
9732-2	7H00064	98	94	80	70	60	0	0	0	0	0
9732-2	7H00065	98	94	80	70	60	0	0	0	0	0
9732-2	7H00066	98	94	80	70	60	0	0	0	0	0
9732-2	7H00067	98	94	80	70	60	0	0	0	0	0
9732-2	7H00068	98	94	80	70	60	0	0	0	0	0
9732-2	7H00069	98	94	80	70	60	0	0	0	0	0
9732-2	7H00070	98	94	80	70	60	0	0	0	0	0
9732-2	7H00071	98	94	80	70	60	0	0	0	0	0
9732-2	7H00072	98	94	80	70	60	0	0	0	0	0
9732-2	7H00073	98	94	80	70	60	0	0	0	0	0
9732-2	7H00074	98	94	80	70	60	0	0	0	0	0
9732-2	7H00075	98	94	80	70	60	0	0	0	0	0
9732-2	7H00076	98	94	80	70	60	0	0	0	0	0
9732-2	7H00077	98	94	80	70	60	0	0	0	0	0
9732-2	7H00078	98	94	80	70	60	0	0	0	0	0
9732-2	7H00079	98	94	80	70	60	0	0	0	0	0
9732-2	7H00080	98	94	80	70	60	0	0	0	0	0
9732-2	7H00081	98	94	80	70	60	0	0	0	0	0
9732-2	7H00082	98	94	80	70	60	0	0	0	0	0
9732-2	7H00083	98	94	80	70	60	0	0	0	0	0
9732-2	7H00084	98	94	80	70	60	0	0	0	0	0
9732-2	7H00085	98	94	80	70	60	0	0	0	0	0
9732-2	7H00086	98	94	80	70	60	0	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	98	94	80	70	60	0	0	0	0	0
9732-2	7H00088	98	94	80	70	60	0	0	0	0	0
9732-2	7H00089	98	94	80	70	60	0	0	0	0	0
9732-2	7H00090	98	94	80	70	60	0	0	0	0	0
9732-2	7H00091	98	94	80	70	60	0	0	0	0	0
9732-2	7H00092	98	94	80	70	60	0	0	0	0	0
9732-2	7H00093	98	94	80	70	60	0	0	0	0	0
9732-2	7H00094	98	94	80	70	60	0	0	0	0	0
9732-2	7H00095	98	94	80	70	60	0	0	0	0	0
9732-2	7H00096	98	94	80	70	60	0	0	0	0	0
9732-2	7H00097	98	94	80	70	60	0	0	0	0	0
9732-2	7H00098	98	94	80	70	60	0	0	0	0	0
9732-2	7H00099	98	94	80	70	60	0	0	0	0	0

Tester: [Signature]

Date: 9/23/97
10/10/97

Q.A.: [Signature]

Date: 9/23/97

A6.1 Test Results P856RA**Pulse Rate Testing****Pulse Oximeter Monitor: Nellcor N200****Finger Probe Model: P856RA****Monitor S/N: 20619220****BioTek Simulator S/N: 125672**

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9732-1	7H00001	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00002	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00003	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00004	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00005	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00006	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00007	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00008	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00009	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00010	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00011	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00012	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00013	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00014	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00015	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00016	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00017	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00018	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00019	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00020	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00021	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00022	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00023	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00024	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00025	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00026	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00027	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00028	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00029	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00030	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00031	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00032	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00033	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00034	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00035	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00036	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00037	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00038	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00039	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00040	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00041	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00042	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00043	249	150	95	80	50	-1	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-1	7H00044	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00045	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00046	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00047	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00048	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00049	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00050	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00051	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00052	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00053	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00054	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00055	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00056	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00057	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00058	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00059	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00060	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00061	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00062	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00063	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00064	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00065	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00066	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00067	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00068	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00069	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00070	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00071	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00072	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00073	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00074	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00075	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00076	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00077	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00078	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00079	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00080	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00081	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00082	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00083	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00084	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00085	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00086	249	149	95	80	50	-1	-1	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00088	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00089	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00090	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00091	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00092	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00093	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00094	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00095	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00096	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00097	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00098	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00099	249	150	95	80	50	-1	0	0	0	0

Tester: *John A. Chen*

Date: 9/23/97

Q.A.: *[Signature]*

Date: 9/23/97

A6.2 Test Results P861RA
Pulse Oximeter Monitor: BCI 71000A2
Monitor S/N: 170276621

Pulse Rate Testing
Finger Probe Model: P861RA
BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9731-2	7H00147	250	150	95	80	50	0	0	0	0	0
9731-2	7H00148	250	150	95	80	50	0	0	0	0	0
9731-2	7H00149	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00150	250	150	95	80	50	0	0	0	0	0
9731-2	7H00151	250	150	95	80	50	0	0	0	0	0
9731-2	7H00152	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00153	250	150	95	80	50	0	0	0	0	0
9731-2	7H00154	250	150	95	80	50	0	0	0	0	0
9731-2	7H00155	250	150	95	80	50	0	0	0	0	0
9731-2	7H00156	250	150	95	80	50	0	0	0	0	0
9731-2	7H00157	250	150	95	80	50	0	0	0	0	0
9731-2	7H00158	250	150	95	80	50	0	0	0	0	0
9731-2	7H00159	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00160	250	150	95	80	50	0	0	0	0	0
9731-2	7H00161	250	150	95	80	50	0	0	0	0	0
9731-2	7H00162	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00163	250	150	95	80	50	0	0	0	0	0
9731-2	7H00164	250	150	95	80	50	0	0	0	0	0
9731-2	7H00165	250	150	95	80	50	0	0	0	0	0
9731-2	7H00166	250	150	95	80	50	0	0	0	0	0
9731-2	7H00167	250	150	95	80	50	0	0	0	0	0
9731-2	7H00168	250	150	95	80	50	0	0	0	0	0
9731-2	7H00169	250	150	95	80	50	0	0	0	0	0
9731-2	7H00170	250	150	95	80	50	0	0	0	0	0
9731-2	7H00171	250	150	95	80	50	0	0	0	0	0
9731-2	7H00172	250	150	95	80	50	0	0	0	0	0
9731-2	7H00173	250	150	95	80	50	0	0	0	0	0
9731-2	7H00174	250	150	95	80	50	0	0	0	0	0
9731-2	7H00175	250	150	95	80	50	0	0	0	0	0
9731-2	7H00176	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00177	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00178	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00179	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00180	250	150	95	80	50	0	0	0	0	0
9731-2	7H00181	250	150	95	80	50	0	0	0	0	0
9731-2	7H00182	250	150	95	80	50	0	0	0	0	0
9731-2	7H00183	250	150	95	80	50	0	0	0	0	0
9731-2	7H00184	250	150	95	80	50	0	0	0	0	0
9731-2	7H00185	250	150	95	80	50	0	0	0	0	0
9731-2	7H00186	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00187	250	150	95	80	50	0	0	0	0	0
9731-2	7H00188	249	150	95	80	50	-1	0	0	0	0
9731-2	7H00189	250	150	95	80	50	0	0	0	0	0
9731-2	7H00190	250	150	95	80	50	0	0	0	0	0
9731-2	7H00191	250	150	95	80	50	0	0	0	0	0
9731-2	7H00192	250	150	95	80	50	0	0	0	0	0
9731-2	7H00193	250	150	95	80	50	0	0	0	0	0
9731-2	7H00194	250	150	95	80	50	0	0	0	0	0

Pulse Oximeter Monitor: BCI 71000A2
Monitor S/N: 170276621

Finger Probe Model: P861RA
BioTek Simulator S/N: 125672

9731-2	7H00195	98	94	80	70	60	0	0	0	0	0
9731-2	7H00196	98	94	80	70	60	0	0	0	0	0
9731-3	7H00197	98	94	80	70	60	0	0	0	0	0
9731-3	7H00198	98	94	80	70	60	0	0	0	0	0
9731-3	7H00199	98	94	80	70	60	0	0	0	0	0
9731-3	7H00200	98	94	80	70	60	0	0	0	0	0
9731-3	7H00201	98	94	80	70	60	0	0	0	0	0
9731-3	7H00202	98	94	80	70	60	0	0	0	0	0
9731-3	7H00203	98	94	80	70	60	0	0	0	0	0
9731-3	7H00204	98	94	80	70	60	0	0	0	0	0
9731-3	7H00205	98	94	80	70	60	0	0	0	0	0
9731-3	7H00206	98	94	80	70	60	0	0	0	0	0
9731-3	7H00207	98	94	80	70	60	0	0	0	0	0
9731-3	7H00208	98	94	80	70	60	0	0	0	0	0
9731-3	7H00209	98	94	80	70	60	0	0	0	0	0
9731-3	7H00210	98	94	80	70	60	0	0	0	0	0
9731-3	7H00211	98	94	80	70	60	0	0	0	0	0
9731-3	7H00212	98	94	80	70	60	0	0	0	0	0
9731-3	7H00213	98	94	80	70	60	0	0	0	0	0
9731-3	7H00214	98	94	80	70	60	0	0	0	0	0
9731-3	7H00215	98	94	80	70	60	0	0	0	0	0
9731-3	7H00216	98	94	80	70	59	0	0	0	0	-1
9731-3	7H00217	98	94	80	70	60	0	0	0	0	0
9731-3	7H00218	98	94	80	70	60	0	0	0	0	0
9731-3	7H00219	98	94	80	70	60	0	0	0	0	0
9731-3	7H00220	98	94	80	70	59	0	0	0	0	-1
9731-3	7H00221	98	94	81	71	61	0	0	1	1	1
9731-3	7H00222	98	94	80	70	60	0	0	0	0	0
9731-3	7H00223	98	94	80	70	60	0	0	0	0	0
9731-3	7H00224	98	94	80	70	60	0	0	0	0	0
9731-3	7H00225	98	94	80	70	60	0	0	0	0	0
9731-3	7H00226	98	94	80	70	60	0	0	0	0	0
9731-3	7H00227	98	94	80	70	60	0	0	0	0	0
9731-3	7H00228	98	94	80	70	60	0	0	0	0	0
9731-3	7H00229	98	94	80	70	60	0	0	0	0	0
9731-3	7H00230	98	94	80	70	60	0	0	0	0	0
9731-3	7H00231	98	94	80	70	60	0	0	0	0	0
9731-3	7H00232	98	94	80	70	60	0	0	0	0	0
9731-3	7H00233	98	94	80	70	60	0	0	0	0	0
9731-3	7H00234	98	94	80	70	60	0	0	0	0	0
9731-3	7H00235	98	94	80	70	60	0	0	0	0	0
9731-3	7H00236	98	94	80	70	60	0	0	0	0	0
9731-3	7H00237	98	94	80	70	60	0	0	0	0	0
9731-3	7H00238	98	94	81	71	61	0	0	1	1	1
9731-3	7H00239	98	94	80	70	60	0	0	0	0	0
9731-3	7H00240	98	94	80	70	60	0	0	0	0	0
9731-3	7H00241	98	94	80	70	60	0	0	0	0	0
9731-3	7H00242	98	94	80	70	60	0	0	0	0	0
9731-3	7H00243	98	94	80	70	60	0	0	0	0	0
9731-3	7H00244	98	94	80	70	60	0	0	0	0	0
9731-3	7H00245	98	94	80	70	60	0	0	0	0	0

Tester: Peter J. Chea

Date: 9/23/97

Q.A.: JB

Date: 9/23/97

Pulse Oximeter Monitor: BCI 71000A2
Monitor S/N: 170276621

Finger Probe Model: P861RA
BioTek Simulator S/N: 125672

9731-2	7H00195	250	150	95	80	50	0	0	0	0	0
9731-2	7H00196	250	150	95	80	50	0	0	0	0	0
9731-3	7H00197	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00198	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00199	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00200	250	150	95	80	50	0	0	0	0	0
9731-3	7H00201	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00202	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00203	250	150	95	80	50	0	0	0	0	0
9731-3	7H00204	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00205	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00206	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00207	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00208	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00209	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00210	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00211	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00212	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00213	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00214	250	150	95	80	50	0	0	0	0	0
9731-3	7H00215	250	150	95	80	50	0	0	0	0	0
9731-3	7H00216	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00217	250	150	95	80	50	0	0	0	0	0
9731-3	7H00218	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00219	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00220	250	150	95	80	50	0	0	0	0	0
9731-3	7H00221	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00222	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00223	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00224	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00225	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00226	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00227	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00228	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00229	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00230	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00231	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00232	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00233	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00234	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00235	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00236	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00237	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00238	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00239	250	150	95	80	50	0	0	0	0	0
9731-3	7H00240	250	150	95	80	50	0	0	0	0	0
9731-3	7H00241	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00242	250	150	95	80	50	0	0	0	0	0
9731-3	7H00243	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00244	249	150	95	80	50	-1	0	0	0	0
9731-3	7H00245	249	150	95	80	50	-1	0	0	0	0

Tester: [Signature]

Date: 7/23/97

Q.A.: [Signature]

Date: 8/23/97

A6.2 Test Results P861RA**Pulse Oximeter Monitor: BCI 71000A2****Monitor S/N: 170276621****SpO₂ Testing****Finger Probe Model: P861RA****BioTek Simulator S/N: 125672**

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9731-2	7H00147	98	94	80	70	60	0	0	0	0	0
9731-2	7H00148	98	94	80	70	60	0	0	0	0	0
9731-2	7H00149	98	94	80	70	60	0	0	0	0	0
9731-2	7H00150	98	94	80	70	60	0	0	0	0	0
9731-2	7H00151	98	94	80	70	60	0	0	0	0	0
9731-2	7H00152	98	94	80	70	60	0	0	0	0	0
9731-2	7H00153	98	94	80	70	60	0	0	0	0	0
9731-2	7H00154	98	94	80	70	60	0	0	0	0	0
9731-2	7H00155	98	94	80	70	60	0	0	0	0	0
9731-2	7H00156	98	94	80	70	60	0	0	0	0	0
9731-2	7H00157	98	94	80	70	60	0	0	0	0	0
9731-2	7H00158	98	94	80	70	60	0	0	0	0	0
9731-2	7H00159	98	94	80	70	60	0	0	0	0	0
9731-2	7H00160	98	94	80	70	60	0	0	0	0	0
9731-2	7H00161	98	94	80	70	60	0	0	0	0	0
9731-2	7H00162	98	94	80	70	60	0	0	0	0	0
9731-2	7H00163	98	94	80	70	60	0	0	0	0	0
9731-2	7H00164	98	94	80	70	60	0	0	0	0	0
9731-2	7H00165	98	94	80	70	60	0	0	0	0	0
9731-2	7H00166	98	94	80	70	60	0	0	0	0	0
9731-2	7H00167	98	94	80	70	60	0	0	0	0	0
9731-2	7H00168	98	94	80	70	60	0	0	0	0	0
9731-2	7H00169	98	94	80	70	60	0	0	0	0	0
9731-2	7H00170	98	94	80	70	60	0	0	0	0	0
9731-2	7H00171	98	94	80	70	60	0	0	0	0	0
9731-2	7H00172	98	94	80	70	60	0	0	0	0	0
9731-2	7H00173	98	94	80	70	60	0	0	0	0	0
9731-2	7H00174	98	94	80	70	60	0	0	0	0	0
9731-2	7H00175	98	94	80	70	60	0	0	0	0	0
9731-2	7H00176	98	94	80	70	60	0	0	0	0	0
9731-2	7H00177	98	94	80	70	60	0	0	0	0	0
9731-2	7H00178	98	94	80	70	60	0	0	0	0	0
9731-2	7H00179	98	94	80	70	60	0	0	0	0	0
9731-2	7H00180	98	94	80	70	60	0	0	0	0	0
9731-2	7H00181	98	94	80	70	60	0	0	0	0	0
9731-2	7H00182	98	94	80	70	60	0	0	0	0	0
9731-2	7H00183	98	94	80	70	60	0	0	0	0	0
9731-2	7H00184	98	94	80	70	60	0	0	0	0	0
9731-2	7H00185	98	94	80	70	60	0	0	0	0	0
9731-2	7H00186	98	94	80	70	60	0	0	0	0	0
9731-2	7H00187	98	94	80	70	60	0	0	0	0	0
9731-2	7H00188	98	94	80	70	60	0	0	0	0	0
9731-2	7H00189	98	94	80	70	60	0	0	0	0	0
9731-2	7H00190	98	94	80	70	60	0	0	0	0	0
9731-2	7H00191	98	94	80	70	60	0	0	0	0	0
9731-2	7H00192	98	94	80	70	60	0	0	0	0	0
9731-2	7H00193	98	94	80	71	60	0	0	0	1	0
9731-2	7H00194	98	94	80	70	60	0	0	0	0	0

A6.3 P862RA Test Results**Oximeter Monitor: CSI 504****Monitor S/N: 291108486****SpO₂ Testing****Finger Probe Model: P862RA****BioTek Simulator S/N: 125672**

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9731-2	7H00880	99	95	82	71	62	1	1	2	1	2
9731-2	7H00881	98	95	82	70	62	0	1	2	0	2
9731-2	7H00882	99	96	82	71	62	1	2	2	1	2
9731-2	7H00883	98	96	81	71	61	0	2	1	1	1
9731-2	7H00884	99	96	82	71	62	1	2	2	1	2
9731-2	7H00885	98	96	80	70	62	0	2	0	0	2
9731-2	7H00886	98	95	80	70	60	0	1	0	0	0
9731-2	7H00887	98	96	82	71	61	0	2	2	1	1
9731-2	7H00888	99	96	82	71	62	1	2	2	1	2
9731-2	7H00889	98	96	81	71	61	0	2	1	1	1
9731-2	7H00890	98	96	82	71	61	0	2	2	1	1
9731-2	7H00891	98	96	82	71	61	0	2	2	1	1
9731-2	7H00893	99	96	80	70	60	1	2	0	0	0
9731-2	7H00894	99	96	82	70	60	1	2	2	0	0
9731-2	7H00895	99	96	81	71	62	1	2	1	1	2
9731-2	7H00897	98	95	82	70	62	0	1	2	0	2
9731-2	7H00898	98	96	81	71	60	0	2	1	1	0
9731-2	7H00899	98	96	82	71	61	0	2	2	1	1
9731-2	7H00900	98	95	80	71	60	0	1	0	1	0
9731-2	7H00901	99	96	82	71	62	1	2	2	1	2
9731-2	7H00902	99	96	80	70	60	1	2	0	0	0
9731-2	7H00903	99	96	82	71	62	1	2	2	1	2
9731-2	7H00904	98	95	81	71	60	0	1	1	1	0
9731-2	7H00905	99	95	81	70	60	1	1	1	0	0
9731-2	7H00906	98	96	82	71	60	0	2	2	1	0
9731-2	7H00907	98	96	81	71	61	0	2	1	1	1
9731-2	7H00908	98	96	81	70	60	0	2	1	0	0
9731-2	7H00909	98	95	80	70	60	0	1	0	0	0
9731-2	7H00910	98	95	82	71	61	0	1	2	1	1
9731-2	7H00911	98	95	82	71	61	0	1	2	1	1
9731-2	7H00912	99	95	82	71	60	1	1	2	1	0
9731-2	7H00913	98	96	82	74	62	0	2	2	4	2
9731-2	7H00914	99	96	82	70	61	1	2	2	0	1
9731-2	7H00915	98	95	82	71	61	0	1	2	1	1
9731-2	7H00916	98	96	82	70	60	0	2	2	0	0
9731-2	7H00917	99	96	82	71	62	1	2	2	1	2

Oximeter Monitor: CSI 504
Monitor S/N: 291108486

Finger Probe Model: P862RA
BioTek Simulator S/N: 125672

9731-1	7I01432	98	96	81	70	61	0	2	1	0	1
9731-1	7I01433	98	96	82	71	62	0	2	2	1	2
9731-1	7I01434	98	96	82	72	61	0	2	2	2	1
9731-1	7I01435	98	96	82	72	61	0	2	2	2	1
9731-1	7I01436	98	96	80	69	59	0	2	0	-1	-1
9731-1	7I01437	98	96	82	72	62	0	2	2	2	2
9731-1	7I01438	98	94	80	70	60	0	0	0	0	0
9731-1	7I01439	98	96	82	71	61	0	2	2	1	1
9731-1	7I01440	98	96	82	72	61	0	2	2	2	0
9731-1	7I01441	99	96	81	70	60	1	2	1	0	-1
9731-1	7I01442	98	95	80	69	59	0	1	0	-1	2
9731-1	7I01443	99	96	83	72	62	1	2	3	2	0
9731-1	7I01444	98	95	81	70	60	0	1	1	0	-1
9731-1	7I01445	99	96	81	71	59	1	2	1	1	2
9731-1	7I01446	98	96	82	72	62	0	2	2	2	2
9731-1	7I01447	98	96	82	71	62	0	2	2	1	2
9731-1	7I01448	99	96	80	70	62	1	2	0	0	1
9731-1	7I01449	98	96	80	70	61	0	2	0	0	-1
9731-1	7I01450	98	96	80	69	59	0	2	0	-1	0
9731-1	7I01451	99	96	80	70	60	1	2	0	0	1
9731-1	7I01452	98	95	82	72	61	0	1	2	2	2
9731-1	7I01453	98	96	82	71	62	0	2	2	1	0
9731-1	7I01454	98	95	80	69	60	0	1	0	-1	0
9731-1	7I01455	98	96	80	72	60	0	2	0	2	0
9731-1	7I01456	98	96	80	70	60	0	2	0	0	0
9731-1	7I01457	96	95	80	70	60	-2	1	0	0	1
9731-1	7I01458	96	94	82	72	61	-2	0	2	2	1
9731-1	7I01459	95	94	82	71	61	-3	0	2	1	1
9731-1	7I01460	95	94	80	71	61	-3	0	0	1	2
9731-1	7I01461	95	94	82	72	62	-3	0	2	2	2
9731-1	7I01462	95	94	80	70	60	-3	0	0	0	0
9731-1	7I01463	96	94	81	71	62	-2	0	1	1	2
9731-1	7I01464	95	94	82	72	61	-3	0	2	2	1
9731-1	7I01465	95	94	81	70	60	-3	0	1	0	0
9731-1	7I01466	96	94	82	71	61	-2	0	2	1	1
9731-1	7I01467	96	94	83	72	61	-2	0	3	2	1
9731-1	7I01468	96	94	82	72	62	-2	0	2	2	2

Tester: [Signature]

Date: 9/24/97

Q.A.: [Signature]

Date: 9/24/97

A6.3 P862RA Test Results

Oximeter Monitor: CSI 504

Monitor S/N: 291108486

Pulse Rate Testing

Finger Probe Model: P862RA

BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9731-2	7H00880	244	149	94	79	49	-6	-1	-1	-1	-1
9731-2	7H00881	244	149	94	79	49	-6	-1	-1	-1	-1
9731-2	7H00882	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00883	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00884	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00885	244	149	94	79	49	-6	-1	-1	-1	-1
9731-2	7H00886	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00887	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00888	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00889	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00890	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00891	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00893	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00894	243	149	94	80	49	-7	-1	-1	0	-1
9731-2	7H00895	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00897	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00898	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00899	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00900	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00901	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00902	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00903	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00904	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00905	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00906	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00907	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00908	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00909	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00910	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00911	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00912	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00913	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00914	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00915	243	149	94	79	49	-7	-1	-1	-1	-1
9731-2	7H00916	242	149	94	79	49	-8	-1	-1	-1	-1
9731-2	7H00917	242	149	94	79	49	-8	-1	-1	-1	-1

Oximeter Monitor: CSI 504
Monitor S/N: 291108486

Finger Probe Model: P862RA
BioTek Simulator S/N: 125672

9731-1	7I01432	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01433	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01434	241	149	94	79	49	-9	-1	-1	-1	-1
9731-1	7I01435	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01436	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01437	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01438	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01439	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01440	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01441	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01442	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01443	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01444	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01445	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01446	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01447	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01448	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01449	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01450	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01451	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01452	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01453	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01454	244	149	94	79	49	-6	-1	-1	-1	-1
9731-1	7I01455	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01456	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01457	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01458	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01459	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01460	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01461	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01462	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01463	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01464	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01465	242	149	94	79	49	-8	-1	-1	-1	-1
9731-1	7I01466	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01467	243	149	94	79	49	-7	-1	-1	-1	-1
9731-1	7I01468	252	149	94	79	49	2	-1	-1	-1	-1

Tester: Peter J. Cha

Date: 9/26/97

Q.A.: James B.

Date: 9/26/97

A6.4 Test Results P867RA

Oximeter Monitor: Ohmeda 3740

Monitor S/N: FMQU02027

SpO₂ Testing

Finger Probe Model: P867RA

BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9731-4	7H00978	98	94	80	70	60	0	0	0	0	0
9731-4	7H00979	98	94	80	70	60	0	0	0	0	0
9731-4	7H00980	98	94	80	70	60	0	0	0	0	0
9731-4	7H00981	98	94	80	70	60	0	0	0	0	0
9731-4	7H00982	98	94	80	70	60	0	0	0	0	0
9731-4	7H00983	98	94	80	70	60	0	0	0	0	0
9731-4	7H00984	98	94	80	70	60	0	0	0	0	0
9731-4	7H00985	98	94	80	70	60	0	0	0	0	0
9731-4	7H00986	98	94	80	70	60	0	0	0	0	0
9731-4	7H00987	98	94	80	70	60	0	0	0	0	0
9731-4	7H00988	98	94	80	70	60	0	0	0	0	0
9731-4	7H00989	98	94	80	70	60	0	0	0	0	0
9731-4	7H00990	98	94	80	70	60	0	0	0	0	0
9731-4	7H00991	98	94	80	70	60	0	0	0	0	0
9731-4	7H00992	98	94	80	70	60	0	0	0	0	0
9731-4	7H00993	98	94	80	70	60	0	0	0	0	0
9731-4	7H00994	98	94	80	70	60	0	0	0	0	0
9731-4	7H00995	98	94	80	70	60	0	0	0	0	0
9731-4	7H00996	98	94	80	70	60	0	0	0	0	0
9731-4	7H00997	98	94	80	70	60	0	0	0	0	0
9731-4	7H00998	98	94	80	70	60	0	0	0	0	0
9731-4	7H00999	98	94	80	70	60	0	0	0	0	0
9731-4	7H01000	98	94	80	70	60	0	0	0	0	0
9731-4	7H01001	98	94	80	70	60	0	0	0	0	0
9731-4	7H01002	99	94	80	70	60	1	0	0	0	0
9731-4	7H01003	97	94	80	70	60	-1	0	0	0	0
9731-4	7H01004	97	94	80	70	60	-1	0	0	0	0
9731-4	7H01005	98	94	80	70	60	0	0	0	0	0
9731-4	7H01006	98	94	80	70	60	0	0	0	0	0
9731-4	7H01007	98	94	80	70	60	0	0	0	0	0
9731-4	7H01008	98	94	80	70	60	0	0	0	0	0
9731-4	7H01009	98	94	80	70	60	0	0	0	0	0
9731-4	7H01010	98	94	80	70	60	0	0	0	0	0
9731-4	7H01011	99	94	80	70	60	1	0	0	0	0
9731-4	7H01012	98	94	80	70	60	0	0	0	0	0
9731-4	7H01013	98	94	80	70	60	0	0	0	0	0
9731-4	7H01014	98	94	80	70	60	0	0	0	0	0
9731-4	7H01015	98	94	80	70	60	0	0	0	0	0
9731-4	7H01016	99	94	80	70	60	1	0	0	0	0
9731-4	7H01017	98	94	80	70	60	0	0	0	0	0
9731-4	7H01018	98	94	80	70	60	0	0	0	0	0
9731-4	7H01019	99	94	80	70	60	1	0	0	0	0
9731-4	7H01020	98	94	80	70	60	0	0	0	0	0

Oximeter Monitor: Ohmeda 3740
Monitor S/N: FMQU02027

Finger Probe Model: P867RA
BioTek Simulator S/N: 125672

9731-4	7H01021	98	94	80	70	60	0	0	0	0	0
9731-4	7H01022	99	94	80	70	60	1	0	0	0	0
9731-4	7H01023	99	94	80	70	60	1	0	0	0	0
9731-4	7H01024	98	94	80	70	60	0	0	0	0	0
9731-4	7H01025	98	94	80	70	60	0	0	0	0	0
9731-4	7H01026	98	94	80	70	60	0	0	0	0	0
9731-4	7H01027	99	94	80	70	60	1	0	0	0	0
9731-2	7H01028	98	94	80	70	60	0	0	0	0	0
9731-2	7H01029	99	94	80	70	60	1	0	0	0	0
9731-2	7H01030	99	94	80	70	60	1	0	0	0	0
9731-2	7H01031	99	94	80	70	60	1	0	0	0	0
9731-2	7H01032	98	94	80	70	60	0	0	0	0	0
9731-2	7H01033	98	94	80	70	60	0	0	0	0	0
9731-2	7H01034	99	94	80	70	60	1	0	0	0	0
9731-2	7H01035	99	94	80	70	60	1	0	0	0	0
9731-2	7H01036	98	94	80	70	60	0	0	0	0	0
9731-2	7H01037	97	94	80	70	60	-1	0	0	0	0
9731-2	7H01038	99	94	80	70	60	1	0	0	0	0
9731-2	7H01039	98	94	80	70	60	0	0	0	0	0
9731-2	7H01040	98	94	80	70	60	0	0	0	0	0
9731-2	7H01041	98	94	80	70	60	0	0	0	0	0
9731-2	7H01042	98	94	80	70	60	0	0	0	0	0
9731-2	7H01043	99	94	80	70	60	1	0	0	0	0
9731-2	7H01044	99	94	80	70	60	1	0	0	0	0
9731-2	7H01045	99	94	80	70	60	1	0	0	0	0
9731-2	7H01046	98	94	80	70	60	0	0	0	0	0
9731-2	7H01047	99	94	80	70	60	1	0	0	0	0
9731-2	7H01048	98	94	80	70	60	0	0	0	0	0
9731-2	7H01049	99	94	80	70	60	1	0	0	0	0
9731-2	7H01050	98	94	80	70	60	0	0	0	0	0
9731-2	7H01051	98	94	80	70	60	0	0	0	0	0
9731-2	7H01052	98	94	80	70	60	0	0	0	0	0
9731-5	7H01053	98	94	80	70	60	0	0	0	0	0
9731-5	7H01054	98	94	80	70	60	0	0	0	0	0
9731-5	7H01055	98	94	80	70	60	0	0	0	0	0
9731-5	7H01056	98	94	80	70	60	0	0	0	0	0
9731-5	7H01057	98	94	80	70	60	0	0	0	0	0
9731-5	7H01058	98	94	80	70	60	0	0	0	0	0
9731-5	7H01059	98	94	80	70	60	0	0	0	0	0
9731-5	7H01060	98	94	80	70	60	0	0	0	0	0

A6.4 P867RA Test Results

Oximeter Monitor: Ohmeda 3740

Monitor S/N: FMQU02027

Pulse Rate Testing

Finger Probe Model: P867RA

BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9731-4	7H00978	250	150	95	80	50	0	0	0	0	0
9731-4	7H00979	250	150	95	80	50	0	0	0	0	0
9731-4	7H00980	250	150	95	80	50	0	0	0	0	0
9731-4	7H00981	250	150	95	80	50	0	0	0	0	0
9731-4	7H00982	250	150	95	80	50	0	0	0	0	0
9731-4	7H00983	250	150	95	80	50	0	0	0	0	0
9731-4	7H00984	250	150	95	80	50	0	0	0	0	0
9731-4	7H00985	250	150	95	80	50	0	0	0	0	0
9731-4	7H00986	250	150	95	80	50	0	0	0	0	0
9731-4	7H00987	250	150	95	80	50	0	0	0	0	0
9731-4	7H00988	250	150	95	80	50	0	0	0	0	0
9731-4	7H00989	250	150	95	80	50	0	0	0	0	0
9731-4	7H00990	252	150	95	80	50	2	0	0	0	0
9731-4	7H00991	250	150	95	80	50	0	0	0	0	0
9731-4	7H00992	250	150	95	80	50	0	0	0	0	0
9731-4	7H00993	250	150	95	80	50	0	0	0	0	0
9731-4	7H00994	250	150	95	80	50	0	0	0	0	0
9731-4	7H00995	250	150	95	80	50	0	0	0	0	0
9731-4	7H00996	250	150	95	80	50	0	0	0	0	0
9731-4	7H00997	250	150	95	80	50	0	0	0	0	0
9731-4	7H00998	250	150	95	80	50	0	0	0	0	0
9731-4	7H00999	250	150	95	80	50	0	0	0	0	0
9731-4	7H01000	250	150	95	80	50	0	0	0	0	0
9731-4	7H01001	250	150	95	80	50	0	0	0	0	0
9731-4	7H01002	250	150	95	80	50	0	0	0	0	0
9731-4	7H01003	250	150	95	80	50	0	0	0	0	0
9731-4	7H01004	250	150	95	80	50	0	0	0	0	0
9731-4	7H01005	251	150	95	80	50	1	0	0	0	0
9731-4	7H01006	250	150	95	80	50	0	0	0	0	0
9731-4	7H01007	250	150	95	80	50	0	0	0	0	0
9731-4	7H01008	250	150	95	80	50	0	0	0	0	0
9731-4	7H01009	250	150	95	80	50	0	0	0	0	0
9731-4	7H01010	250	150	95	80	50	0	0	0	0	0
9731-4	7H01011	250	150	95	80	50	0	0	0	0	0
9731-4	7H01012	250	150	95	80	50	0	0	0	0	0
9731-4	7H01013	250	150	95	80	50	0	0	0	0	0
9731-4	7H01014	250	150	95	80	50	0	0	0	0	0
9731-4	7H01015	250	150	95	80	50	0	0	0	0	0
9731-4	7H01016	252	150	95	80	50	2	0	0	0	0
9731-4	7H01017	252	150	95	80	50	2	0	0	0	0
9731-4	7H01018	250	150	95	80	50	0	0	0	0	0
9731-4	7H01019	250	150	95	80	50	0	0	0	0	0
9731-4	7H01020	250	150	95	80	50	0	0	0	0	0

Oximeter Monitor: Ohmeda 3740
Monitor S/N: FMQU02027

Finger Probe Model: P867RA
BioTek Simulator S/N: 125672

9731-4	7H01021	250	150	95	80	50	0	0	0	0	0
9731-4	7H01022	250	150	95	80	50	0	0	0	0	0
9731-4	7H01023	250	150	95	80	50	0	0	0	0	0
9731-4	7H01024	250	150	95	80	50	0	0	0	0	0
9731-4	7H01025	250	150	95	80	50	0	0	0	0	0
9731-4	7H01026	250	150	95	80	50	0	0	0	0	0
9731-4	7H01027	250	150	95	80	50	0	0	0	0	0
9731-2	7H01028	250	150	95	80	50	0	0	0	0	0
9731-2	7H01029	250	150	95	80	50	0	0	0	0	0
9731-2	7H01030	250	150	95	80	50	0	0	0	0	0
9731-2	7H01031	250	150	95	80	50	0	0	0	0	0
9731-2	7H01032	250	150	95	80	50	0	0	0	0	0
9731-2	7H01033	250	150	95	80	50	0	0	0	0	0
9731-2	7H01034	250	150	95	80	50	0	0	0	0	0
9731-2	7H01035	250	150	95	80	50	0	0	0	0	0
9731-2	7H01036	250	150	95	80	50	0	0	0	0	0
9731-2	7H01037	250	150	95	80	50	0	0	0	0	0
9731-2	7H01038	250	150	95	80	50	0	0	0	0	0
9731-2	7H01039	250	150	95	80	50	0	0	0	0	0
9731-2	7H01040	250	150	95	80	50	0	0	0	0	0
9731-2	7H01041	250	150	95	80	50	0	0	0	0	0
9731-2	7H01042	250	150	95	80	50	0	0	0	0	0
9731-2	7H01043	250	150	95	80	50	0	0	0	0	0
9731-2	7H01044	250	150	95	80	50	0	0	0	0	0
9731-2	7H01045	250	150	95	80	50	0	0	0	0	0
9731-2	7H01046	250	150	95	80	50	0	0	0	0	0
9731-2	7H01047	250	150	95	80	50	0	0	0	0	0
9731-2	7H01048	250	150	95	80	50	0	0	0	0	0
9731-2	7H01049	250	150	95	80	50	0	0	0	0	0
9731-2	7H01050	250	150	95	80	50	0	0	0	0	0
9731-2	7H01051	250	150	95	80	50	0	0	0	0	0
9731-2	7H01052	250	150	95	80	50	0	0	0	0	0
9731-5	7H01053	252	150	95	80	50	2	0	0	0	0
9731-5	7H01054	251	150	95	80	50	1	0	0	0	0
9731-5	7H01055	250	150	95	80	50	0	0	0	0	0
9731-5	7H01056	250	150	95	80	50	0	0	0	0	0
9731-5	7H01057	250	150	95	80	50	0	0	0	0	0
9731-5	7H01058	251	150	95	80	50	1	0	0	0	0
9731-5	7H01059	252	150	95	80	50	2	0	0	0	0
9731-5	7H01060	250	150	95	80	50	0	0	0	0	0

Oximeter Monitor: Ohmeda 3740
Monitor S/N: FMOU02027

Finger Probe Model: P867RA
BioTek Simulator S/N: 125672

9731-5	7H01061	98	94	80	70	60	0	0	0	0	0
9731-5	7H01062	98	94	80	70	60	0	0	0	0	0
9731-5	7H01063	98	94	80	70	60	0	0	0	0	0
9731-5	7H01064	98	94	80	70	60	0	0	0	0	0
9731-5	7H01065	99	94	80	70	60	1	0	0	0	0
9731-5	7H01066	98	94	80	70	60	0	0	0	0	0
9731-5	7H01067	98	94	80	70	60	0	0	0	0	0
9731-5	7H01068	98	94	80	70	60	0	0	0	0	0
9731-5	7H01069	98	94	80	70	60	0	0	0	0	0
9731-5	7H01070	98	94	80	70	60	0	0	0	0	0
9731-5	7H01071	98	94	80	70	60	0	0	0	0	0
9731-5	7H01072	98	94	80	70	60	0	0	0	0	0
9731-5	7H01073	98	94	80	70	60	0	0	0	0	0
9731-5	7H01074	98	94	80	70	60	0	0	0	0	0
9731-5	7H01075	98	94	80	70	60	0	0	0	0	0
9731-5	7H01076	98	94	80	70	60	0	0	0	0	0

Tester: 

Date: 9/24/97

Q.A.: 

Date: 9/24/97

Oximeter Monitor: Ohmeda 3740
Monitor S/N: FMQU02027

Finger Probe Model: P867RA
BioTek Simulator S/N: 125672

9731-5	7H01061	252	150	95	80	50	2	0	0	0	0
9731-5	7H01062	251	150	95	80	50	1	0	0	0	0
9731-5	7H01063	250	150	95	80	50	0	0	0	0	0
9731-5	7H01064	251	150	95	80	50	1	0	0	0	0
9731-5	7H01065	250	150	95	80	50	0	0	0	0	0
9731-5	7H01066	252	150	95	80	50	2	0	0	0	0
9731-5	7H01067	250	150	95	80	50	0	0	0	0	0
9731-5	7H01068	250	150	95	80	50	0	0	0	0	0
9731-5	7H01069	250	150	95	80	50	0	0	0	0	0
9731-5	7H01070	252	150	95	80	50	2	0	0	0	0
9731-5	7H01071	250	150	95	80	50	0	0	0	0	0
9731-5	7H01072	250	150	95	80	50	0	0	0	0	0
9731-5	7H01073	250	150	95	80	50	0	0	0	0	0
9731-5	7H01074	251	150	95	80	50	1	0	0	0	0
9731-5	7H01075	250	150	95	80	50	0	0	0	0	0
9731-5	7H01076	251	150	95	80	50	1	0	0	0	0

Tester:

[Signature]

Date:

9/26/97

Q.A.:

[Signature]

Date:

9/26/97

A6.5 P876RA Test Results
Oximeter Monitor: Novamatrix 515A
 Monitor S/N: 69-1010A

SpO₂ Testing
Finger Probe Model: P876RA
 BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9731-1	7101380	98	94	80	70	59	0	0	0	0	-1
9731-1	7101382	98	94	80	70	60	0	0	0	0	0
9731-1	7101385	98	94	80	70	60	0	0	0	0	0
9731-1	7101386	98	94	80	70	59	0	0	0	0	-1
9731-1	7101387	98	94	80	70	60	0	0	0	0	0
9731-1	7101388	98	94	80	70	60	0	0	0	0	0
9731-1	7101389	98	94	80	70	60	0	0	0	0	0
9731-1	7101390	98	94	80	70	60	0	0	0	0	0
9731-1	7101391	98	94	80	70	60	0	0	0	0	0
9731-1	7101392	98	94	80	70	59	0	0	0	0	-1
9731-1	7101393	98	94	80	70	60	0	0	0	0	0
9731-1	7101394	98	94	80	70	60	0	0	0	0	0
9731-1	7101395	98	94	80	70	60	0	0	0	0	0
9731-1	7101396	98	94	80	70	60	0	0	0	0	0
9731-1	7101397	98	94	80	70	60	0	0	0	0	0
9731-1	7101398	98	94	80	70	60	0	0	0	0	0
9731-1	7101399	98	94	80	70	59	0	0	0	0	-1
9731-1	7101400	98	94	80	70	59	0	0	0	0	-1
9731-1	7101401	98	94	80	70	60	0	0	0	0	0
9731-1	7101402	98	94	80	70	60	0	0	0	0	0
9731-1	7101403	98	94	80	70	59	0	0	0	0	-1
9731-1	7101404	98	94	80	70	59	0	0	0	0	-1
9731-1	7101405	98	94	80	70	59	0	0	0	0	-1
9731-1	7101406	98	94	80	70	59	0	0	0	0	-1
9731-1	7101407	98	94	80	70	60	0	0	0	0	0
9731-1	7101408	98	94	80	70	60	0	0	0	0	0
9731-1	7101409	98	94	80	70	60	0	0	0	0	0
9731-1	7101410	98	94	80	70	60	0	0	0	0	0
9731-1	7101411	98	94	80	70	60	0	0	0	0	0
9731-1	7101412	98	94	80	70	59	0	0	0	0	-1
9731-1	7101413	98	94	80	70	60	0	0	0	0	0
9731-1	7101414	98	94	80	70	59	0	0	0	0	-1
9731-1	7101415	98	94	80	70	59	0	0	0	0	-1
9731-1	7101416	98	94	80	70	60	0	0	0	0	0
9731-1	7101417	98	94	80	70	60	0	0	0	0	0
9731-1	7101418	98	94	80	70	60	0	0	0	0	0
9731-1	7101419	98	94	80	70	59	0	0	0	0	-1
9731-1	7101420	98	94	80	70	60	0	0	0	0	0
9731-1	7101421	98	94	80	70	59	0	0	0	0	-1
9731-1	7101422	98	94	80	70	60	0	0	0	0	0
9731-1	7101423	98	94	80	70	60	0	0	0	0	0
9731-1	7101424	98	94	80	70	60	0	0	0	0	0
9731-1	7101425	98	94	80	70	60	0	0	0	0	0
9731-1	7101426	98	94	80	70	59	0	0	0	0	-1
9731-1	7101427	98	94	80	70	59	0	0	0	0	-1
9731-1	7101428	98	94	80	70	60	0	0	0	0	0
9731-1	7101429	98	94	80	70	60	0	0	0	0	0
9731-1	7101430	98	94	80	70	60	0	0	0	0	0
9731-1	7101431	98	94	80	70	60	0	0	0	0	0

Tester: *[Signature]*

Date: 9/26/97

Q.A.: *[Signature]*

Date: 9/26/97

A6.5 P876RA Test Results

Oximeter Monitor: Novamatrix 515A

Monitor S/N: 69-1010A

Pulse Rate Testing

Finger Probe Model: P876RA

BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9731-1	7101380	199	160	100	80	60	-1	0	0	0	0
9731-1	7101382	200	160	100	80	60	0	0	0	0	0
9731-1	7101385	200	160	100	80	60	0	0	0	0	0
9731-1	7101386	200	160	100	80	60	0	0	0	0	0
9731-1	7101387	200	160	100	80	60	0	0	0	0	0
9731-1	7101388	199	160	100	80	60	-1	0	0	0	0
9731-1	7101389	200	160	100	80	60	0	0	0	0	0
9731-1	7101390	199	160	100	80	60	-1	0	0	0	0
9731-1	7101391	199	160	100	80	60	-1	0	0	0	0
9731-1	7101392	199	160	100	80	60	-1	0	0	0	0
9731-1	7101393	200	160	100	80	60	0	0	0	0	0
9731-1	7101394	200	160	100	80	60	0	0	0	0	0
9731-1	7101395	200	160	100	80	60	0	0	0	0	0
9731-1	7101396	200	160	100	80	60	0	0	0	0	0
9731-1	7101397	200	160	100	80	60	0	0	0	0	0
9731-1	7101398	199	160	100	80	60	-1	0	0	0	0
9731-1	7101399	200	160	100	80	60	0	0	0	0	0
9731-1	7101400	200	160	100	80	60	0	0	0	0	0
9731-1	7101401	200	160	100	80	60	0	0	0	0	0
9731-1	7101402	200	160	100	80	60	0	0	0	0	0
9731-1	7101403	199	160	100	80	60	-1	0	0	0	0
9731-1	7101404	200	160	100	80	60	0	0	0	0	0
9731-1	7101405	199	160	100	80	60	-1	0	0	0	0
9731-1	7101406	200	160	100	80	60	0	0	0	0	0
9731-1	7101407	200	160	100	80	60	0	0	0	0	0
9731-1	7101408	200	160	100	80	60	0	0	0	0	0
9731-1	7101409	199	160	100	80	60	-1	0	0	0	0
9731-1	7101410	200	160	100	80	60	0	0	0	0	0
9731-1	7101411	200	160	100	80	60	0	0	0	0	0
9731-1	7101412	200	160	100	80	60	0	0	0	0	0
9731-1	7101413	200	160	100	80	60	0	0	0	0	0
9731-1	7101414	200	160	100	80	60	0	0	0	0	0
9731-1	7101415	200	160	100	80	60	0	0	0	0	0
9731-1	7101416	199	160	100	80	60	-1	0	0	0	0
9731-1	7101417	200	160	100	80	60	0	0	0	0	0
9731-1	7101418	200	160	100	80	60	0	0	0	0	0
9731-1	7101419	200	160	100	80	60	0	0	0	0	0
9731-1	7101420	200	160	100	80	60	0	0	0	0	0
9731-1	7101421	200	160	100	80	60	0	0	0	0	0
9731-1	7101422	200	160	100	80	60	0	0	0	0	0
9731-1	7101423	200	160	100	80	60	0	0	0	0	0
9731-1	7101424	199	160	100	80	60	-1	0	0	0	0
9731-1	7101425	200	160	100	80	60	0	0	0	0	0
9731-1	7101426	200	160	100	80	60	0	0	0	0	0
9731-1	7101427	200	160	100	80	60	0	0	0	0	0
9731-1	7101428	200	160	100	80	60	0	0	0	0	0
9731-1	7101429	200	160	100	80	60	0	0	0	0	0
9731-1	7101430	199	160	100	80	60	-1	0	0	0	0
9731-1	7101431	200	160	100	80	60	0	0	0	0	0

Tester: *[Signature]*

Date: 7/26/97

Q.A.: *[Signature]*

Date: 7/26/97

A6.1 Test Results P856RA
Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

SpO₂ Testing
Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		98	94	80	70	60	98	94	80	70	60
9732-1	7H00001	98	94	80	70	60	0	0	0	0	0
9732-1	7H00002	98	94	80	70	60	0	0	0	0	0
9732-1	7H00003	98	94	80	70	60	0	0	0	0	0
9732-1	7H00004	98	94	80	70	60	0	0	0	0	0
9732-1	7H00005	98	94	80	70	60	0	0	0	0	0
9732-1	7H00006	98	94	80	70	60	0	0	0	0	0
9732-1	7H00007	98	94	80	70	60	0	0	0	0	0
9732-1	7H00008	98	94	80	69	60	0	0	0	-1	0
9732-1	7H00009	98	94	80	70	60	0	0	0	0	0
9732-1	7H00010	98	94	80	70	60	0	0	0	0	0
9732-1	7H00011	98	94	80	70	60	0	0	0	0	0
9732-1	7H00012	98	94	80	70	60	0	0	0	0	0
9732-1	7H00013	98	94	80	70	60	0	0	0	0	0
9732-1	7H00014	98	94	80	70	60	0	0	0	0	0
9732-1	7H00015	98	94	80	70	60	0	0	0	0	0
9732-1	7H00016	98	94	80	70	60	0	0	0	0	0
9732-1	7H00017	98	94	80	70	60	0	0	0	0	0
9732-1	7H00018	98	94	80	70	60	0	0	0	0	0
9732-1	7H00019	98	94	80	70	60	0	0	0	0	0
9732-1	7H00020	98	94	80	70	60	0	0	0	0	0
9732-1	7H00021	98	94	80	70	60	0	0	0	0	0
9732-1	7H00022	98	94	80	70	60	0	0	0	0	0
9732-1	7H00023	98	94	80	70	60	0	0	0	0	0
9732-1	7H00024	98	94	80	70	60	0	0	0	0	0
9732-1	7H00025	98	94	80	70	60	0	0	0	0	0
9732-1	7H00026	98	94	80	70	60	0	0	0	0	0
9732-1	7H00027	98	94	80	70	60	0	0	0	0	0
9732-1	7H00028	98	94	80	70	60	0	0	0	0	0
9732-1	7H00029	98	94	80	70	60	0	0	0	0	0
9732-1	7H00030	98	94	80	70	60	0	0	0	0	0
9732-1	7H00031	98	94	80	70	60	0	0	0	0	0
9732-1	7H00032	98	94	80	70	60	0	0	0	0	0
9732-1	7H00033	98	94	80	70	60	0	0	0	0	0
9732-1	7H00034	98	94	80	70	60	0	0	0	0	0
9732-1	7H00035	98	94	80	70	60	0	0	0	0	0
9732-1	7H00036	98	94	80	70	60	0	0	0	0	0
9732-1	7H00037	98	94	80	70	60	0	0	0	0	0
9732-1	7H00038	98	94	80	70	60	0	0	0	0	0
9732-1	7H00039	98	94	80	70	60	0	0	0	0	0
9732-1	7H00040	98	94	80	70	60	0	0	0	0	0
9732-1	7H00041	98	94	80	70	60	0	0	0	0	0
9732-1	7H00042	98	94	80	70	60	0	0	0	0	0
9732-1	7H00043	98	94	80	70	60	0	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-1	7H00044	98	94	80	70	60	0	0	0	0	0
9732-1	7H00045	98	94	80	70	60	0	0	0	0	0
9732-1	7H00046	98	94	80	70	60	0	0	0	0	0
9732-1	7H00047	98	94	80	70	60	0	0	0	0	0
9732-1	7H00048	98	94	80	70	60	0	0	0	0	0
9732-1	7H00049	98	94	80	70	60	0	0	0	0	0
9732-1	7H00050	98	94	80	70	60	0	0	0	0	0
9732-2	7H00051	98	94	80	70	60	0	0	0	0	0
9732-2	7H00052	98	94	80	70	60	0	0	0	0	0
9732-2	7H00053	98	94	80	70	60	0	0	0	0	0
9732-2	7H00054	98	94	80	70	60	0	0	0	0	0
9732-2	7H00055	98	94	80	70	60	0	0	0	0	0
9732-2	7H00056	98	94	80	70	60	0	0	0	0	0
9732-2	7H00057	98	94	80	70	60	0	0	0	0	0
9732-2	7H00058	98	94	80	70	60	0	0	0	0	0
9732-2	7H00059	98	94	80	70	60	0	0	0	0	0
9732-2	7H00060	98	94	80	70	60	0	0	0	0	0
9732-2	7H00061	98	94	80	70	60	0	0	0	0	0
9732-2	7H00062	98	94	80	70	60	0	0	0	0	0
9732-2	7H00063	98	94	80	70	60	0	0	0	0	0
9732-2	7H00064	98	94	80	70	60	0	0	0	0	0
9732-2	7H00065	98	94	80	70	60	0	0	0	0	0
9732-2	7H00066	98	94	80	70	60	0	0	0	0	0
9732-2	7H00067	98	94	80	70	60	0	0	0	0	0
9732-2	7H00068	98	94	80	70	60	0	0	0	0	0
9732-2	7H00069	98	94	80	70	60	0	0	0	0	0
9732-2	7H00070	98	94	80	70	60	0	0	0	0	0
9732-2	7H00071	98	94	80	70	60	0	0	0	0	0
9732-2	7H00072	98	94	80	70	60	0	0	0	0	0
9732-2	7H00073	98	94	80	70	60	0	0	0	0	0
9732-2	7H00074	98	94	80	70	60	0	0	0	0	0
9732-2	7H00075	98	94	80	70	60	0	0	0	0	0
9732-2	7H00076	98	94	80	70	60	0	0	0	0	0
9732-2	7H00077	98	94	80	70	60	0	0	0	0	0
9732-2	7H00078	98	94	80	70	60	0	0	0	0	0
9732-2	7H00079	98	94	80	70	60	0	0	0	0	0
9732-2	7H00080	98	94	80	70	60	0	0	0	0	0
9732-2	7H00081	98	94	80	70	60	0	0	0	0	0
9732-2	7H00082	98	94	80	70	60	0	0	0	0	0
9732-2	7H00083	98	94	80	70	60	0	0	0	0	0
9732-2	7H00084	98	94	80	70	60	0	0	0	0	0
9732-2	7H00085	98	94	80	70	60	0	0	0	0	0
9732-2	7H00086	98	94	80	70	60	0	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	98	94	80	70	60	0	0	0	0	0
9732-2	7H00088	98	94	80	70	60	0	0	0	0	0
9732-2	7H00089	98	94	80	70	60	0	0	0	0	0
9732-2	7H00090	98	94	80	70	60	0	0	0	0	0
9732-2	7H00091	98	94	80	70	60	0	0	0	0	0
9732-2	7H00092	98	94	80	70	60	0	0	0	0	0
9732-2	7H00093	98	94	80	70	60	0	0	0	0	0
9732-2	7H00094	98	94	80	70	60	0	0	0	0	0
9732-2	7H00095	98	94	80	70	60	0	0	0	0	0
9732-2	7H00096	98	94	80	70	60	0	0	0	0	0
9732-2	7H00097	98	94	80	70	60	0	0	0	0	0
9732-2	7H00098	98	94	80	70	60	0	0	0	0	0
9732-2	7H00099	98	94	80	70	60	0	0	0	0	0

Tester: [Signature]

Date: 9/23/97
10/10/97

Q.A.: [Signature]

Date: 9/23/97

A6.1 Test Results P856RA

Pulse Rate Testing

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

Lot Number	Serial Number	Test Results					Bias Error				
		250	150	95	80	50	250	150	95	80	50
9732-1	7H00001	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00002	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00003	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00004	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00005	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00006	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00007	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00008	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00009	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00010	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00011	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00012	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00013	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00014	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00015	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00016	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00017	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00018	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00019	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00020	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00021	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00022	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00023	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00024	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00025	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00026	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00027	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00028	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00029	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00030	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00031	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00032	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00033	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00034	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00035	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00036	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00037	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00038	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00039	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00040	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00041	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00042	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00043	249	150	95	80	50	-1	0	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-1	7H00044	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00045	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00046	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00047	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00048	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00049	249	150	95	80	50	-1	0	0	0	0
9732-1	7H00050	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00051	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00052	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00053	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00054	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00055	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00056	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00057	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00058	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00059	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00060	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00061	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00062	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00063	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00064	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00065	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00066	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00067	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00068	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00069	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00070	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00071	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00072	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00073	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00074	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00075	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00076	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00077	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00078	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00079	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00080	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00081	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00082	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00083	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00084	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00085	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00086	249	149	95	80	50	-1	-1	0	0	0

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00088	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00089	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00090	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00091	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00092	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00093	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00094	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00095	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00096	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00097	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00098	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00099	249	150	95	80	50	-1	0	0	0	0

Tester: [Signature]

Date: 9/23/97

Q.A.: [Signature]

Date: 9/23/97

Pulse Oximeter Monitor: Nellcor N200
Monitor S/N: 20619220

Finger Probe Model: P856RA
BioTek Simulator S/N: 125672

9732-2	7H00087	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00088	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00089	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00090	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00091	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00092	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00093	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00094	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00095	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00096	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00097	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00098	249	150	95	80	50	-1	0	0	0	0
9732-2	7H00099	249	150	95	80	50	-1	0	0	0	0

Tester: [Signature]

Date: 9/23/97

Q.A.: [Signature]

Date: 9/23/97