

Service Repair Sheet SRS68390

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Carol Hoy

Royal Veterinary College

Senior Anaesthesia Nurse

01707666366

01707666366

00012213

choy@rvc.ac.uk

21/Feb/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC
Date: 21/02/23

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 14/Feb/2023 Kate Griffiths

14/Feb/2023 Kate Griffiths

Would like to send in 4410540 VM2500S, s/n 81040139 "the capnograph indicator light is red and not reading even after changing consumables". She also says "in the past we have arranged a service exchange with you"

21/Feb/2023 Robert Connor

Received 1 x VM2500-S-Vet s/n 81040139, with li-poly battery

Ready For quote

Green 6423

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35677	Capnograph	4410540	81040139	N

4440000 x 1/2hr @ £45
S/N, SRS, SRN

4420504 x 1 @ £1,049.75
SRS, SRN

4430010 x 1 @ £2.76.
SRS, SRN

20

UPS x 1 @ £12.



Queen Mother Hospital for Animals

Royal Veterinary College

As Arranged with Kate Griffiths
for assessment & possible repair of
capnograph

Reference: SR568390

With compliments

Hawkshead Lane
North Mymms
Hatfield
Hertfordshire AL9 7TA
United Kingdom
T +44 (0)1707 666366
F +44 (0)1707 649384
E qmhreception@rvc.ac.uk
W www.rvc.ac.uk

Delivered by	
	
Royal Mail	
Postage paid GB	
Franked Parcels	
	
278GB	
RETURN TO:	ROYAL VETERINARY COLLEGE
	HAWKSHEAD LANE, NORTH MYMMS
	HATFIELD
	AL9 7TA GB
NE4000256	
Post by the end of	
22.02.2023	