

| Spectrometer | Prema SPM 9001 |
|--------------------------------|---------------------------------|
| | |
| Manufacturer of probe | Epic |
| Type of probe | Finger |
| Part number | E203-01 |
| Serial number | 9760101052 |
| Cable length | 3ft |
| Connector | D type 9 pin male |
| Connects to | Monitor or extension cable |
| | |
| For use on | Nellcor and compatible monitors |
| Tested on | Nellcor N-200 |
| | |
| Red peak wavelength (nm) | 658 |
| Infra red peak wavelength (nm) | 915.4 |
| | |
| Emitter configuration | 2 lead (parallel, reverse) |
| | |
| Carried out by (name printed) | S Nixon |
| Signed | |
| Date | 31 July 1998 |