





Spectrometer	Prema SPM 9001
Part number	
Serial number	42 23031
Cable length	3 m
Connector	KPK1
Connects to	Monitor
For use on	Kontron
Tested on	Kontron 7840
Red peak wavelength	660
Infra red peak wavelength	890
Emitter configuration	2 lead
Carried out by (name printed)	S Nixon
Signed	
Date	29 July 1998



Spectrometer	Prema SPM 9001
Manufacturer of probe	Kontron
Type of probe	Ear
Part number	
Serial number	
Cable length	2.78 m
Connector	KPK1
Connects to	Monitor
For use on	Kontron
Tested on	Kontron 7840
Red peak wavelength	660
Infra red peak wavelength	880
Emitter configuration	2 lead
Carried out by (name printed)	S Nixon
Signed	
Date	28 July 1998