





Flexible Y Probe Testing DL3000

Probe Type	P976YA	Date	23/7/99
Serial Number	8F07912	Tester	RJS
DL 3000 S/N	0059997	OEM S/N	66-1065A
Resistance	Kohm	OEM Model	Novametric SIS
Cable Length	120" $\frac{1}{2}$		

DL300 setting	OEM Reading
99	
97	97
95	
90	90
85	
80	81
75	
70	
65	
60	

Visual Check	Pass	Fail
Emitter window quality	✓	
Detectors window quality	✓	
Alignment	✓	✓
Mould Quality	✓	
Serial Number	8F07912	
Colour Dot	Yellow	
Cable Length	120" $\frac{1}{2}$	
Housing Colour	black	




Flexible Y Probe Testing DL3000

Probe Type	P976YA	Date	23/7/98
Serial Number	8F07913	Tester	RST
DL 3000 S/N	0059997	OEM S/N	66-1065A
Resistance	Kohm	OEM Model	Novametric 515
Cable Length	120" $\frac{1}{2}$		

DL300 setting	OEM Reading
99	
97	97
95	
90	90
85	
80	81
75	
70	
65	
60	

Visual Check	Pass	Fail
Emitter window quality	✓	
Detectors window quality	✓	
Alignment	✓	
Mould Quality	✓	
Serial Number	8F07913	
Colour Dot	yellow	
Cable Length	120" $\frac{1}{2}$	
Housing Colour	white/black	




Flexible Y Probe Testing DL3000

Probe Type	PS 76YA	Date	23/7/98
Serial Number	8F07414	Tester	NSS
DL 3000 S/N	0034497	OEM S/N	66-10654
Resistance	Kohm	OEM Model	Neurometric SIS
Cable Length	120" $\frac{1}{2}$		

DL300 setting	OEM Reading
99	
97	917
95	
90	910
85	
80	91
75	
70	
65	
60	

Visual Check	Pass	Fail
Emitter window quality	✓	
Detectors window quality	✓	
Alignment	✓	
Mould Quality	✓	
Serial Number	8F07414	
Colour Dot	Yellow	
Cable Length	120" $\frac{1}{2}$	
Housing Colour	black	




Flexible Y Probe Testing DL3000

Probe Type	PS7674	Date	23/7/98
Serial Number	870311	Tester	1235
DL 3000 S/N	054997	OEM S/N	66-10654
Resistance	Kohm	OEM Model	Newcometrix S12
Cable Length	120" ±		

DL300 setting	OEM Reading
99	98
97	96
95	94
90	92
85	
80	90
75	88
70	
65	
60	

Visual Check	Pass	Fail
Emitter window quality	/	
Detectors window quality	/	
Alignment		/
Mould Quality	✓	
Serial Number	870311	
Colour Dot	YELLOW	
Cable Length	120"	
Housing Colour	BLACK	




Flexible Y Probe Testing DL3000

Probe Type	P8764A	Date	23/7/99
Serial Number	8F07916	Tester	RJS
DL 3000 S/N	0039997	OEM S/N	66-1065A
Resistance	Kohm	OEM Model	Novametrix SIS
Cable Length	120" $\frac{1}{2}$		

DL300 setting	OEM Reading
99	
97	17
95	
90	90
85	
80	80
75	
70	
65	
60	

Visual Check	Pass	Fail
Emitter window quality	✓	
Detectors window quality	✓	
Alignment		✓
Mould Quality	✓	
Serial Number	8F07916	
Colour Dot	yellow	
Cable Length	120" $\frac{1}{2}$	
Housing Colour	white/black	




Flexible Y Probe Testing DL3000

Probe Type	PF764A	Date	23/3/98
Serial Number	YF03913	Tester	A.S
DL 3000 S/N	0057993	OEM S/N	66-106519
Resistance	Kohm	OEM Model	NOVA MATRIX
Cable Length	120"		

DL300 setting	OEM Reading
99	98
97	96
95	95
90	90
85	84
80	80
75	74
70	69
65	64
60	59

Visual Check	Pass	Fail
Emitter window quality	✓	
Detectors window quality	✓	
Alignment		✓
Mould Quality	✓	
Serial Number	YF03913	
Colour Dot	YELLOW	
Cable Length	120"	
Housing Colour	WHITE	



Flexible Y Probe Testing DL3000

Probe Type	P8761A	Date	23/7/99
Serial Number	8F07918	Tester	RSS
DL 3000 S/N	0039997	OEM S/N	66-1065A
Resistance	Kohm	OEM Model	Novametric SIS
Cable Length	120" $\frac{1}{2}$		

DL300 setting	OEM Reading
99	
97	97
95	
90	90
85	
80	80
75	
70	
65	
60	

Visual Check	Pass	Fail
Emitter window quality	✓	
Detectors window quality	✓	
Alignment	✓	
Mould Quality	✓	
Serial Number	8F07918	
Colour Dot	yellow	
Cable Length	120" $\frac{1}{2}$	
Housing Colour	black	