

Compatible SpO₂ Probes for Spacelabs Equipment

Pulse Oximetry Sensors SpaceLabs Medical Equipment.						
Part Number	90419 90501 90651A-03	90419-01 90651A-07	90351 D/G 90465 / 66 / 67 90489 / 90651A-08	Spacelabs Extension Cable	MCI probe Number	Comments
Reusable Sensors						
369083-001 Adult Finger	♦				P875RA	7 pin Hypertronic
369083-002 Ear	♦					
369083-004 Flexible	♦					
Spacelabs Medical Sensors						
015-0130-00 Adult Finger		♦	♦		P876RA	7 pin Hypertronic
015-0133-00 Universal Reusable		♦	♦			
Nellcor Sensors - Reusable						
690-0003-00 Adult Finger			♦		P856RA	DB9
690-0004-00 Oxi A/N			♦			
690-0039 Oxi P/I			♦			
Nellcor Sensors - Disposable						
690-0001-00 Adult Finger (D-25)			♦			
690-0007-00 Peadiatric Finger (D-20)			♦			
690-0002-00 Infant (I-20)			♦			
690-0006-00 Neonatal (N-25)			♦			
690-0005-00 Adult Nasal (R-15)			♦			
Viamed one piece cable			*		PP857RA	12 pin Hypertronics

PULSE OXIMETRY SENSORS Spacelabs Medical Equipment			
PART NUMBER	90419 90301 90631A-03	90419-01 90651A-07	90351 D/G 90465/66/67 90489 90651A-08
REUSABLE SENSORS			
369083-001 Adult Finger-Reusable	•		
369083-002 Ear-Reusable	•		
369083-004 Flexible-Reusable	•		
SPACELABS MEDICAL SENSORS			
015-0130-00 Adult Finger-Reusable		•	•
015-0133-00 Universal-Reusable		•	•
NELLCOR SENSORS-REUSABLE			
690-0003-00 Adult Finger (DS100)			•
690-0004-00 OXI A/N			•
690-0039-00 OXI P/I			•
NELLCOR SENSORS-DISPOSABLE			
690-0001-00 Adult Finger (D-25)			•
690-0007-00 Pediatric Finger (D-20)			•
690-0002-00 Infant (I-20)			•
690-0006-00 Neonatal (N-25)			•
690-0005-00 Adult Nasal (R-15)			•