

 <p>Vandagraph Ltd. 15, Station Road, Crosshills, Keighley. W.Yorks BD207DT</p>		Action Required	
		Customer	
		Date Out	
		Invoice out	
		Comments	
Equipment			
Serial number			
Accessories	Serial Number		
Oxygen Sensor Type			
Power supply unit			
Tests Carried out			
Battery Check			
Soak Test			
Check Function			
Software Version			
Calibration			
Gas	Oxygen Reading	Checks with D.P.M.	
Set 100% Oxygen	Helium Reading =	Oxygen Level	mV
Air Reading		100.00%	
Set 100% Helium	Oxygen Reading =	75.00%	
Heliox/Trimix		50.00%	
Oxygen/Air Ratio (4.78)		25.00%	
Range 4.64 - 4.79		10.00%	
Job Carried out by		Comments	
Date Completed			
Service Time			