



Honeywell Healthcare Solutions GmbH

✓CID4348 Opera 00007282 Type:Company Supplier Customer
 Honeywell Healthcare Solutions GmbH Alter Holzhafen 18, Wismar,
 , D-23966, Germany
 Tel 0049 3841 360204 Euro Distributor (0004) [Log Call](#)

[Add Issue](#)

VAT DE137437213

Account Home	Main Details	Account Memos	Sales History	Purchase Orders	Supplier Returns	Stock References	Cust./Supp. Returns	Service Visits
Distributor Agreements	Documents Files	Linked Issues	Regulatory Agreements					

BOX:910

Manufacturer RMA : 300262460

☒ Repaired ☐ No Fault Found

[Returned Barcode](#)
[Close File 910](#)
[Tick All Replaced](#)

Review	Stock Ref	Sup Ref	B/C	S/N	Fault	Replaced	No Fault Found	Repaired	Credited	Out Warranty	Returned Unrepaired	Notes
Review 0110021 Customer Stats	0110021	0110021	2028842	V118584	No Output Country of Origin: Germany Custom Value:5.00 euro	☐	☐	☐	☒	☐	☐	

[Update Returns](#)

Customer Failure Statistics

Start From 02/05/2021 End at 02/05/2023

COMPLETED

cg 2-5-23
complete.

Viamed Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA: 300262460

Andre Schliemann

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 31 Mar 2023

Viamed RN: RTS1100910

Shipper Address:

Viamed

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634542

+44 (0) 1535 635582

info@viamed.co.uk

Returns Under Warranty Value for Customs : 5.00 Euro

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
0110021	0110021	E1002722 OOM103 Oxygen Sensor	1
Tariff Code: 9019200000			

12A847290454753467

CG 31-3-23

Shipped with box 909

waiting for
report email
sent. CG. 2-5-23

COMPLETE

Serial Number	RMA	Fault Description
V118584	300262460	No Output Country of Origin: Germany Custom Value:5.00 euro

ORIGINAL

Correction 5263412999

27.04.2023

Honeywell Healthcare Solutions GmbH
 Alter Holzhafen 18
 23966 Wismar
 Germany

RTS1100910

1797550

Viamed Ltd. Cross Hills
 15 Station Road
 KEIGHLEY
 BD20 7DT
 United Kingdom

Sold to:

1797550

Viamed Ltd. Cross Hills
 15 Station Road
 KEIGHLEY
 BD20 7DT
 United Kingdom

Order Information:

Credit Memo Request: 64064850
 Sales Office: 2170 DE SPS Wismar
 Incoterms: FCA Origin
 Payment Terms: Net 30 Days After Invoice Date

Ship to:

1797550

Viamed Ltd. Cross Hills
 15 Station Road
 KEIGHLEY
 BD20 7DT
 United Kingdom

EM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
00010	E1002722 Oxygen Sensor OOM103 VIA Customer Part Nbr: 0110021 Credit note for SN:118584	1	EA	22,28	-22,28	0%

Totals

Total Before VAT -22,28 EUR
 VAT-Total 0,00 EUR
 Total Amount -22,28 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods
 Fulfilment Date: 27.04.2023

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.help@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

We should ~~not~~
 have received it
 €18.11 not €22.28

Steve Nixon requested
 for me not to get
 this adjusted.

Cg 2.5.23

Ins #:

Bank Details: No Payment Due Account Name

Account No: Bank Key:

AN: Payment method

Registered No: Schwerin HRB 1838 VAT Reg No: DE137437213

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar

Credit Analyst: Xenia Musat

Customer Care Rep: Jessica Wagner +49 3841 360 206

Unless otherwise agreed in writing by the parties, all sales of products and/or services in this document shall be governed by Honeywell, a copy of which can be found at <https://sps.honeywell.com/us/en/support/safety/legal-documents/sst>. Continuing with this transaction, you (Buyer) agree to those Honeywell Terms and Conditions, which are subject to the performance of this Purchase Order to the terms in the Honeywell Terms and Conditions and expressly rejects any conditions governed by the Terms and Conditions of the Honeywell legal entity listed on this document.

Base Order is
 in force. By
 fulfillment and
 s. This sale shall

PRÜFBERICHT - TEST REPORT Sauerstoffsensoren MED / Oxygen Sensors Medical Use

Vorgangsnummer/ Process Number	48557781	Benachrichtigungsnummer/ Notification Number	300262460
-----------------------------------	----------	---	-----------



Kunde / Customer <i>Via med</i>	Kunden - Auftragsnummer / Customer Order No.	RTS1100910
Auftragsdatum / Date of Order	28.03.2023	
Lief.an code/ Delivery Code	Aufgenommen von / Received by	<i>V.G.</i>

Beanstandung / Objection

0-Signal

Artikelnummer / Part No.	Herstell. Datum / Date of Manufact.	Code + Seriennummer / Code + Serial No.	Garantie bis / End of Warranty Date	Beschreibung / Description
E1002722	<i>06/22</i>	118584	<i>09/23</i>	Oxygen Sensor OOM103 VIA

Prüfkriterien unter Standardbedingungen / Test Parameters under Standard Conditions

- | | | |
|---|---|---------------|
| 1 | Ausgangssignal an Luft / Output Signal in Atmosphere | [20,95 % vol] |
| 2 | Linearitätsfehler, Meßfehler / Linearity Error, Error of Measurment | [100 % vol.C] |
| 3 | Nullsignal / Zero Signal | [100 % vol.N] |

Messprotokoll / Test Report

Prüfkriterien / Test Parameters	Spezifiziert / Specified	Gemessen / Measured	
		Signal1	Signal1
1 [mV]	<i>11.2</i>	<i>0</i>	
2 [%]	<i>63</i>		
3 [µV]	<i>4200</i>		

Prüfbericht / Checklist

Fehler / Defect:	O2-S001..Luftwert außerh.Toleranz / Signal Out Of Tolerance
	O2-S002..Signal Instabil / Signal Not Stable
	O2-S003..Offset Außerh. Toleranz / Offset Out Of Tolerance
	O2-S004..Signal 0mV / Signal 0mV
	O2-S005..Gehäuse Defekt / Housing Defective
	O2-S006..Elektrolytaustritt / Electrolyte Leakage
	O2-S007..Sonstiges / Others
Bemerkungen / Remark	kein Fehler festgestellt, Sensor innerhalb Spezifikation / No failure found, Sensor within specifications
	außerhalb der Garantie / out of warranty

Anerkannt / Accepted ☒	Nicht anerkannt / Not Accepted ☐
--	---

Austausch / Replacement	Rücksendung / Return
VB-Nr. / VB No	Lieferdatum / Delivery Date
Seriennummer / Serial No.:	
Lieferdatum / Delivery Date	Geprüft / Inspected: <i>Buy</i>
Bearbeitet / Processed <i>12.04.23</i>	