

Kate Griffiths <viamed.kate.griffiths@gmail.com>

Re: MAX O2 AE and Oxygen Sensor, R-47V

2 messages

Supattra Polnikorn <bb_bio_chem2010@yahoo.com> To: kate.griffiths@viamed.co.uk

11 May 2023 at 09:43

Dear Kate Griffiths,

My customer told it cannot pass 100% calibration because it can read only up to 80%.

Best regards, Pongnarin Yoosomboon

Sent from my iPhone

On 11 May BE 2566, at 15:10, Kate Griffiths <office@viamed.co.uk> wrote:

Good morning Supattra,

Thank you for your email. Please could I ask what is the nature of the claim?

Many thanks.

Kind regards

Kate Griffiths

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

http://www.viamed.co.uk

Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542 Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Thu, 11 May 2023 at 03:29, Supattra Polnikorn

bb_bio_chem2010@yahoo.com> wrote: Dear Kate Griffiths,

My customer would like to claim 1pc of MAX-250K, S/N:HC52799120 invoice no. MVM138793 and arrived date was 21-11-2022.

I would like to inform you that Mr.Pongnarin Yoosomboon will start to handle Viamed's orders instead of me from May 1st.

Best regards,

Ponsorn Sengthan