

Service Repair Sheet SRS68453

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Suhail Ahmed

Northern General Hospital

Clinical Engineering

Clinical Engineer

0114 2715076

01142434343

00004560

suhail.ahmed1@nhs.net

004759705

10/May/2023

Robert Connor

Viamed

Repair

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 03/May/2023 Steve Hardaker

03/May/2023 Steve Hardaker

Returning a Tom Thumb for repair, which is reading incorrectly.

03/May/2023 Steve Hardaker

s/n 040393, supplied 01 Apr 2010.

04/May/2023 Aqib Majeed

PO received

10/May/2023 Robert Connor

Received 1 x tom thumb s/n 040393, gauge is not zeroed.

Ready For quote

CGreen 11-5-23

Repair Complete Signed

CGreen 11-5-23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35775	Tom Thumb	0310030	040393	N

03800000 - edited by 11-5-23

0310030 x 1 @ £80

SIN, SRS, SRN

UPS x 1 @ £12.