Service Repair Sheet SRS68466

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Purchasing

Eden Medical (UK) Ltd

01314406500

01314406500

00010170

00012178

info@edenmedical.co.uk

DPO-01414

17/May/2023

Robert Connor

Viamed

For Servicing

Cleaned by Viamed, if no declaration certificate from customer Signed: アメ

Notes 15/May/2023 Aqib Majeed 15/May/2023 Aqib Majeed Already sent in PO attached 15/May/2023 Aqib Majeed S/N PR02643A11 17/May/2023 Robert Connor Received 1 x V1000 s/n PR02643/

Received 1 x V1000 s/n PR02643A11, with 4 x AA battery and hard plastic case.

Ready For quote

Repair Complete Signed

CGreen 17-5.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35796	Foetal Simulator	1410000	PR02643A11	2

14800000 × 100 £38.25 5/N, SRS, SRN

1430309 x 10 FOC 5RS, SRN

UPS X 1 D E12

PURCHASE ORDER



Purchase Order To:

Viamed Limited 15 Station Road Cross Hills Keighley WEST YORKSHIRE BD20 7DT

United Kingdom

For Delivery To:

Eden Medical (UK) Limited
7 Dryden Place
Bilston Glen Industrial Estate
Loanhead
Midlothian
EH20 9HP
United Kingdom

Date

15/05/2023

Purchase Order No.

DPO-01414

VAT Registration No.

GB 316 5811 10

Product Code	Order Qty	Product Description	Unit Cost	Unit Qty	Total
CUSTEQT	2 1 2	Customer Equipment for Service/Repair Foetal Heart Rate Simulator V1000 [SN PR02643A11]	45.00	EACH	45.00

ADDITIONAL COSTS

Amount	Total
12.00	12.00
	57.00

Important information:

We reserve the right to reject any invoice or application for payment if any the details above have been altered without our agreement. Please note our order number must be quoted on all correspondence and a delivery note must accompany each consignment.



Service Repair Sheet 68466

Contact Name

Purchasing

Company/ Hospital Name

Eden Medical (UK) Ltd

Direct Phone General Phone 01314406500

Opera Account

01314406500 00012178

Email

info@edenmedical.co.uk

Order Number Date Received Booked in By DPO-01414

17/May/2023 Robert Connor

Repair

Ref

S/N

Equipment Type

SRN35796

1410000 PR02643A11

Foetal Simulator

Complete - Repaired

Time: 0 Hour(s)

Parts Replaced

Qty: 1

1480000

Qty: 1

1430309

V1000 has had transducer interface cushion replaced and calibration has been checked