

VOP			
Viamed Operating sub Process			
VM3/COP/02.01			
<u>Exclusions to ISO13485:2003</u>			

Current Exclusions from 13485:2003:

7.5.3.2.2 Exclusion active implantable devices

Not Applicable, No implantable devices.

7.5.2.2 Exclusion for sterile medical devices

Not Applicable, no sterile devices.

7.5.1.3 Particular requirement for sterile devices

Not Applicable, no sterile devices.