

Service Repair Sheet SRS68474

Contact Name  
Company/ Hospital Name  
Department  
Position  
Direct Phone  
General Phone  
Opera Account  
Email  
Order Number  
Date Received  
Booked in By  
Main Company  
Type Return

Hocine Benabdelhafid  
Homerton Hospital  
Medical Electronics & Physics  
Calibration Officer  
02085107935  
02085107935  
00002950  
hocine.benabdelhafid@nhs.net  
  
18/May/2023  
Robert Connor  
Viamed  
Quote

**VIAMEDclean**

**Goods In Only**

Decontamination  
certificate provided  
by customer ☐

Cleaned by Viamed,  
if no declaration certificate  
from customer ☒

Signed: RC  
Date: 18/5/23

**Goods Out Only**

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_  
Date: \_\_\_\_\_

Notes 17/May/2023 Kate Griffiths

17/May/2023 Kate Griffiths

service and calib for V1000, s/n PR01702A12, PO 334085930 attached. Sent in PO before asking for SRS

18/May/2023 Robert Connor

Received 1 x V1000 s/n PR01702A12, with 4 x AA battery and blue fabric carry case.

P/O attached

Ready For quote

CCNEON 23.5.23

Repair Complete Signed

CCNEON 26.5.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35800	Foetal Simulator	1410000	PR01702A12	N

1480000 x 1 @ £45  
S/N, SRS, SRN

1430309 x 1 FOC  
SRS, SRN

UPS x 1 @ £12

Service Repair Sheet 68474

Contact Name	Hocine Benabdelhafid
Company/ Hospital Name	Homerton Hospital
Department	Medical Electronics & Physics
Position	Calibration Officer
Direct Phone	02085107935
General Phone	02085107935
Opera Account	00002950
Email	hocine.benabdelhafid@nhs.net
Date Received	18/May/2023
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN35800	1410000	PR01702A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

---