



TELEDYNE
ANALYTICAL INSTRUMENTS
Everywhere you look

Teledyne Instruments, Inc.
16830 Chestnut Street
City of Industry, California 91748-1020
Phone: 626.934.1500 Fax: 626.961.2538
www.teledyne.ai.com

EU Declaration of Conformity

Product Identification

Product Name	Model Number
Oxygen Analyzer	AX 300 and AX 300-I

Manufacturer

Company Name	Address	Management Representative
Teledyne Analytical Instruments a business unit of Teledyne Instruments, Inc	16830 Chestnut Street City of Industry, CA 91748, USA	Vasu Narasimhan Director of Regulatory Affairs & Quality Assurance

Authorized Representative

Company Name	Address	Contact Information
Viamed Limited	15 Station Road, Cross Hills Kleghley, W. Yorkshire BD20 7DT United Kingdom	Phone: +44 (0) 1535 634542 URL: www.viamed.co.uk E-mail: Jean@Viamed.co.uk

Registration Information

Notified Body / ID #	CE Certificate Number	Date CE Marking first applied
BSI, UK / 0086	CE 02000	09 July 1998

Conformity Assessment

Device Classification	Route of Compliance	Standards Applied
Class IIa (Annex IX, Rule 10)	Annex II, Section 3.2, of MDD 93/42/EEC, as amended by Directive 2007/47/EC	EN 60601-1, 3 rd Edition EN 60601-1-2 EN 60601-1-6 EN 14971:2007 EN ISO 21647

We, Teledyne Analytical Instruments, a business unit of Teledyne Instruments, Inc, declare that the above mentioned products meet the provision of the Council Directive 93/42/EEC, as amended by Council Directive 2007/47/EC, for Medical Devices. All supporting documents are retained at the premises of the manufacturer.

Signature: _____

Name: Vasu Narasimhan

Title: Director of RA & QA

Date: June 1, 2012

Signature: _____

Name: Stephen Broy

Title: Director of Engineering

Date: June 1, 2012