



Honeywell Healthcare Solutions GmbH

/CID4348 Opera 00007282 Type:Company Supplier Customer
Honeywell Healthcare Solutions GmbH Alter Holzhafen 18, Wismar,
D-23966, Germany
Tel 0049 3841 360204 Euro Distributor (0004) Log Call

Add Issue
VAT DE137437213

Account Home	Main Details	Account Memos	Sales History	Purchase Orders	Supplier Returns	Stock References	Cust./Supp. Returns	Service Visits
Distributor Agreements	Documents Files	Linked Issues	Regulatory Agreements					

BOX:920

Manufacturer RMA : 300264772

☒ Repaired ☐ No Fault Found

Returned Barcode

Close File 920

Tick All Replaced

Review	Stock Ref	Sup Ref	B/C	S/N	Fault	Replaced	No Fault Found	Repaired	Credited	Out Warranty	Returned Unrepaired	Notes
Last Rev Jun 23 0110040 Customer Stats	0110040	0110040	2164226	V107288	No Output Country of Origin: Germany Customs Value: 5.00 Euro	☐	☐	☐	☒	☐	☐	

Update Returns

Customer Failure Statistics

Start From 04/07/2021 End at 04/07/2023

Invoice Correction 5264008960

ORIGINAL

Invoice Correction Date: 27.06.2023

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

Purchase Order No: RTS1100920

23966 Wismar

Germany

Bill to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

Sold to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

Order Information:

Credit Memo Request: 64095064
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797550

Viamed Ltd. Cross Hills
15 Station Road
KEIGHLEY
BD20 7DT
United Kingdom

EM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
00010	E1002731	1	EA	37,73	-37,73 ✓	0%

Oxygen Sensor OOM110 VIA
Customer Part Nbr: 0110040

credit note for SN:107288

Totals

Total Before VAT -37,73 EUR
VAT-Total 0,00 EUR
Total Amount -37,73 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods
Fulfilment Date: 27.06.2023

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.helpline@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

Invoice #:

Link Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

AN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner +49 3841 360 206

Unless otherwise agreed in writing by the parties, all sales of products and/or services in this document shall be governed solely by the applicable Honeywell Terms and Conditions in effect at the time a Purchase Order is accepted by Honeywell, a copy of which can be found at <https://sps.honeywell.com/us/en/support/safety/legal-documents/sst>. Honeywell's acknowledgment of receipt of a Purchase Order shall not constitute acceptance. By continuing with this transaction, you (Buyer) agree to those Honeywell Terms and Conditions, which are subject to change at Honeywell's discretion without notice. Honeywell expressly limits its acceptance, fulfillment and performance of this Purchase Order to the terms in the Honeywell Terms and Conditions and expressly rejects any different, conflicting or additional terms in the Purchase Order or any other Buyer documents. This sale shall be governed by the Terms and Conditions of the Honeywell legal entity listed on this document.

PRÜFBERICHT - TEST REPORT

Sauerstoffsensoren MED / Oxygen Sensors Medical Use

Vorgangsnummer/ Process Number	48660905	Benachrichtigungsnummer/ Notification Number	300264772
-----------------------------------	----------	---	-----------



Kunde / Customer <i>Via med</i>	Kunden - Auftragsnummer / Customer Order No. RTS1100920
Auftragsdatum / Date of Order 07.06.2023	
Lief.an code/ Delivery Code	Aufgenommen von / Received by <i>U. G.</i>

Beanstandung / Objection

0-Signal

Artikelnummer / Part No.	Herstell. Datum / Date of Manufact.	Code + Seriennummer / Code + Serial No.	Garantie bis / End of Warranty Date	Beschreibung / Description
E1002731	<i>01-23</i>	107288	<i>03-23</i>	Oxygen Sensor OOM110 VIA

Prüfkriterien unter Standardbedingungen / Test Parameters under Standard Conditions

- | | | |
|---|---|-------------------------------|
| 1 | Ausgangssignal an Luft / Output Signal in Atmosphere | [20,95 % vol.O ₂] |
| 2 | Linearitätsfehler, Meßfehler / Linearity Error, Error of Measurment | [100 % vol.O ₂] |
| 3 | Nullsignal / Zero Signal | [100 % vol.N ₂] |

Messprotokoll / Test Report

Prüfkriterien / Test Parameters	Spezifiziert / Specified	Gemessen / Measured	
		Signal1	Signal1
1 [mV]	<i>11.1</i>	<i>0</i>	
2 [%]	<i>63</i>		
3 [µV]	<i>6200</i>		

Prüfbericht / Checklist

Fehler / Defect:	O2-S001..Luftwert außerh.Toleranz / Signal Out Of Tolerance
	O2-S002..Signal Instabil / Signal Not Stable
	O2-S003..Offset Außerh. Toleranz / Offset Out Of Tolerance
	O2-S004..Signal 0mV / Signal 0mV
	O2-S005..Gehäuse Defekt / Housing Defective
	O2-S006..Elektrolytaustritt / Electrolyte Leakage
	O2-S007..Sonstiges / Others
Bemerkungen / Remark	kein Fehler festgestellt, Sensor innerhalb Spezifikation / No failure found, Sensor within specifications
	außerhalb der Garantie / out of warranty

Anerkannt / Accepted ☒	Nicht anerkannt / Not Accepted ☐
--	---

Austausch / Replacement	Rücksendung / Return
VB-Nr. / VB No	Lieferdatum / Delivery Date
Seriennummer / Serial No.:	Geprüft / Inspected: <i>Bull</i>
Lieferdatum / Delivery Date	
Bearbeitet / Processed <i>18.06.23</i>	