

Service Repair Sheet SRS68498

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Jake Sales

Kings College Hospital (use this o

02032991607

00002990

jake.sales@nhs.net

05/Jul/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC
Date: 5/7/23

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 26/Jun/2023 Janine Gill

26/Jun/2023 Janine Gill

Emailed customer with SRS number and instructions on sending the device back to us.

MaxO2 ME with a faulty screen, the serial number is GK17799001.

05/Jul/2023 Robert Connor

Received 1 x MAXO2 ME s/n GK17799001, coiled cable s/n 1875422, MAX550E s/n HH65699192, with 4 x AA battery.

Ready For quote

Repair Complete Signed

CC Green 7.7.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35870	Oxygen Monitor	0111263	GK17799001	Yes
SRN35871	Oxygen Monitor Cable	0131700	Tracking10	
SRN35872	Oxygen Sensor	0110452	HH65699192	Yes

3 part of unit.

0140050 x 1 FOC
old S/N, SRS, SRN

0111263 x 1 FOC
New S/N, SRS, SRN
includes replacement cable as part of unit

0146011 x 1 FOC
S/N, SRS, SRN

UPS x 1 FOC

COMPANY: MAXTEC - Viamed Ltd

Viamed Ltd
ADDRESS: 15 Station Road
Cross Hills
Keighley
BD20 7DT

King's College Hospital **NHS**

NHS Trust

Department of Medical Engineering & Physics
Kings College Hospital
Faraday Building
124-126 Denmark Hill
LONDON SE5 8RX

Tel: 020 3299 3538

DATE: 26/06/2023

TEL: +44 (0)1535 634542

FAX:

Dear Sir or Madam,

Please find enclosed the following equipment:

Returns Number: **SRS68498**

DESCRIPTION: OXYGEN MONITOR	
Model: MaxO2 ME	Serial No: GK17799001
ID Code: 77-1023	Call No: 291839 MTO: Jake.sales@nhs.net
PROBLEM: Screen faulty	
COMMENTS: REF: SRS68498	

We would be grateful if you could:-

Provide an estimate for its repair: EMAIL TO - jake.sales@nhs.net	N/A
Repair it against our official order	N/A
Repair it under warranty	YES

ORDER No:	
------------------	--

REQUISITION No:	
------------------------	--

Yours sincerely,

Engel Bejarin
e.bejarin@nhs.net

ALL EQUIPMENT TO BE RETURNED TO THE ADDRESS BELOW

Medical Engineering & Physics – MEMS
Faraday Building
124-126 Denmark Hill
LONDON
SE5 8RX

Tel: Direct Line 020 3299 3538

COMPANY: MAXTEC - Viamed Ltd

ADDRESS: Viamed Ltd
15 Station Road
Cross Hills
Keighley
BD20 7DT

King's College Hospital **NHS**

NHS Trust

Department of Medical Engineering & Physics
Kings College Hospital
Faraday Building
124-126 Denmark Hill
LONDON SE5 8RX

Tel: 020 3299 3538

DATE: 26/06/2023

TEL: +44 (0)1535 634542

FAX:

Dear Sir or Madam,

Please find enclosed the following equipment:

Returns Number: **SRS68498**

DESCRIPTION: OXYGEN MONITOR	
Model: MaxO2 ME	Serial No: GK17799001
ID Code: 77-1023	Call No: 291839 MTO: Jake.sales@nhs.net
PROBLEM: Screen faulty	
COMMENTS: REF: SRS68498	

We would be grateful if you could:-

Provide an estimate for its repair: EMAIL TO - jake.sales@nhs.net	N/A
Repair it against our official order	N/A
Repair it under warranty	YES

ORDER No:	
------------------	--

REQUISITION No:	
------------------------	--

Yours sincerely,

Engel Bejarin
e.bejarin@nhs.net

ALL EQUIPMENT TO BE RETURNED TO THE ADDRESS BELOW

Medical Engineering & Physics – MEMS
Faraday Building
124-126 Denmark Hill
LONDON
SE5 8RX

Tel: Direct Line 020 3299 3538

Service Repair Sheet 68498

Contact Name	Jake Sales
Company/ Hospital Name	Kings College Hospital (use this one)
Direct Phone	02032991607
Opera Account	00002990
Email	jake.sales@nhs.net
Date Received	05/Jul/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN35870	0111263	GK17799001	Oxygen Monitor	

Parts Replaced Time :0 Hour(s)

Qty: 1 0111263

Replaced under warranty with MAXO2ME Monitor S/N HJ80299023 that includes Sensor Cable barcode 2112240.

Repair	Ref	S/N	Equipment Type
SRN35871	0131700	Tracking10	Oxygen Monitor Cable

Time :0 Hour(s)

Replaced under warranty with MAXO2ME Monitor S/N HJ80299023 that includes Sensor Cable barcode 2112240.

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN35872	0110452	HH65699192	Oxygen Sensor	

No Fault Found
Time :0 Hour(s)

No fault found.
