

Compatible SpO₂ Finger Probes Cross Reference

Original Equipment Manufacturer	Cross Reference of Model and Part Numbers	Length *Metres	Part Number
Baxter	Explorer Monarch	0.9	P856
Baxter	600333-003 ASAT	3.7	P860RA
Biochem (BCI)	3044 3300 3301 3302	0.9	P861RA
Bruker	Bruker	0.9	P861RA
Corometrics	556	0.9	P856RA
Criticare	511SD/934SD 1100 1100-3 501 503 504 507 508 581 Poet Plus	3.0	P862RA
Criticare	511SD/934SD 1100 1100-3 501 503 504 507 508 581 Poet Plus (Requires P968E8 or P969E8)	0.9	P869RA
Critikon	9700 Series	0.9	P856RA
Critikon	9700 Series	3.7	P858RA
Critikon	9084 8700 Series	0.9	P865RA
Critikon	8997 Dinamap Oxytrak	0.9	P877RA
Datascope	0600-00-0026-02 Passport (Requires P963E10)	0.9	P863RA
Datascope	0600-00-0026-01 Accusat Accutor	3.7	P864RA
Datex	Fits Most Datex Models	3.7	P872RA
Datex	Fits Most Datex Models (Requires P973E10)	0.9	P873RA
Dräger	Vitalert 1000 Narkomed Cicero	0.9	P856RA
Elmed	6006194 PSO1000	0.9	P861RA
Entri Medical	Propaq 102 104 106	0.9	P856RA
Fukuda Denshi	HG302 Module	0.9	P856RA
Gambro	BP806	0.9	P856RA
Invivio	4500 Plus 2	0.9	P856F
Invivio	9383 Scout ES 4500 Scout Scout + Scout +1 Scout +2 Scout +3 Scout +4	3.0	P878K
Ivy Medical	405A	0.9	P856RA
Kontron	7840	3.7	P886RA
Marquette	Eagle 3100	0.9	P856RA
Marquette	Tram System Modular System	3.7	P867RA
Medical Data	2800-50 OP11A OP30 Escort 100 Escort 11	0.9	P856RA
Mennen	800-060-730 Horizon XL (Nellcor)	0.9	P856RA
Mennen	800-060-540 Horizon XL (Ohmeda)	3.7	P867RA
Nellcor	C-13-200M N-200 DS100A (Compatible with Nellcor derivatives. Requires P956E4 or P956E8)	0.9	P856RA
Nellcor	N6000 N180E N180	3.7	P858RA
Nonin	8500 8604K (Requires P971E3 or P971E10)	0.9	P871RA
Novamatrix	8660 500 Series	3.0	P875RA
Novamatrix	8774 8776 505 511 515 520 7000 Oxypleth	3.0	P876RA
Ohmeda	3700 4700 Oxystat Plus 4700 380-1000-042 6051-0000-049 Oxystat Modular CD Anac	3.7	P867RA
Pacetek	520	3.7	P859RA
Propaq	Propaq	0.9	P856RA
S & W Vickers	4621643 4700708 DSNT	0.9	P856RA
Sensormedics	766721 767367 Oxyshuttle OMNI-SAT	0.9	P856RA
Siemens	1281	0.9	P856RA
Simed	S-100c	3.7	P860RA
Spacelabs	015-0150-00 90465/66/67/89 90651A-08 690-0003-00 (Nellcor). Requires P926E10	0.9	P856F
Spacelabs	90651A-08 90419-01 90485 90489 90467 90466 90465 90351-D/G	3.0	P857RA
Spacelabs	369-083-001 90419 90501 90651A-03 (Novamatrix)	3.0	P875RA
	Extension Cables		
Critikon/Sensormedics	9084 8700 Series **P865RA)	3.0	P965E10
Criticare	518LD 1100 1100-3 Poet Plus 503 504 507 508 581 (DB9 to DB9) **P869RA	2.4	P968E8
Criticare	518DD 503 Spot (Lemo to DB9) **P869RA	2.4	P969E8
Datascope	012-00-0516-12 Passport (Switchcraft to DB9) **P863RA	3.0	P963E10
Datex	Nicolet to DB9 **P873RA	3.0	P973E10
Nellcor	EC-4 N180 N180E N6000 (DB9 to DB9) **P856RA	1.2	P956E4
Nellcor	EC-4 N180 N180E N6000 (DB9 to DB9) **P856RA	2.4	P956E8
Nonin	85001 (DB9 to DB9) **P871RA	0.9	P971E3
Nonin	8604x-10 (DB9 to DB9) **P871RA	3.0	P971E10
Spacelabs	175-0646-00 90465/66/67/89 90651A-08 90351-D/G (7 Pin Hypertronic to 12 Pin) **P876RA	0.6	P925E2
Spacelabs	175-0706-00 90465/66/67/89 90651A-08 690-0003-00 90351-DG (7 Pin Hypertronic to 12 Pin) **P876RA	3.0	P926E10

VIAMED

15 Station Road, Cross Hills, Keighley,
West Yorkshire Bd20 7DT UK
Tel: 01535 634542 Fax: 01535 635582
www.Viamed.co.uk Email: info@viamed.co.uk

CE0086

For any manufacturer not mentioned please contact us.
All Manufacturers' trade marks and registered names recog

* Lengths are approximate

** Indicates suitable for use with stated finger probe.

Distmci97 20/7/99