










2 UNLESS OTHERWISE SPECIFIED, ALL PARTS ARE ODS
1 REMOVE ALL BURRS & BEVEL SHARP EDGES.



SECTION VIEW - RECEPTACLE


DETAIL A
(DB9F)

Sheet 3 Of 3

Title			
Criticare Extension cable P968E8		Dim in	Inches
		Tol	: 0.01
Scale	Not To Scale	Drawn	J.Nirwan
Date	30/01/02	Dwg No.	SPE-010.1

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	Details	Approv	Material :
				Part No0038697




DETAIL B
(DB9M)

Sheet 2 of 3

Title			
Criticare service cable P968E8		Dim in	Inches
		Tol	: 0.01
Scale	Not To Scale	Drawn	J.Nirwan
Date	30/01/02	Dwg No.	SPE-010.1

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

				Material :
REV	Date	Details	Approv	Part No0038697



8. 0.02" RECESSED PANEL.

8. CABLE/CONNECTOR SHOULD WITHSTAND MIN. 20 LB AXIAL PULL FORCE.

5. FINISH CABLE ASSEMBLY TO BE 100% ELECTRICALLY TESTED.

3. MAT'L

CONNECTOR / STRAIN RELIEF :

PVC, SHORE 80-85 A, MEDICAL GRADE.
COLOR: WHITE, MATCH CABLE COLOR.

CONTACTS PINS:

PHOSPHOR BRONZE, 0.040 TYP IN.
FINISH: GOLD FLASH OVER 50 MIN. NICKLE.

CABLE:

P/N 75-00100 (CUSTOM FURNISHED)

2. UNLESS OTHERWISE SPECIFIED, ALL RADI ARE 0.03

1. REMOVE ALL BURRS & BREAK SHARP EDGES.

Sheet 1 Of 3




Title			
Criticare service cable P968E8		Dim in	Inches
Scale		Tol	± 0.01
Not To Scale		Drawn	J.Nirwan
Date	30/01/02	Dwg No.	SPE-010.1

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	Details	Approv	Material :
				Part No0038697



D-Type Female 9 Pin

D-Type Male 9 Pin






D-Type Female 9 Pin

D-Type Male 9 Pin



D-Type Female 9 Pin

D-Type Male 9 Pin



Title P968E8
Criticare

Dim in mm
Tol ± 0.2
Drawn J.Nirwan



VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

REV	Date	N°	Drawn	Approv	Material :
					Part No. 0019688


Scale	Not To Scale	Date	05/12/01	Dwg No.	SPE-010
-------	--------------	------	----------	---------	---------

D-Type Female 9 Pin

D-Type Male 9 Pin



Title P968E8
Criticare



Dim in mm

Tol ± 0.2

Drawn J.Nirwan

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

					Material :
REV	Date	N°	Drawn	Approv	Part No. 0019688

Scale	Not To Scale	Date	05/12/01	Dwg No.	SPE-010.1
-------	--------------	------	----------	---------	-----------