




Hypertronics Male 9 Pin

Y- Probe




Hypertronics Male 9 Pin

Y- Probe




Hypertronics male 9 pin




Hypertronics Male 9 Pin

Y- Probe




Enlarged front view




Hypertronics Male 9 Pin

Y- Probe




Enlarged front view




Hypertronics Male 9 Pin

Y- Probe



Pin Housing Rear View

(3 & 4) R1 = 1.98 k



Title P866YA
Simed
Baxter



Dim in mm

Tol ± 0.2

Drawn J.Nirwan


VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

					Material :
REV	Date	N°	Drawn	Approv	Part No. 0018662


Scale	Not To Scale	Date	13/05/02	Dwg No.	SPY-860
-------	--------------	------	----------	---------	---------

Hypertronics Male 9 Pin


Y- Probe



Enlarged front view




Hypertronics male 9 pin



R1 - 1.58 K OHM 1%, 1/8W RESISTOR

* INNER SHIELD AND OUTER SHIELD TO BE TWISTED TOGETHER.

Title P866YA
Smed
Baxter



Dim in mm

Tol ± 0.2

Drawn J.Nirwan

VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT



Ltd.

REV	Date	N°	Drawn	Approv	Material :
					Part No. 0018662

Scale	Not To Scale	Date	13/05/02	Dwg No.	SPY-860.1
-------	--------------	------	----------	---------	-----------

Hypertronics Male 9 Pin

Y- Probe



Title: P866YA
Simed Baxter



Dim in	mm
Tol	± 0.2
Drawn	J.Nirwan

Scale: Not To Scale

Date: 13/05/02

Dwg No.: SPY-860



VIAMED Ltd.
15 Station Rd
Cross Hills, Keighley
West Yorkshire
BD20 7DT

Material:

Part No. 0018662

REV	Date	N°	Drawn	Approv
-----	------	----	-------	--------