

Service Repair Sheet SRS68525

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Mark Taylor

Gloucestershire Royal Hospital - C

Medical Engineering

Senior Engineer

0300 422 6116

0300 422 2222

00001902

mark.taylor15@nhs.net

27/Jul/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 25/Jul/2023 Catrin Hird

25/Jul/2023 Catrin Hird

s/n 50940153

Returning Capnograph, faulty display.

27/Jul/2023 Robert Connor

Received 1 x VM-2500-S s/n 50940153, with silicone cover, li-poly battery, orange fabric case, power supply v3, 1 x masimo nomoline.

Ready For quote

CGneon 1-8-23

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35938	Capnograph	4410520	50940153	N

4440000 x 0.5hr @ £90 per hr = £45

S/N, SRS, SRN

4430120 x 1 £80

SRS, SRN

UPS x 1 @ £12

MEDICAL EQUIPMENT FOR TRANSPORT AND/OR REPAIRGloucestershire Hospitals **NHS**
NHS Foundation Trust
Medical Engineering Department

Please attach this form to any Medical Equipment which requires transportation and/or repair

HELPDESK CALL REF:		Med Eng Use Only (Optim Job No):	
SECTION 1 : Equipment			
ME-38098		DESCRIPTION Screen	
SECTION 2 : Fault (if applicable) – please give as much detail as possible, including who experienced the fault			
Screen Pixels missing			
If this equipment has been involved in an incident, please state the Datix Incident Number:			
SECTION 3 : Decontamination			
STATE ANY CONTAMINATION:		None	
(eg: blood, body fluids, respired gases, pathological samples, chemicals, cytotoxic drugs, radioactive material or any other hazard)			
METHOD OF DECONTAMINATION:		Clinel Wipes	
SECTION 4 : Destination (if transporting)			
Medical Engineering <input checked="" type="checkbox"/>		Equipment Library <input type="checkbox"/>	Ward/Department <input type="text"/>
SECTION 5 : Declaration			
I declare that I have taken all reasonable steps to ensure the accuracy of the above information, the equipment has been cleaned & decontaminated as far as is practically possible and has been prepared to ensure safe handling and transportation			
SITE: GRH		WARD/DEPT: Medical Engineering	TEL: 0300 4226116
PRINT NAME: Mark Taylor		POSITION: Specialist Engineer	
SIGNED: <i>M Taylor</i>			DATE: 25 / 07 / 2023

Catherine Spence **1 Calls** 34 Issues 24 **Unread** 22 ISO Tasks **22.00**


°C

Search



Gloucestershire Royal Hospital - Gloucestershire Managed Services orders only

✓CID4187 Opera 00001902
Type:NHS Hospital End User
**Gloucestershire Royal Hospital -
Gloucestershire Managed
Services orders only , c/o
Distribution Stores, Great Western
Road, Gloucester, , GL1 3NN, UK
Tel 0300 422 2222 GBP End User
(0001)** [Log Call](#) [Add Issue](#)
Credit Credit Cr Credit Limit
0.00GBP



Use :0001 GBP End User

Repair SRS68525 Values in Pounds @ £1: £1.00

SRN35938	Capnograph	4410520			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4440000	0.5	£90.00	£45.00
		4430120	1	£80.00	£80.00
				Parts Cost	£125.00

**Page updated, TRipple check prices to price
in intrastats main search**

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent