
October 28, 1998

Mr. Peter Lamb
UK Sales Manager
Viamed Limited
15, Station Road, Cross Hills
BD20 7DT Keighley, West Yorkshire
United Kingdom

RE: "Wrap Arounds" Quotations



Dear Mr. Lamb:

Thank you for this opportunity to quote you on the manufacture of the "Wrap Arounds". You will find several quotations below which reflect the fabric (TIB# 0398006) and foam (TIB# 0296007) with and without a printed linear Please note, the quotation is per sheet of 5 "Wrap Arounds".

Product Description:

Wrap Arounds - sheet of 5 pieces Sheet Size = 87mm x 120mm

Size and Shape:

Part Size = 15mm x 120mm

Materials:

Two Materials Quoted Below

Fabric: TIB# 0398006 Foam: TIB# 0296007

	TIB# 0398006 w/ printed liner	TIB# 0398006 w/out printed liner	TIB# 0296007 w/ printed liner	TIB# 0296007 w/out printed liner
5000±35"	\$0.3157	\$0.2550	\$0.3058	\$0.2452
10.000807	\$0.2017	\$0.1664	\$0.1920	\$0.1566
ESOROF.	\$0:1638	\$0.1369	\$0.1540	\$0.1271
20(000=	\$0.1448	\$0.1222	\$0.1350	\$0.1123
25,000=	\$0:1334	\$0 .1133	\$0.1236	\$0.1035
30,000€	\$0.1258	\$0.1074	\$0.1160	\$0.0976

Tooling (2 mols): \$1175 + 8.25% CA Sales Tax

Printing Plates: \$2,80 + 8.25% CA Sales Tax (only for printed liner options)

Packaging 3. 3. Bulk packed in pole bees

Property of the Property of th

Will came once again that the of this opportunity to quote on the manufacture of the "Wrap Arounds".

Show that the thing of the "Case do not hesitate to contact the little of the "Wrap Arounds".

e Oraling Iso yillin



ACUTEK	ITEM NO. 0398006	REVISION LEVEL
Adhesive Specialt MATERIAL DESCRIPTION	White elastic woven fabric, ctd. on one side with a latex free medical grade rubber adh., supplied on 60# liner.	PAGE: 1 of 1

GENERAL MATERIAL INFORMATION

1.0. MATERIAL DESCRIPTION:

A 5.5mil latex free rubber based adhesive coated on one side of a white elastic woven fabric with a 60# semibleached densified Kraft release liner. Designed for flexible bandage applications, high stretch and excellent conformability.

Features:

High Stretch, excellent conformability. Adhesive system provides good adhesion to skin yet removes cleanly without causing skin trauma.

2.0. PRODUCT PROFILE:

2.1. Backing:

White woven fabric, 24.5mils (.6223mm)

2.2. Adhesive:

Clear Synethic Rubber based Adhesive, 5.5 mils (.12 mm)

2.3. Liner:

60 lb. bleached Kraft, 3.5 mils (.089mm)

3.0. PHYSICAL PROPERTIES:

Typical Value

Test Method

Peel Adhesion

Initial to Stainless Steel (1min @ RT)

120.0 oz/in (32 N/25mm)

14 hours

PSTC #7 Mod/178°

PSTC #1 Mod/180°

2.2psi (1" x 1" x 1000g) @ RT

4.0. PRODUCT SAFETY:

Holding Power (load)

The product has been tested by an independant laboratory for Primary Skin Irritation in accordance with the guidelines of the ISO Part 10: Tests for Irritation and Sensitization. Under the conditions of this study, no irritation was observed on the skin of rabbits. The Cumulative Irritation Index for the test article was calculated to be 0.0 and the Cumulative Irritation Index for the control article was calculated to be 0.0. The corresponding Response Category was negligible.

Values stated are typical and are not intended for specification development. Material is sold with the understanding that the BUYER will determine the suitability for their purposes

FDA Master Files for Devices# Available Upon Request

TIB# 0398006