



<b>Spectrometer.</b>	SPM 9001.
<b>Manufacturer of Probe.</b>	BCI.
<b>Type.</b>	Finger.
<b>Part no.</b>	Model 3044.
<b>Serial no.</b>	170271350
<b>Connector.</b>	9 pin male, D connector.
<b>Connect to.</b>	Oximeter or extension.
<b>For use on.</b>	
<b>Tested on.</b>	Nellcor N-200
<b>Red peak wavelength (nm).</b>	661.3.
<b>Infrared peak wavelength (nm).</b>	<del>892.7</del> 876.5 - 892.7.
<b>Internal resistance value.</b>	Unknown.
<b>Emitter configuration.</b>	Unknown.
<b>Carried out by.</b>	S Watmough. I Eng., B Eng.
<b>Date.</b>	05-01-00.