

VIA MED SpO₂ FINGER PROBES

Part number	*Compatible oximeter manufacturer	Cable length		Connector		
		Feet	Metres	Type	Number of contacts	Gender
P856RA	Nellcor	3	0.90	D	9	M
P857RA	Spacelabs	10	3.00	Hypertronics	12	M
P858RA	Nellcor	12	3.65	D	9	M
P859RA	Pace Tech	12	3.65	Switchcraft	5	F
P860RA	Simed, Baxter	12	3.65	Hypertronics	9	M
P861RA	BCI	3	0.90	D	9	M
P862RA	Criticare	10	3.00	Lemo	5	M
P863RA	Datascope	3	0.90	D	9	M
P864RA	Datascope	12	3.65	DIN	8	M
P865RA	Sensormedics	3	0.90	Compu-shield	6	M
P866RA	Simed, Baxter	12	3.65	Hypertronics	9	M
P867RA	Ohmeda	12	3.65	Hypertronics	9	M
P869RA	Criticare	3	0.90	D	9	M
P871RA	Nonin	3	0.90	D	9	M
P872RA	Datex	12	3.65	DPK1	10	F
P873RA	Datex	3	0.90	D	9	M
P875RA	Novamatrix	10	3.00	Hypertronics	7	M
P876RA	Novamatrix	10	3.00	Hypertronics	7	M
P877RA	Critikon	3	0.90	D	9	M
P878RA	Invivo	10	3.00	Lemo	7	M
P886RA	Kontron	12	3.65	DPK1	12	M
P887RA	Spacelabs	10	3.00	Hypertronics	12	M
P888RA	S & W	12	3.65	SPK1	10	F
P889RA	Criticare	10	3.00	D	9	M
P890RA	Critikon/Sensormedics	10	3.00	Thorkom	12	F
P891RA	Critikon	10	3.00	Lemo	7	M

CE 0086

* Main compatible oximeter manufacturer for guidance purposes only. For full details of other compatible manufacturers, please refer to cross reference list. Cable lengths are approximate.