

Service Repair Sheet SRS68537

Contact Name

Guyan Perera

Company/ Hospital Name

Southmead Hospital

Department

Position

Direct Phone

General Phone

0117 9505050

Opera Account

00000740

Email

guyan.perera@nbt.nhs.uk

Order Number

Date Received

14/Aug/2023

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 07/Aug/2023 Sophie Lines

07/Aug/2023 Sophie Lines

Guyan Perera called and advised he has a faulty MAX O2 ME Monitor - serial number HG36099001 - unit is still under warranty. I have emailed Guyan the SRS number and instructions for return.

14/Aug/2023 Robert Connor

Received 1 x MAXO2ME s/n HG36099001, with coiled cable s/n 2076515, no batteries or sensor. Customer states unit not switch off.

Ready For quote

Repair Complete Signed

CGNEON 15.8.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35965	Oxygen Monitor	0111263	HG36099001	Yes
SRN35966	Oxygen Monitor Cable	0131700	Tracking10	Yes

3 Part of
complete
unit.

0140050 x 1 FOC

Old S/N, SRS, SRN

0111263 x 1 FOC

New S/N, SRS, SRN

Warranty Replacement

UPS x 1 @ £12.

→ should be £0 my mistake I added
chargeable carriage by accident.
cg. 17.8.23.