



Honeywell Healthcare Solutions GmbH

✓CID4348 Opera 00007282 Type:Company Supplier Customer
Honeywell Healthcare Solutions GmbH Alter Holzhafen 18, Wismar
, D-23966, Germany
Tel 0049 3841 360204 Euro Distributor (0004) [Log Call](#)

[Add Issue](#)

VAT DE137437213

- [Account Home](#) [Main Details](#) [Account Memos](#) [Sales History](#) [Purchase Orders](#) [Supplier Returns](#) [Stock References](#) [Cust./Supp. Returns](#) [Service Visits](#)
- [Distributor Agreements](#) [Documents Files](#) [Linked Issues](#) [Regulatory Agreements](#)

MHRA Registration number: 29372

BOX:924

Manufacturer RMA : 300265364

☒ Repaired ☐ No Fault Found[Returned Barcode](#)[Close File 924](#)[Tick All Replaced](#)

Review	Stock Ref	Sup Ref	B/C	S/N	Fault	Replaced	No Fault Found	Repaired	Credited	Out Warranty	Returned Unrepaired	Notes
Review 8010004 Customer Stats	8010004		2164604	JJ143173	Unstable output signal Country of Origin: Germany Customs value: 5.00 euro	☐	☐	☐	☒	☐	☐	
Review 8010007 Customer Stats	8010007		2048585	S01116602	Unstable output signal Country of Origin: Germany Customs value: 5.00 euro	☐	☐	☐	☒	☐	☐	
Review 8010018 Customer Stats	8010018		2210974	M3S104197	Low Output Country of Origin: Germany Customs value: 5.00 euro	☐	☐	☐	☒	☐	☐	

[Update Returns](#)

Customer Failure Statistics

Start From 18/08/2021 End at 18/08/2023

Invoice Correction 5264416848

ORIGINAL

Invoice Correction Date: 11.08.2023

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

Purchase Order No: RTS3100924

23966 Wismar

Germany

Bill to: 1797448

Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

Sold to: 1797448

Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

Order Information:

Credit Memo Request: 64117321
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797448

Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

EM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
00010	E1002736 Oxygen Sensor OOD103 VAN 8010018 Customer Part Nbr: 8010018 credit note for SN: 104197	1	EA	20,39	-20,39	0%
00020	E1001581 Oxygen Sensor OOD103-1V Customer Part Nbr: 8010007 credit note for SN: 116602	1	EA	24,15	-24,15	0%
00030	E1001721 Oxygen Sensor OOD103-2V Customer Part Nbr: 8010004 credit note for SN: 143173	1	EA	28,63	-28,63	0%

Ins #:

Link Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

AN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat


Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner +49 3841 360 206

Unless otherwise agreed in writing by the parties, all sales of products and/or services in this document shall be governed solely by the applicable Honeywell Terms and Conditions in effect at the time a Purchase Order is accepted by Honeywell, a copy of which can be found at <https://sps.honeywell.com/us/en/support/safety/legal-documents/sst>. Honeywell's acknowledgment of receipt of a Purchase Order shall not constitute acceptance. By continuing with this transaction, you (Buyer) agree to those Honeywell Terms and Conditions, which are subject to change at Honeywell's discretion without notice. Honeywell expressly limits its acceptance, fulfillment and performance of this Purchase Order to the terms in the Honeywell Terms and Conditions and expressly rejects any different, conflicting or additional terms in the Purchase Order or any other Buyer documents. This sale shall be governed by the Terms and Conditions of the Honeywell legal entity listed on this document.

PRÜFBERICHT – TEST REPORT

Sauerstoffsensoren/Oxygen Sensors OOA, OOI, OOD



Kunde/ Customer	Vendagraph	Retour/ RMA- No.	Kunde/Customer	RTS 3100914	Envitec	390265364
Artikel/ Article	000103;000103-14;000103-24	Anzahl/ Quantity	11.1.1	Eing.Datum/ Inc.Date	12.07.23	
Bearbeitet/ processed	U. B. M.	Prüfdatum/ Insp. Date	14.07.23	Geprüft/ Inspected	3. M.	

Produkt/ Identification			Messwert/ Test Result				Bewertung/ Evaluation			
Nr.	Seriennummer/ Serial no.	Lieferlos/ Manufact. code	Luftwert/ Air signal @ 20,9% O ₂ [mV]	Nullsignal/ Offset @ 0% O ₂ [µV]	Linearität/ Linearity [%]	Bemerkung/ Remark	außerhalb Garantie/ out of warranty	kein Fehler festgestellt/ no failure found	Fehlerkode/ Failure- code	Austausch / Replacement
1	104 197	04123	0						04	
2	116602	07122	11.3	35	i.o.			+		
3	143173	02123	11,3	35	fehlerhaft				02	
4										
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										
16										
17										
18										
19										
20										

Fehler O2-S001: Luftwert außerhalb Toleranz/ Air signal out of tolerance
 Fehler O2-S003: Offset außerhalb Toleranz/ Offset out of tolerance
 Fehler O2-S005: Gehäuse defekt/ Housing defective

Fehler O2-S002: Signal instabil / Signal not stable
 Fehler O2-S004: Signal 0mV / Signal 0mV
 Fehler O2-S006: Elektrolyt Austritt / Electrolyte leakage