

Service Repair Sheet SRS68449

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Owen Daniels

Royal Albert Edward Infirmary

Medical Electronics

Medical Engineering Technician

01942822002

01942822002

00005363

Owen.Daniels@wwl.nhs.uk

14/Aug/2023

Robert Connor

Viamed

Quote

VIAMEDclear

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 26/Apr/2023 Kate Griffiths

26/Apr/2023 Kate Griffiths

Would like to send in V1000, s/n: PR02724A19 for service and calibration. I asked for decontamination cert.

08/Aug/2023 Steve Hardaker

Owen contacted us again, is sending now this back for service. Advised:

- p/n 1480000 - Foetal Heart Simulator V1000 Service @ £60.00+VAT

- Return carriage: £12.00+VAT

14/Aug/2023 Robert Connor

Received 1 x V1000 s/n PR02724A19, with 4 x AA battery, grey plastic carry case, metal cylinder, 2-pin connector.

Ready For quote

CGreen 15-8-23

Repair Complete Signed

CGreen 18-8-23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35967	Foetal Simulator	1410000	PR02724A19	

1480000 x 1 @ £60

S/N, SRS, SRN

1430309 x 1 @ FOC

SRS, SRN

UPS x 1 @ £12

Service Repair Sheet SRS68449

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Owen Daniels

Royal Albert Edward Infirmary

Medical Electronics

Medical Engineering Technician

01942822002

01942822002

00005363

Owen.Daniels@wwl.nhs.uk

14/Aug/2023

Robert Connor

Viamed

Quote

VIAMEDclear

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 26/Apr/2023 Kate Griffiths

26/Apr/2023 Kate Griffiths

Would like to send in V1000, s/n: PR02724A19 for service and calibration. I asked for decontamination cert.

08/Aug/2023 Steve Hardaker

Owen contacted us again, is sending now this back for service. Advised:

- p/n 1480000 - Foetal Heart Simulator V1000 Service @ £60.00+VAT

- Return carriage: £12.00+VAT

14/Aug/2023 Robert Connor

Received 1 x V1000 s/n PR02724A19, with 4 x AA battery, grey plastic carry case, metal cylinder, 2-pin connector.

Ready For quote

CGreen 15-8-23

Repair Complete Signed

CGreen 18-8-23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35967	Foetal Simulator	1410000	PR02724A19	

1480000 x 1 @ £60

S/N, SRS, SRN

1430309 x 1 @ FOC

SRS, SRN

UPS x 1 @ £12

Service Repair Sheet 68449

Contact Name	Owen Daniels
Company/ Hospital Name	Royal Albert Edward Infirmary
Department	Medical Electronics
Position	Medical Engineering Technician
Direct Phone	01942822002
General Phone	01942822002
Opera Account	00005363
Email	Owen.Daniels@wwl.nhs.uk
Order Number	319167052
Date Received	14/Aug/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35967	1410000	PR02724A19	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

PURCHASE ORDER

Page: 1 of 1

WRIGHTINGTON, WIGAN & LEIGH TEACHING



Supplier:

VIAMED LTD
15 STATION ROAD
CROSS HILLS
KEIGHLEY
BD20 7DT

GLN:

Buyer DANIELLE RRF JOHNSON
Telephone
Email danielle.johnson@wwl.nhs.uk

RRF9009 RAEI MEDICAL ELECTRONICS

Deliver to:

RAEI GENERAL STORES
ROYAL ALBERT EDWARD INFIRMARY
WIGAN LANE
WIGAN, WN1 2NN

Invoice to:

WRIGHTINGTON, WIGAN & LEIGH
TEACHING HOSPITALS NHS FOUNDATION
RRF PAYABLES F615
PO BOX 312
LEEDS, LS11 1HP

0303 123 1177
GLN:

Order Number 319167052

Date 14-AUG-23

Purchase Order issued in accordance with the appropriate NHS Terms & Conditions for the Supply of Goods and for the Provision of Services

Should the cost of any item requested within differ from that stated, if any item is currently out of stock, or if a carriage charge needs to be added, please contact the named buyer immediately before proceeding

Please note any specific delivery requirement below and ensure delivery is made to the "Deliver To" address NOT the invoice address
All goods supplied which are covered by COSHH and CHIP Regulations MUST be accompanied by the relevant safety data sheet upon each despatch

Payment Terms: Net Terms Monthly

Quantity Required	U.O.M.	Supplier Part Number	Description	Delivery Date	Unit Price Including Discount	Line Value GBP
----------------------	--------	----------------------	-------------	------------------	----------------------------------	-------------------

1	EACH	1480000	Maintenance. Foetal Heart Simulator V1000 Service.	17-AUG-23	60.00	60.00
Serial Number=PR02724A19						
Job/Quote Ref=SRS68449						

Total Value of Order (Exc VAT)

60.00

PURCHASE ORDER

Page: 1 of 1

WRIGHTINGTON, WIGAN & LEIGH TEACHING



Supplier:

VIAMED LTD
15 STATION ROAD
CROSS HILLS
KEIGHLEY
BD20 7DT

GLN:

Buyer DANIELLE RRF JOHNSON
Telephone
Email danielle.johnson@wwl.nhs.uk

RRF9009 RAEI MEDICAL ELECTRONICS

Deliver to:

RAEI GENERAL STORES
ROYAL ALBERT EDWARD INFIRMARY
WIGAN LANE
WIGAN, WN1 2NN

Invoice to:

WRIGHTINGTON, WIGAN & LEIGH
TEACHING HOSPITALS NHS FOUNDATION
RRF PAYABLES F615
PO BOX 312
LEEDS, LS11 1HP

0303 123 1177
GLN:

Order Number 319167052

Date 14-AUG-23

Purchase Order issued in accordance with the appropriate NHS Terms & Conditions for the Supply of Goods and for the Provision of Services

Should the cost of any item requested within differ from that stated, if any item is currently out of stock, or if a carriage charge needs to be added, please contact the named buyer immediately before proceeding

Please note any specific delivery requirement below and ensure delivery is made to the "Deliver To" address NOT the invoice address
All goods supplied which are covered by COSHH and CHIP Regulations MUST be accompanied by the relevant safety data sheet upon each despatch

Payment Terms: Net Terms Monthly

Quantity Required	U.O.M.	Supplier Part Number	Description	Delivery Date	Unit Price Including Discount	Line Value GBP
-------------------	--------	----------------------	-------------	---------------	-------------------------------	----------------

1	EACH	1480000	Maintenance. Foetal Heart Simulator V1000 Service.	17-AUG-23	60.00	60.00
			Serial Number=PR02724A19			
			Job/Quote Ref=SRS68449			

Total Value of Order (Exc VAT)

60.00