

10:35

Hi Ryan, Can you tell me if we have an O2 Sensor compatible with a SIPAP Ventilator Carefusion Model (Infantflow) - I can not find anything on the system. Did look on the website and a MAX-250E came up but I have no idea if this is correct. Please advise - Modern Technology Company are asking for a price



Ryan, 10:47

Hi Janine, they need the R-22MEDV - 0110021

