

Service Repair Sheet SRS68557

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Matthew Geddes
Queen Elizabeth University Hospital
Medical Physics
Senior Assistant Technical Officer
01414523781

00001853
matthew.geddes@ggc.scot.nhs.uk

30/Aug/2023
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒
Signed: RC
Date: 30/08/23

Goods Out Only
Cleaned by Viamed before
returning to customer
Signed: _____
Date: _____

Notes 29/Aug/2023 Sophie Lines

29/Aug/2023 Sophie Lines

Customer has a V1000 (serial number PR02041A10) they would like to send for service/calibration. I have sent him instructions on how to return and given him SRS number.

30/Aug/2023 Kate Griffiths

PO GMPS13277546 rec d, attached

30/Aug/2023 Robert Connor

Received 1 x V1000 s/n PR02041A10, with 4 x AA battery

Ready For quote

CGreen 1-9-23

Repair Complete Signed

CGreen 7-9-23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35976	Foetal Simulator	1410000	PR02041A10	<u>N</u>

1480000 x 1 £60
S/N, SRS, SRN

1430309 x 1 fOC
SRS, SRN

ups > 1 £12.

Service Repair Sheet 68557

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
Opera Account
Email
Order Number
Date Received
Booked in By

Matthew Geddes
Queen Elizabeth University Hospital
Medical Physics
Senior Assistant Technical Officer
01414523781
00001853
matthew.geddes@ggc.scot.nhs.uk
GMPS13277546
30/Aug/2023
Robert Connor

Repair	Ref	S/N	Equipment Type
SRN35976	1410000	PR02041A10	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked

PURCHASE ORDER: GMPS13277546

Buyer NHS GREATER GLASGOW & CLYDE EnquiriesTo: Support@ggcprocurement customerservices.zendesk.com If Urgent Please Contact -, 0141 2111200	Order Date 29 Aug 2023	Invoice To NHS GREATER GLASGOW AND CLYDE PAYMENTS DEPARTMENT (nhsgg&c.payments@ggc.scot.nhs.uk) PO BOX 7388 GLASGOW, G51 9BS
Supplier Vamed 15 Station Road Crosshills Keighley, West Yorks BD20 7DT	Order Contact Name: Matthew Geddes Phone: 0141 452 3282 matthew.geddes@ggc.scot.nhs.uk hs.uk	Delivery Matthew Geddes DEL NOT DEFINED MEDICAL PHYSICS DEPT, 2ND FLOOR TOWER BLOCK SOUTH GLASGOW UNIVERSITY HOSPITAL 1345 GOVAN ROAD GLASGOW, G51 4TF

Delivery Information

Order Type:	Direct Ship
Carrier:	Not Selected -Not Selected
FOB - Delivery Terms:	Not Selected -Not Selected

Payment Information

Customer Number:	GGC3142-0139
Payment Terms:	Not Selected

Additional Order Information

VAT Number:	VAT Number 654850811
EORI Number:	EORI Number GB654850811000

Line	Item Type	Item No	Manufacturer No	UOM	Pack Size	Qty	Unit Price	VAT Type	Extended Amt
	Description							Est VAT	
1	Non-Catalogue	unknown		Each		1	£60.00	SI	£60.00
	service and calibration of the V1000 Foetal Heart Simulator SN: PR02041A10							£12.00	
2	Non-Catalogue	unknown		Each		1	£12.00	SI	£12.00
	carriage (SRS68557)							£2.40	
							Total Extended Amount:		£72.00
							Total Estimated VAT:		£14.40
							Estimated Gross Amount:		£86.40

VAT Types

Key	Description	Estimated VAT
SI	SI - STD IRRECOVERABLE	£14.40