

Service Repair Sheet SRS68562

Contact Name

Chris Kavanagh

Company/ Hospital Name

Just Inc Limited

Department

Position

Direct Phone

07738 828787

General Phone

Opera Account

CID14856

Email

chris@thinairsports.com

Order Number

DVM138640-1

Date Received

12/Sep/2023

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 05/Sep/2023 Aqib Majeed

05/Sep/2023 Aqib Majeed

Customer sending back 1 x 0014850 Pulse Oximetry Ear Sensor. S/N: HAR10018

Spring has come loose and do not attach to ear

12/Sep/2023 Robert Connor

Received 1 x ear clip s/n HAR10018, with blue cable guide

Ready For quote

Repair Complete Signed

CGreen 20.9.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN35998	Pulse Oximetry Ear Clip Sensor	0014850	HAR10018	Yes

0040017 - CG 22.9.23 - did not see this one before CG-22.9.23
~~0040000~~ x 1 FOC
 S/N, SRS, SRN

UPS x 1 FOC