



Kate Griffiths <viamed.kate.griffiths@gmail.com>

FW: Oxygen Sensor - záruka?

2 messages

Oskar Šlechta <o.slechta@chromservis.eu>

27 September 2023 at 07:58

To: "kate.griffiths@viamed.co.uk" <kate.griffiths@viamed.co.uk>

Cc: "viamedinbox@gmail.com" <viamedinbox@gmail.com>

Dear Kate,

unfortunately, we received a new claim for **R-22A sensor, s/n 919838**, received in March 2023 under our order DE2023-124 (we ordered 50 sensors together under this order). See enclosed picture for sensor details.

After installation, the value was 3mV only. This sensor doesn't work. I would like to ask you for start of claim process. Of course, we'll send faulty sensor you to back.

Thanks so much, best regards

Oskar

Oskar Šlechta

Senior Internal Sales Advisor

Chromservis [Praha](#)[Jakobiho 327](#)[CZ-10900 Praha 10](#)

Czech Republic

Tel.: +420 274 021 221

Fax: +420 274 021 220

E-mail: o.slechta@chromservis.euwww.chromservis.eu