



John Lamb <viamed.john.lamb@gmail.com>

Phototherapy eye masks Project 29 447072 messages

John Lamb <john.lamb@viamed.co.uk>
To: Alexander Yin <alexander99@yahoo.com>
Cc: derek@viamed.co.uk

Mon, Sep 2, 2013 at 11:55 AM

Dear Alexander

Have you carried out a full Risk Analysis?.

Did you use the EN14971 : 2009
If yes is it possible to have a copy

You should also have in your Technical Files a Risk Analysis Certificate which you generate and sign to confirm you have carried out a Risk Assessment.

I have found that using the prior EN14971 Annex A and Annex D made the exercise much easier to understand and complete. If you do not have these I can send them too you.

If your analysis is only in Chinese just send it we will construct our own Annex A and Annex D in English.

Kind regards

John

Alexander Yin <alexander99@yahoo.com>
Reply-To: Alexander Yin <alexander99@yahoo.com>
To: John Lamb <john.lamb@viamed.co.uk>
Cc: "derek@lamb.uk.net" <derek@lamb.uk.net>

Tue, Sep 3, 2013 at 10:55 AM

Dear Mr. John,

Yes, we carried out the full risk analysis according to EN14971:2009.

Enclosed please find the risk analysis report for your reference.

Thanks & best regards,

Alexander

From: John Lamb <john.lamb@viamed.co.uk>
To: Alexander Yin <alexander99@yahoo.com>
Cc: derek@viamed.co.uk
Sent: Monday, September 2, 2013 6:55 PM
Subject: Phototherapy eye masks Project 29 44707

[Quoted text hidden]

This message has been scanned for malware.

 **risk analysis report.doc**
150K