

Resuscitation Cabinet 2002



**Viamed Limited. 15 Station Road, Cross Hills,
Keighley, West Yorkshire BD20 7DT, United Kingdom**

Tel: +44 (0)1535 634542 Fax: +44 (0)1535 635582

E-mail: info@viamed.co.uk Web site: www.viamed.co.uk

**Viamed Limited, 15 Station Road, Cross Hills
Keighley, West Yorkshire, BD20 7DT, United Kingdom
Tel: +44 (0)1535 634542 Fax: +44 (0)1535 635582
E-mail: info@viamed.co.uk Website: www.viamed.co.uk**

**Infant Resuscitation Cabinet
Serial number: E201**



Caution

Do not apply excessive force to the treatment platform.

Maximum loading: 25Kg

Ensure that the door is free of obstructions before closing.

Do not attempt to close the clasp if force is required to close the door.